

mmB

**Department of Forensic Science
Breath Alcohol Section
Instrument Maintenance History
1/1/1990 To 3/14/2014**

Instrument Serial Number: 010602

As of 14-Mar-14

Maintenance Date	Certification Date	Technician Initials	Remarks
14-Mar-14	22-Nov-13	MMB (26417)	REPLACED PRINTER.
22-Nov-13	22-Nov-13	MMB (26417)	CERTIFIED.
30-Sep-13	30-Sep-13	TLN (25387)	FIRMWARE/SOFTWARE UPGRADE. CALIBRATED AND CERTIFIED.
08-Jan-13	08-Jan-13	TLN (25387)	CERTIFIED. REPLACED DRY GAS STANDARD.
04-Sep-12	19-Jul-12	DBH (21250)	REPLACED DRY GAS STANDARD.
19-Jul-12	19-Jul-12	SED (24178)	CERTIFIED. REPLACED DRY GAS STANDARD.
02-Feb-12	02-Feb-12	TLN (25387)	CERTIFIED
17-Aug-11	17-Aug-11	DBH (21250)	REPLACED DRY GAS STANDARD, CERTIFIED
09-Mar-11	09-Mar-11	DBH (21250)	CERTIFIED
21-Sep-10	21-Sep-10	DBH (21250)	CALIBRATED AND CERTIFIED.
12-Jul-10	06-Apr-10	DBH (21250)	REPLACED DRY GAS STANDARD
06-Apr-10	06-Apr-10	DBH (21250)	CERTIFIED
20-Oct-09	20-Oct-09	DBH (21250)	CERTIFIED.
14-May-09	14-May-09	DBH (21250)	FIRMWARE/SOFTWARE UPGRADE, CALIBRATED AND CERTIFIED.
12-Jan-09	12-Jan-09	DBH (21250)	CERTIFIED.
25-Nov-08	25-Nov-08	NMJ (22392)	CALIBRATED, CERTIFIED AND ACCEPTED.

mmB

INTOX EC/IR II Quality Assurance Worksheet

Instrument Serial Number: 010602 Worksheet Start Date: 3/14/2014

Location: Prince William Garfield

Address: 15948 Donald Curtis Dr Woodbridge VA 22191

DFS Technician: Michelle Billeter License No.: 26417

Laboratory On-Site

Site Specification: No detrimental environmental conditions exist.

Instrument Barometer (mm HG): 762 Reference Barometer (mm HG): 765

Reference Barometer(RB)Serial #: 009111 RB Calibration Due: 5/23/2014

Measurement Assurance Check

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.300		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.100		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Dry gas standard Lot No. (with tank no.) AG320502-11

- Replaced dry gas standard (+O-ring)
- Installed at Location
- Removed to DFS-Central

Supplies
Mouthpieces
Certificates of Analysis
Operator Worksheet
Other:



Notes:

Replaced Printer.

Instrument Serial Number

010602

Certification Date

Calibrated

Certified

Measurement Assurance Check

Instrument Test

Troubleshooting/Maintenance Worksheet Completed

(Successful completion denotes satisfactory condition of the item.)

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.250		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.150		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.080		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.020		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

All measurements are in g/210L

Estimation of Uncertainty of Measurement and traceability records are located within the Breath Alcohol Section

DFS Technician

Michelle Bellita

Date

3/14/14

Issuing Analyst

Michelle Bellita

Date

3/14/14

msk

mmb



COMMONWEALTH OF VIRGINIA
DEPARTMENT OF FORENSIC SCIENCE

CERTIFICATE OF BLOOD ALCOHOL ANALYSIS
AS DETERMINED BY A CHEMICAL TEST OF THE ACCUSED'S BREATH

NAME OF ACCUSED INSTRUMENT, TEST,	NAME OF COURT DFS
--	--------------------------

BREATH ANALYSIS

SAMPLE EXAMINED AND TEST CONDUCTED BY BILLETER, MICHELLE, M		AGENCY DFS Central Lab
DFS LICENSE NUMBER 26417	LICENSE EXPIRES 10/01/2014	DATE TEST CONDUCTED 03/14/2014
TEST EQUIPMENT NUMBER 010602	[REDACTED]	

RESULTS: TIME SAMPLE TAKEN 11:30 EDT

SAMPLE'S ALCOHOL CONTENT 0.00 GRAMS PER 210 LITERS OF BREATH

ATTEST:

I CERTIFY THAT THE ABOVE IS AN ACCURATE RECORD OF THE TEST CONDUCTED; THAT THE TEST WAS CONDUCTED WITH THE TYPE OF EQUIPMENT AND IN ACCORDANCE WITH THE METHODS APPROVED BY THE DEPARTMENT OF FORENSIC SCIENCE; THAT THE TEST WAS CONDUCTED IN ACCORDANCE WITH THE DEPARTMENT'S SPECIFICATIONS; THAT PRIOR TO ADMINISTRATION OF THE TEST THE ACCUSED WAS ADVISED OF HIS RIGHT TO OBSERVE THE PROCESS AND SEE THE BLOOD ALCOHOL READING ON THE EQUIPMENT USED TO PERFORM THE BREATH TEST, AND THAT I POSSESS A VALID LICENSE TO CONDUCT SUCH TEST, GIVEN UNDER MY HAND THIS _____ DAY OF _____, 20_____.

BREATH TEST OPERATOR

I HAVE RECEIVED A COPY OF THIS CERTIFICATE OF ANALYSIS _____
SUBJECT'S SIGNATURE

SUBJECT REFUSED TO SIGN FOR COPY OF CERTIFICATE OF ANALYSIS _____
OPERATOR'S SIGNATURE

mm8



IntoxNet MIS Report

Report Generated 14 Mar 2014 at 14:25

Test Results

Instrument Serial Number 010602

Test # 000665 **Subject Test**

Test Location 1 Department of
Test Date 14 Mar 2014

Test Location 2 Forensic Science
Test Time 11:23
Remote/Local Local

Test Location 3
System Check Passed

Operator's Last Name BILLETER
Agency DFS Central Lab

Operator's First Name MICHELLE

Operator's Middle Initial M

License Number 26417

Card Serial Number 126417

Effective Date 10/01/2012

Expiration Date 10/01/2014

Subject's Last Name INSTRUMENT

Subject's First Name TEST

Subject's Middle Initial

Subject's Date of Birth 00/00/0000

Subject's Sex Male

Driver's License Number

Driver's License Expiration 00/00/0000

Driver's License State

Court Name DFS

End Date 14 Mar 2014

End Time 11:31

Result Time 11:30

Result Date 14 Mar 2014

Result 0.00

Data Type DIAG

Sample Value Pass

Sample Time 11:24

Data Type BLK

Sample Value 0.000

Sample Time 11:24

Data Type CHK

Sample Value 0.100

Sample Time 11:25

Data Type BLK

Sample Value 0.000

Sample Time 11:26

Data Type SUBJ

Sample Value 0.000

Sample Time 11:27

Data Type BLK

Sample Value 0.000

Sample Time 11:28

Data Type SUBJ

Sample Value 0.000

Sample Time 11:30

Data Type BLK

Sample Value 0.000

Sample Time 11:31

Standard Type Dry Gas Std

Standard Value 0.100

Standard Lot Number AG320502-11

Standard Expiration Date 07/24/2015

Tank Pressure 330

Barometric Pressure 762 mmHg

Blow Sample Number 1

Blow Duration 3.72 sec

Blow Volume 1692 cc

End-of-Blow Time 11:27

Blow Sample Number 2

Blow Duration 3.85 sec

Blow Volume 1801 cc

End-of-Blow Time 11:30

Tamper Evident Stamp 337c0365

Test Status Code 0

Test Status Success

Department of Forensic Science
Breath Alcohol Section
Trouble Call Log
1/1/1990 To 3/13/2014

MMB
MBL

Instrument Serial Number: 010602

13-Mar-14

Date: 02-Sep-12 Time: 4:17 AM Remote: True Tech: DBH (21250) Location: Middlesex County Sheriff's Office

Indication: OTHER

Problem: OTHER. INSTRUMENT DISPLAYING "DRY GAS PRESSURE LOW" MESSAGE.

Technician Response: SITE VISIT CONDUCTED. ISSUE RESOLVED.

Resolution NO FURTHER ACTION REQUIRED

Date: 15-Apr-13 Time: 11:40 AM Remote: True Tech: MMB (26417) Location: Middlesex County Sheriff's Office

Indication: OPERATOR QUESTION / COMMENT

Problem: OPERATOR QUESTION / COMMENT. INSTRUMENT DISPLAYING INVALID SAMPLE MESSAGES.

Technician Response: INSTRUMENT CONTACTED VIA MODEM. HAD OPERATOR CONDUCT PRACTICE TEST. INSTRUMENT FUNCTIONING PROPERLY. SITE VISIT CONDUCTED.

Resolution REMOVED TO DFS-CENTRAL.

Date: 13-Mar-14 Time: 8:18 AM Remote: True Tech: MMB (26417) Location: Prince William County PD

Indication: PRINTER NOT FUNCTIONING PROPERLY

Problem: PRINTER NEEDS RIBBON REPLACED.

Technician Response: SITE VISIT CONDUCTED. ISSUE RESOLVED.

Resolution NO FURTHER ACTION REQUIRED
