

Department of Forensic Science
Breath Alcohol Section
Instrument Location History
1/1/1990 To 3/20/2014

DBH
gud

Instrument Serial Number: 010617 As of 20-Mar-14

Start Date	End Date	Agency Name
19-Mar-14		DFS Central Lab
30-Jul-13	19-Mar-14	Virginia Beach PD
14-Jan-13	30-Jul-13	DFS Central Lab
28-Dec-12	14-Jan-13	Prince William County PD
30-Nov-12	28-Dec-12	DFS Central Lab
05-Jan-09	30-Nov-12	King William County Sheriff's Office
15-Aug-08	05-Jan-09	DFS Central Lab

DD11
red

INTOX EC/IR II Quality Assurance Worksheet

Instrument Serial Number **010617** Worksheet Start Date **3/19/2014**
 Location **Virginia Beach Jail**
 Address **2501 James Madison Blvd Virginia Beach VA 23456**
 DFS Technician **Donald Hall** License No. **21250**

Laboratory On-Site

Site Specification: No detrimental environmental conditions exist.

Instrument Barometer (mm HG) **763** Reference Barometer (mm HG) **764**

Reference Barometer(RB)Serial # **009109** RB Calibration Due **8/1/2014**


Measurement Assurance Check

Standard (sea level)	PA Target	minimum	maximum	Sample 1	Sample 2	Sample 3
0.300	0.301	0.291	0.310	0.298	0.298	0.297
Precision		sample min	sample max			
0.001		0.297	0.298			

Standard (sea level)	PA Target	minimum	maximum	Sample 1	Sample 2	Sample 3
0.100	0.100	0.097	0.103	0.099	0.099	0.099
Precision		sample min	sample max			
0		0.099	0.099			

Dry gas standard Lot No. (with tank no.) **AG227501-28**

- Replaced dry gas standard (+O-ring)
 - Installed at Location
 - Removed to DFS-Central
- Supplies
 Mouthpieces
 Certificates of Analysis
 Operator Worksheet
 Other:



Notes:

[Large textured area, likely a redacted note or signature]

Instrument Serial Number

010617

Certification Date

Calibrated

Certified

Measurement Assurance Check

Instrument Test

Troubleshooting/Maintenance Worksheet Completed

(Successful completion denotes satisfactory condition of the item.)

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.250		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.150		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.080		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.020		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

All measurements are in g/210L

Estimation of Uncertainty of Measurement and traceability records are located within the Breath Alcohol Section

DFS Technician

Donald B Hall

Date

3/20/14

Issuing Analyst

gjl

Date

3/25/14

DBH

sd



COMMONWEALTH OF VIRGINIA
DEPARTMENT OF FORENSIC SCIENCE

CERTIFICATE OF BLOOD ALCOHOL ANALYSIS
AS DETERMINED BY A CHEMICAL TEST OF THE ACCUSED'S BREATH

NAME OF ACCUSED INSTRUMENT, TEST,	NAME OF COURT DFS
--	--------------------------

BREATH ANALYSIS

SAMPLE EXAMINED AND TEST CONDUCTED BY HALL, DONALD, B		AGENCY DFS Central Lab
DFS LICENSE NUMBER 21250	LICENSE EXPIRES 10/01/2014	DATE TEST CONDUCTED 03/19/2014
TEST EQUIPMENT NUMBER 010617		

RESULTS: TIME SAMPLE TAKEN 12:27 EDT

SAMPLE'S ALCOHOL CONTENT 0.00 GRAMS PER 210 LITERS OF BREATH

ATTEST:

I CERTIFY THAT THE ABOVE IS AN ACCURATE RECORD OF THE TEST CONDUCTED; THAT THE TEST WAS CONDUCTED WITH THE TYPE OF EQUIPMENT AND IN ACCORDANCE WITH THE METHODS APPROVED BY THE DEPARTMENT OF FORENSIC SCIENCE; THAT THE TEST WAS CONDUCTED IN ACCORDANCE WITH THE DEPARTMENT'S SPECIFICATIONS; THAT PRIOR TO ADMINISTRATION OF THE TEST THE ACCUSED WAS ADVISED OF HIS RIGHT TO OBSERVE THE PROCESS AND SEE THE BLOOD ALCOHOL READING ON THE EQUIPMENT USED TO PERFORM THE BREATH TEST, AND THAT I POSSESS A VALID LICENSE TO CONDUCT SUCH TEST, GIVEN UNDER MY HAND THIS _____ DAY OF _____, 20 _____.

BREATH TEST OPERATOR

I HAVE RECEIVED A COPY OF THIS CERTIFICATE OF ANALYSIS _____ SUBJECT'S SIGNATURE

SUBJECT REFUSED TO SIGN FOR COPY OF CERTIFICATE OF ANALYSIS _____ OPERATOR'S SIGNATURE



IntoxNet MIS Report
Report Generated 19 Mar 2014 at 13:22

HMS
DBH
SJD

Test Results

Instrument Serial Number 010617

Test # 001662 Subject Test

Test Location 1 Department of
Test Date 19 Mar 2014

Test Location 2 Forensic Science
Test Time 12:21
Remote/Local Local

Test Location 3
System Check Passed

Operator's Last Name HALL
Agency DFS Central Lab

Operator's First Name DONALD

Operator's Middle Initial B

Card Serial Number 121250

License Number 21250

Effective Date 10/01/2012

Expiration Date 10/01/2014

Subject's Last Name INSTRUMENT

Subject's First Name TEST

Subject's Middle Initial

Subject's Date of Birth 00/00/0000

Subject's Sex Male

Driver's License Number

Driver's License Expiration 00/00/0000

Driver's License State

Court Name DFS

End Date 19 Mar 2014 End Time 12:29

Result Time 12:27

Result Date 19 Mar 2014 Result 0.00

Data Type DIAG

Sample Value Pass

Sample Time 12:21

Data Type BLK

Sample Value 0.000

Sample Time 12:22

Data Type CHK

Sample Value 0.100

Sample Time 12:22

Data Type BLK

Sample Value 0.000

Sample Time 12:24

Data Type SUBJ

Sample Value 0.000

Sample Time 12:25

Data Type BLK

Sample Value 0.000

Sample Time 12:26

Data Type SUBJ

Sample Value 0.000

Sample Time 12:27

Data Type BLK

Sample Value 0.000

Sample Time 12:28

Standard Type Dry Gas Std

Standard Value 0.100

Standard Lot Number AG227501-28

Standard Expiration Date 10/01/2014

Tank Pressure 620

Barometric Pressure 762 mmHg

Blow Sample Number 1 Blow Duration 3.27 sec

Blow Volume 1784 cc End-of-Blow Time 12:25

Blow Sample Number 2 Blow Duration 3.07 sec

Blow Volume 1784 cc End-of-Blow Time 12:27

Tamper Evident Stamp fe722d59

Test Status Code 0

Test Status Success

DBH

sed

Intox EC/IR-II: Accuracy Check

Department of Forensic Science

Serial Number: 010617 Test Number: 1663

Test Date: 03/19/2014 Test Time: 12:32 EDT

Dry Gas Target: 0.100

Lot Number: AG227501-28 Exp Date: 10/01/2014

Tank Pressure: 615 psi Barometric Pressure: 762 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	12:33
CHK	0.099	12:34
BLK	0.000	12:35
CHK	0.099	12:36
BLK	0.000	12:38
CHK	0.099	12:38

Test Status: *Success*

Calibration CRC: 57CA468F

DB11
Sed

Intox EC/IR-II: Accuracy Check

Department of Forensic Science

Serial Number: 010617 Test Number: 1664

Test Date: 03/19/2014 Test Time: 12:42 EDT

Dry Gas Target: 0.301

Lot Number: AG322702-06 Exp Date: 08/15/2015

Tank Pressure: 213 psi Barometric Pressure: 762 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	12:43
CHK	0.298	12:43
BLK	0.000	12:45
CHK	0.298	12:46
BLK	0.000	12:48
CHK	0.297	12:49

Test Status: *Success*

Calibration CRC: 57CA468F