

mmb

Department of Forensic Science
Breath Alcohol Section
Instrument Location History
1/1/1990 To 5/21/2014

Instrument Serial Number: 010621 **As of 21-May-14**

Start Date	End Date	Agency Name
20-May-14		DFS Central Lab
13-Jan-09	20-May-14	Mathews County SO
04-Aug-08	13-Jan-09	DFS Central Lab

mmB

INTOX EC/IR II Quality Assurance Worksheet

Instrument Serial Number **010621** Worksheet Start Date **5/20/2014**
 Location **Mathews Co SO**
 Address **10622 Buckley Hall Rd Mathews VA 23109**
 DFS Technician **Michelle Billeter** License No. **26417**
 Laboratory On-Site

Site Specification: No detrimental environmental conditions exist.

Instrument Barometer (mm HG) **764** Reference Barometer (mm HG) **765**
 Reference Barometer(RB)Serial # **009109** RB Calibration Due **8/1/2014**

Measurement Assurance Check


Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.307
0.300	0.302	0.292	0.311	Sample 2	0.307
Precision		sample min	sample max	Sample 3	0.306
0.001		0.306	0.307		

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.102
0.100	0.101	0.097	0.104	Sample 2	0.102
Precision		sample min	sample max	Sample 3	0.102
0		0.102	0.102		

Dry gas standard Lot No. (with tank no.) **AG320502-29**

- Replaced dry gas standard (+O-ring)
- Installed at Location
- Removed to DFS-Central

Supplies
 Mouthpieces
 Certificates of Analysis
 Operator Worksheet
 Other:



Notes:

[Large shaded area for notes]

Instrument Serial Number

010621

Certification Date

Calibrated

Certified

Measurement Assurance Check

Instrument Test

Troubleshooting/Maintenance Worksheet Completed

(Successful completion denotes satisfactory condition of the item.)

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.250		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.150		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.080		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.020		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

All measurements are in g/210L

Estimation of Uncertainty of Measurement and traceability records are located within the Breath Alcohol Section

DFS Technician

Michelle Bullock

Date

5/21/14

Issuing Analyst

Michelle Bullock

Date

5/21/14

TLN

MMB



COMMONWEALTH OF VIRGINIA
DEPARTMENT OF FORENSIC SCIENCE

CERTIFICATE OF BLOOD ALCOHOL ANALYSIS
AS DETERMINED BY A CHEMICAL TEST OF THE ACCUSED'S BREATH

NAME OF ACCUSED INSTRUMENT, TEST,	NAME OF COURT DES
---	-----------------------------

BREATH ANALYSIS

SAMPLE EXAMINED AND TEST CONDUCTED BY BILLETER, MICHELLE, M		AGENCY DES Central Lab
DFS LICENSE NUMBER 26417	LICENSE EXPIRES 10/01/2014	DATE TEST CONDUCTED 05/20/2014
TEST EQUIPMENT NUMBER 010621		

RESULTS: TIME SAMPLE TAKEN 10:33 EDT

SAMPLE'S ALCOHOL CONTENT 0.00 GRAMS PER 210 LITERS OF BREATH

ATTEST:

I CERTIFY THAT THE ABOVE IS AN ACCURATE RECORD OF THE TEST CONDUCTED; THAT THE TEST WAS CONDUCTED WITH THE TYPE OF EQUIPMENT AND IN ACCORDANCE WITH THE METHODS APPROVED BY THE DEPARTMENT OF FORENSIC SCIENCE; THAT THE TEST WAS CONDUCTED IN ACCORDANCE WITH THE DEPARTMENT'S SPECIFICATIONS; THAT PRIOR TO ADMINISTRATION OF THE TEST THE ACCUSED WAS ADVISED OF HIS RIGHT TO OBSERVE THE PROCESS AND SEE THE BLOOD ALCOHOL READING ON THE EQUIPMENT USED TO PERFORM THE BREATH TEST, AND THAT I POSSESS A VALID LICENSE TO CONDUCT SUCH TEST, GIVEN UNDER MY HAND THIS _____ DAY OF _____, 20_____.

BREATH TEST OPERATOR

I HAVE RECEIVED A COPY OF THIS CERTIFICATE OF ANALYSIS _____ SUBJECT'S SIGNATURE

SUBJECT REFUSED TO SIGN FOR COPY OF CERTIFICATE OF ANALYSIS _____ OPERATOR'S SIGNATURE

mmB



IntoxNet MIS Report

Report Generated 21 May 2014 at 08:52

Test Results

Instrument Serial Number 010621

Test # 000299 **Subject Test**

Test Location 1 Department of
Test Date 20 May 2014

Test Location 2 Forensic Science
Test Time 10:27
Remote/Local Local

Test Location 3
System Check Passed

Operator's Last Name BILLETER
Agency DFS Central Lab

Operator's First Name MICHELLE

Operator's Middle Initial M

Card Serial Number 126417

License Number 26417

Effective Date 10/01/2012

Expiration Date 10/01/2014

Subject's Last Name INSTRUMENT

Subject's First Name TEST

Subject's Middle Initial

Subject's Date of Birth 00/00/0000

Subject's Sex Male

Driver's License Number

Driver's License Expiration 00/00/0000

Driver's License State

Court Name DFS

End Date 20 May 2014 **End Time** 10:35

Result Time 10:33

Result Date 20 May 2014 **Result** 0.00

Data Type DIAG

Sample Value Pass

Sample Time 10:28

Data Type BLK

Sample Value 0.000

Sample Time 10:28

Data Type CHK

Sample Value 0.103

Sample Time 10:29

Data Type BLK

Sample Value 0.000

Sample Time 10:30

Data Type SUBJ

Sample Value 0.000

Sample Time 10:31

Data Type BLK

Sample Value 0.000

Sample Time 10:32

Data Type SUBJ

Sample Value 0.000

Sample Time 10:33

Data Type BLK

Sample Value 0.000

Sample Time 10:34

Standard Type Dry Gas Std

Standard Value 0.101

Standard Lot Number AG320502-29

Standard Expiration Date 07/24/2015

Tank Pressure 1090

Barometric Pressure 764 mmHg

Blow Sample Number 1 **Blow Duration** 4.03 sec

Blow Volume 1691 cc **End-of-Blow Time** 10:31

Blow Sample Number 2 **Blow Duration** 4.01 sec

Blow Volume 1698 cc **End-of-Blow Time** 10:33

Tamper Evident Stamp 8adc32e4

Test Status Code 0

Test Status Success

MmB

Intox EC/IR-II: Accuracy Check

Department of Forensic Science

Serial Number: 010621 Test Number: 300

Test Date: 05/20/2014 Test Time: 10:37 EDT

Dry Gas Target: 0.101

Lot Number: AG320502-29 Exp Date: 07/24/2015

Tank Pressure: 1090 psi Barometric Pressure: 764 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	10:38
CHK	0.102	10:38
BLK	0.000	10:40
CHK	0.102	10:41
BLK	0.000	10:42
CHK	0.102	10:43

Test Status: *Success*

Calibration CRC: 34FC2254

Intox EC/IR-11: Accuracy Check*Department of Forensic Science*

Serial Number: 010621 Test Number: 301

Test Date: 05/20/2014 Test Time: 10:46 EDT

Dry Gas Target: 0.302

Lot Number: AG334603-07 Exp Date: 12/12/2015

Tank Pressure: 541 psi Barometric Pressure: 764 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	10:46
CHK	0.307	10:47
BLK	0.000	10:49
CHK	0.307	10:50
BLK	0.000	10:52
CHK	0.306	10:53

Test Status: *Success*

Calibration CRC: 34FC2254