

**Department of Forensic Science
Breath Alcohol Section
Instrument Maintenance History
1/1/1990 To 6/27/2014**

HMS
mmb

Instrument Serial Number: 010455 **As of 27-Jun-14**

Maintenance Date	Certification Date	Technician Initials	Remarks
24-Jun-14	06-Mar-14	HMS (18910)	REPLACED DRY GAS STANDARD.
06-Mar-14	06-Mar-14	DBH (21250)	CALIBRATED AND CERTIFIED.
18-Sep-13	18-Sep-13	DBH (21250)	FIRMWARE/SOFTWARE UPGRADE. CALIBRATED AND CERTIFIED. REPLACED DRY GAS STANDARD.
24-Jun-13	24-Jun-13	HMS (18910)	CERTIFIED
23-Oct-12	23-Oct-12	DBH (21250)	FIRMWARE/SOFTWARE UPGRADE, CALIBRATED AND CERTIFIED.
13-Sep-12	13-Jun-12	HMS (18910)	REPLACED DRY GAS STANDARD.
13-Jun-12	13-Jun-12	HMS (18910)	CERTIFIED
04-Jun-12	29-Dec-11	HMS (18910)	REPLACED E-SEEK CARD READER.
11-May-12	29-Dec-11	HMS (18910)	REPLACED DRY GAS STANDARD.
29-Dec-11	29-Dec-11	HMS (18910)	CERTIFIED, REPLACED DRY GAS STANDARD.
10-Nov-11	14-Jul-11	HMS (18910)	REPLACED DRY GAS STANDARD.
14-Jul-11	14-Jul-11	HMS (18910)	CERTIFIED
12-Jul-11	21-Mar-11	HMS (18910)	REPLACED DRY GAS STANDARD.
21-Mar-11	21-Mar-11	HMS (18910)	CALIBRATED AND CERTIFIED.
09-Mar-10	09-Mar-10	NMJ (22392)	ADJUSTED AND TESTED SET SOLENOID, CALIBRATED AND CERTIFIED.
26-Oct-09	26-Oct-09	HMS (18910)	REPLACED TANK TOP LID, CALIBRATED AND CERTIFIED.
26-Mar-09	26-Mar-09	HMS (18910)	FIRMWARE/SOFTWARE UPGRADE, CALIBRATED AND CERTIFIED.
04-Nov-08	04-Nov-08	RJP (22428)	CERTIFIED.
22-Oct-08	22-Oct-08	NMJ (22392)	CALIBRATED, CERTIFIED AND ACCEPTED.

Handwritten initials: HMB, mmb

INTOX EC/IR II Quality Assurance Worksheet

Instrument Serial Number: 010455 Worksheet Start Date: 6/24/2014

Location: Fairfax County Adult Detention Center

Address: 10520 Judicial Dr., Fairfax, VA 22030

DFS Technician: Heather Stanton License No.: 18910

Laboratory On-Site

Site Specification: No detrimental environmental conditions exist.

Instrument Barometer (mm HG): 753 Reference Barometer (mm HG): 753

Reference Barometer(RB)Serial #: 009113 RB Calibration Due: 4/28/2015

Measurement Assurance Check


Standard (sea level)	PA Target	minimum	maximum	Sample 1	Sample 2	Sample 3
0.300	0.298	0.289	0.306	0.305	0.304	0.303
Precision		sample min	sample max			
0.002		0.303	0.305			

Standard (sea level)	PA Target	minimum	maximum	Sample 1	Sample 2	Sample 3
0.100	0.099	0.096	0.102	0.101	0.101	0.101
Precision		sample min	sample max			
0		0.101	0.101			


Dry gas standard Lot No. (with tank no.) AG230501-45

- Replaced dry gas standard (+O-ring)
- Installed at Location
- Removed to DFS-Central

Supplies
Mouthpieces
Certificates of Analysis
Operator Worksheet
Other:



Notes:



Instrument Serial Number

010455

Certification Date

Calibrated

Certified

Measurement Assurance Check

Instrument Test

Troubleshooting/Maintenance Worksheet Completed

(Successful completion denotes satisfactory condition of the item.)

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.250		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.150		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.080		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.020		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

All measurements are in g/210L

Estimation of Uncertainty of Measurement and traceability records are located within the Breath Alcohol Section

DFS Technician

Heath W. Starn

Date

6/27/14

Issuing Analyst

Michelle Bellin

Date

6/27/14

MSK

AMS
mmb

Intox EC/IR-II: Accuracy Check

Department of Forensic Science

Serial Number: 010455 Test Number: 1823

Test Date: 06/24/2014 Test Time: 10:00 EDT

Dry Gas Target: 0.298

Lot Number: AG334603-02 Exp Date: 12/12/2015

Tank Pressure: 286 psi Barometric Pressure: 754 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	10:01
CHK	0.305	10:02
BLK	0.000	10:04
CHK	0.304	10:05
BLK	0.000	10:07
CHK	0.303	10:08

Test Status: *Success*

Calibration CRC: A997A77C

AMS
mmB

Intox EC/IR-II: Accuracy Check

Department of Forensic Science

Serial Number: 010455 Test Number: 1824

Test Date: 06/24/2014 Test Time: 10:11 EDT

Dry Gas Target: 0.099

Lot Number: AG230501-45 Exp Date: 10/31/2014

Tank Pressure: 840 psi Barometric Pressure: 754 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	10:12
CHK	0.101	10:13
BLK	0.000	10:14
CHK	0.101	10:15
BLK	0.000	10:17
CHK	0.101	10:17

Test Status: *Success*

Calibration CRC: A997A77C

IMS
mmB



COMMONWEALTH OF VIRGINIA
DEPARTMENT OF FORENSIC SCIENCE

CERTIFICATE OF BLOOD ALCOHOL ANALYSIS
AS DETERMINED BY A CHEMICAL TEST OF THE ACCUSED'S BREATH

NAME OF ACCUSED INSTRUMENT, TEST,	NAME OF COURT DFS
---	-----------------------------

BREATH ANALYSIS

SAMPLE EXAMINED AND TEST CONDUCTED BY STANTON, HEATHER, M		AGENCY DFS Central Lab
DFS LICENSE NUMBER 18910	LICENSE EXPIRES 10/01/2014	DATE TEST CONDUCTED 06/24/2014
TEST EQUIPMENT NUMBER 010455		

RESULTS: TIME SAMPLE TAKEN 10:26 EDT

SAMPLE'S ALCOHOL CONTENT 0.00 GRAMS PER 210 LITERS OF BREATH

ATTEST:

I CERTIFY THAT THE ABOVE IS AN ACCURATE RECORD OF THE TEST CONDUCTED; THAT THE TEST WAS CONDUCTED WITH THE TYPE OF EQUIPMENT AND IN ACCORDANCE WITH THE METHODS APPROVED BY THE DEPARTMENT OF FORENSIC SCIENCE; THAT THE TEST WAS CONDUCTED IN ACCORDANCE WITH THE DEPARTMENT'S SPECIFICATIONS; THAT PRIOR TO ADMINISTRATION OF THE TEST THE ACCUSED WAS ADVISED OF HIS RIGHT TO OBSERVE THE PROCESS AND SEE THE BLOOD ALCOHOL READING ON THE EQUIPMENT USED TO PERFORM THE BREATH TEST, AND THAT I POSSESS A VALID LICENSE TO CONDUCT SUCH TEST, GIVEN UNDER MY HAND THIS _____ DAY OF _____, 20_____.

BREATH TEST OPERATOR

I HAVE RECEIVED A COPY OF THIS CERTIFICATE OF ANALYSIS _____ SUBJECT'S SIGNATURE

SUBJECT REFUSED TO SIGN FOR COPY OF CERTIFICATE OF ANALYSIS _____ OPERATOR'S SIGNATURE



IntoxNet MIS Report
Report Generated 27 Jun 2014 at 08:42

HMS
mmb

Test Results

Instrument Serial Number 010455

Test # 001825 Subject Test

Test Location 1 Department of
Test Date 24 Jun 2014

Test Location 2 Forensic Science
Test Time 10:19
Remote/Local Local

Test Location 3
System Check Passed

Operator's Last Name STANTON
Agency DFS Central Lab

Operator's First Name HEATHER

Operator's Middle Initial M

Card Serial Number 118910

Effective Date 10/01/2012

License Number 18910

Expiration Date 10/01/2014

Subject's Last Name INSTRUMENT

Subject's First Name TEST

Subject's Middle Initial

Subject's Date of Birth 00/00/0000

Subject's Sex Male

Driver's License Number

Driver's License Expiration 00/00/0000

Driver's License State

Court Name DFS

End Date 24 Jun 2014 End Time 10:27

Result Time 10:26

Result Date 24 Jun 2014 Result 0.00

Data Type DIAG

Sample Value Pass

Sample Time 10:19

Data Type BLK

Sample Value 0.000

Sample Time 10:20

Data Type CHK

Sample Value 0.100

Sample Time 10:21

Data Type BLK

Sample Value 0.000

Sample Time 10:22

Data Type SUBJ

Sample Value 0.000

Sample Time 10:23

Data Type BLK

Sample Value 0.000

Sample Time 10:24

Data Type SUBJ

Sample Value 0.000

Sample Time 10:26

Data Type BLK

Sample Value 0.000

Sample Time 10:27

Standard Type Dry Gas Std

Standard Value 0.099

Standard Lot Number AG230501-45

Standard Expiration Date 10/31/2014

Tank Pressure 830

Barometric Pressure 754 mmHg

Blow Sample Number 1 Blow Duration 3.41 sec

Blow Volume 1738 cc End-of-Blow Time 10:23

Blow Sample Number 2 Blow Duration 3.42 sec

Blow Volume 1879 cc End-of-Blow Time 10:26

Tamper Evident Stamp 3f83a57b

Test Status Code 0

Test Status Success