

Department of Forensic Science
Breath Alcohol Section
Instrument Location History
1/1/1990 To 6/2/2014

DBH
mmb

Instrument Serial Number: 010461 As of 02-Jun-14

Start Date	End Date	Agency Name
29-May-14		DFS Central Lab
26-Jul-12	29-May-14	Appomattox County SO
10-Jul-12	26-Jul-12	DFS Central Lab
09-Dec-11	10-Jul-12	Henrico County PD
09-Dec-11	09-Dec-11	DFS Central Lab
17-Aug-10	09-Dec-11	Henrico County PD
25-Mar-09	17-Aug-10	DFS Central Lab
28-Oct-08	25-Mar-09	Russell County SO
08-May-08	28-Oct-08	DFS Central Lab

DBH
mmB

INTOX EC/IR II Quality Assurance Worksheet

Instrument Serial Number **010461** Worksheet Start Date **5/29/2014**
 Location **Appomattox CO SO**
 Address **179 Morton Lane Appomattox VA 24522**
 DFS Technician **Donald Hall** License No. **21250**

Laboratory On-Site

Site Specification: No detrimental environmental conditions exist.

Instrument Barometer (mm HG) **742** Reference Barometer (mm HG) **743**
 Reference Barometer(RB)Serial # **009109** RB Calibration Due **8/1/2014**

Measurement Assurance Check

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.298
0.300	0.293	0.284	0.301	Sample 2	0.297
Precision		sample min	sample max	Sample 3	0.297
0.001		0.297	0.298		

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.099
0.100	0.098	0.095	0.101	Sample 2	0.099
Precision		sample min	sample max	Sample 3	0.099
0		0.099	0.099		

Dry gas standard Lot No. (with tank no.) **AG320502-13**

- Replaced dry gas standard (+O-ring)
- Installed at Location
- Removed to DFS-Central

- Supplies
- Mouthpieces
- Certificates of Analysis
- Operator Worksheet
- Other:



Notes:

[Large shaded area for notes]

Instrument Serial Number

010461

Certification Date

Calibrated

Certified

Measurement Assurance Check

Instrument Test

Troubleshooting/Maintenance Worksheet Completed

(Successful completion denotes satisfactory condition of the item.)

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.250		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.150		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.080		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.020		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

All measurements are in g/210L

Estimation of Uncertainty of Measurement and traceability records are located within the Breath Alcohol Section

DFS Technician

David B. Hall

Date

6/2/14

Issuing Analyst

Michelle Bellis

Date

6/4/14

max

DBH
mmB



COMMONWEALTH OF VIRGINIA
DEPARTMENT OF FORENSIC SCIENCE

CERTIFICATE OF BLOOD ALCOHOL ANALYSIS
AS DETERMINED BY A CHEMICAL TEST OF THE ACCUSED'S BREATH

NAME OF ACCUSED INSTRUMENT, TEST,	NAME OF COURT DFS
---	-----------------------------

BREATH ANALYSIS

SAMPLE EXAMINED AND TEST CONDUCTED BY HALL, DONALD, B		AGENCY DFS Central Lab
DFS LICENSE NUMBER 21250	LICENSE EXPIRES 10/01/2014	DATE TEST CONDUCTED 05/29/2014
TEST EQUIPMENT NUMBER 010461		

RESULTS: TIME SAMPLE TAKEN 11:00 EDT

SAMPLE'S ALCOHOL CONTENT 0.00 GRAMS PER 210 LITERS OF BREATH

ATTEST:

I CERTIFY THAT THE ABOVE IS AN ACCURATE RECORD OF THE TEST CONDUCTED; THAT THE TEST WAS CONDUCTED WITH THE TYPE OF EQUIPMENT AND IN ACCORDANCE WITH THE METHODS APPROVED BY THE DEPARTMENT OF FORENSIC SCIENCE; THAT THE TEST WAS CONDUCTED IN ACCORDANCE WITH THE DEPARTMENT'S SPECIFICATIONS; THAT PRIOR TO ADMINISTRATION OF THE TEST THE ACCUSED WAS ADVISED OF HIS RIGHT TO OBSERVE THE PROCESS AND SEE THE BLOOD ALCOHOL READING ON THE EQUIPMENT USED TO PERFORM THE BREATH TEST, AND THAT I POSSESS A VALID LICENSE TO CONDUCT SUCH TEST, GIVEN UNDER MY HAND THIS _____ DAY OF _____, 20 _____.

BREATH TEST OPERATOR

I HAVE RECEIVED A COPY OF THIS CERTIFICATE OF ANALYSIS _____ SUBJECT'S SIGNATURE

SUBJECT REFUSED TO SIGN FOR COPY OF CERTIFICATE OF ANALYSIS _____ OPERATOR'S SIGNATURE

FOR COURT



IntoxNet MIS Report
Report Generated 29 May 2014 at 11:26

DBA
mmB

Test Results

Instrument Serial Number 010461

Test # 002345 Subject Test

Test Location 1 Department of
Test Date 29 May 2014

Test Location 2 Forensic Science
Test Time 10:54
Remote/Local Local

Test Location 3
System Check Passed

Operator's Last Name HALL
Agency DFS Central Lab

Operator's First Name DONALD

Operator's Middle Initial B

Card Serial Number 121250

License Number 21250

Effective Date 10/01/2012

Expiration Date 10/01/2014

Subject's Last Name INSTRUMENT

Subject's First Name TEST

Subject's Middle Initial

Subject's Date of Birth 00/00/0000

Subject's Sex Male

Driver's License Number

Driver's License Expiration 00/00/0000

Driver's License State

Court Name DFS

End Date 29 May 2014 End Time 11:01

Result Time 11:00

Result Date 29 May 2014 Result 0.00

Data Type DIAG

Sample Value Pass

Sample Time 10:54

Data Type BLK

Sample Value 0.000

Sample Time 10:55

Data Type CHK

Sample Value 0.099

Sample Time 10:56

Data Type BLK

Sample Value 0.000

Sample Time 10:57

Data Type SUBJ

Sample Value 0.000

Sample Time 10:58

Data Type BLK

Sample Value 0.000

Sample Time 10:59

Data Type SUBJ

Sample Value 0.000

Sample Time 11:00

Data Type BLK

Sample Value 0.000

Sample Time 11:01

Standard Type Dry Gas Std

Standard Value 0.098

Standard Lot Number AG320502-13

Standard Expiration Date 07/24/2015

Tank Pressure 1036

Barometric Pressure 742 mmHg

Blow Sample Number 1 Blow Duration 3.62 sec

Blow Volume 1830 cc End-of-Blow Time 10:58

Blow Sample Number 2 Blow Duration 4.08 sec

Blow Volume 2030 cc End-of-Blow Time 11:00

Tamper Evident Stamp 9794c1b3

Test Status Code 0

Test Status Success

DBH
mmB

Intox EC/IR-II: Accuracy Check

Department of Forensic Science

Serial Number: 010461 Test Number: 2346

Test Date: 05/29/2014 Test Time: 11:03 EDT

Dry Gas Target: 0.098

Lot Number: AG320502-13 Exp Date: 07/24/2015

Tank Pressure: 1031 psi Barometric Pressure: 742 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	11:04
CHK	0.099	11:05
BLK	0.000	11:06
CHK	0.099	11:07
BLK	0.000	11:08
CHK	0.099	11:09

Test Status: *Success*

Calibration CRC: 69B6385F

DBH
mmB

Intox EC/IR-II: Accuracy Check

Department of Forensic Science

Serial Number: 010461 Test Number: 2347

Test Date: 05/29/2014 Test Time: 11:12 EDT

Dry Gas Target: 0.293

Lot Number: AG322702-09 Exp Date: 08/15/2015

Tank Pressure: 85 psi Barometric Pressure: 742 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	11:12
CHK	0.298	11:13
BLK	0.000	11:15
CHK	0.297	11:16
BLK	0.000	11:17
CHK	0.297	11:18

Test Status: *Success*

Calibration CRC: 69B6385F