

mms

**Department of Forensic Science
Breath Alcohol Section
Instrument Location History
1/1/1990 To 5/23/2014**

Instrument Serial Number: 010539 **As of 23-May-14**

Start Date	End Date	Agency Name
23-May-14		DFS Central Lab
05-Jan-09	23-May-14	West Point PD
16-Jun-08	05-Jan-09	DFS Central Lab

mmB

INTOX EC/IR II Quality Assurance Worksheet

Instrument Serial Number **010539** Worksheet Start Date **5/23/2014**
 Location **West Point PD**
 Address **433 12th St West Point VA 23181**
 DFS Technician **Michelle Billeter** License No. **26417**

Laboratory On-Site

Site Specification: No detrimental environmental conditions exist.

Instrument Barometer (mm HG) **760** Reference Barometer (mm HG) **761**
 Reference Barometer(RB)Serial # **009111** RB Calibration Due **5/19/2015**

Measurement Assurance Check

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.305
0.300	0.300	0.291	0.309	Sample 2	0.305
Precision		sample min	sample max	Sample 3	0.305
0		0.305	0.305		

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.102
0.100	0.100	0.097	0.103	Sample 2	0.102
Precision		sample min	sample max	Sample 3	0.102
0		0.102	0.102		


Dry gas standard Lot No. (with tank no.) **AG320502-10**

- Replaced dry gas standard (+O-ring)
- Installed at Location
- Removed to DFS-Central

Supplies
 Mouthpieces
 Certificates of Analysis
 Operator Worksheet
 Other:



Notes:



Instrument Serial Number

010539

Certification Date

Calibrated

Certified

Measurement Assurance Check

Instrument Test

Troubleshooting/Maintenance Worksheet Completed

(Successful completion denotes satisfactory condition of the item.)

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.250		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.150		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.080		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.020		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

All measurements are in g/210L

Estimation of Uncertainty of Measurement and traceability records are located within the Breath Alcohol Section

DFS Technician

Michelle Bullock

Date

5/23/14

Issuing Analyst

Michelle Bullock

Date

5/23/14

mbk

mmb



COMMONWEALTH OF VIRGINIA
DEPARTMENT OF FORENSIC SCIENCE

CERTIFICATE OF BLOOD ALCOHOL ANALYSIS
AS DETERMINED BY A CHEMICAL TEST OF THE ACCUSED'S BREATH

NAME OF ACCUSED INSTRUMENT, TEST,	NAME OF COURT DFS
---	-----------------------------

BREATH ANALYSIS

SAMPLE EXAMINED AND TEST CONDUCTED BY BILLETTER, MICHELLE, M		AGENCY DFS Central Lab
DFS LICENSE NUMBER 26417	LICENSE EXPIRES 10/01/2014	DATE TEST CONDUCTED 05/23/2014
TEST EQUIPMENT NUMBER 010539		

RESULTS: TIME SAMPLE TAKEN 11:30 EDT

SAMPLE'S ALCOHOL CONTENT 0.00 GRAMS PER 210 LITERS OF BREATH

ATTEST:

I CERTIFY THAT THE ABOVE IS AN ACCURATE RECORD OF THE TEST CONDUCTED; THAT THE TEST WAS CONDUCTED WITH THE TYPE OF EQUIPMENT AND IN ACCORDANCE WITH THE METHODS APPROVED BY THE DEPARTMENT OF FORENSIC SCIENCE; THAT THE TEST WAS CONDUCTED IN ACCORDANCE WITH THE DEPARTMENT'S SPECIFICATIONS; THAT PRIOR TO ADMINISTRATION OF THE TEST THE ACCUSED WAS ADVISED OF HIS RIGHT TO OBSERVE THE PROCESS AND SEE THE BLOOD ALCOHOL READING ON THE EQUIPMENT USED TO PERFORM THE BREATH TEST, AND THAT I POSSESS A VALID LICENSE TO CONDUCT SUCH TEST, GIVEN UNDER MY HAND THIS _____ DAY OF _____, 20_____.

BREATH TEST OPERATOR

I HAVE RECEIVED A COPY OF THIS CERTIFICATE OF ANALYSIS _____ SUBJECT'S SIGNATURE

SUBJECT REFUSED TO SIGN FOR COPY OF CERTIFICATE OF ANALYSIS _____ OPERATOR'S SIGNATURE

mmB



IntoxNet MIS Report

Report Generated 23 May 2014 at 11:55

Test Results

Instrument Serial Number 010539

Test # 000343 **Subject Test**

Test Location 1 Department of
Test Date 23 May 2014

Test Location 2 Forensic Science
Test Time 11:24
Remote/Local Local

Test Location 3
System Check Passed

Operator's Last Name BILLETER
Agency DFS Central Lab

Operator's First Name MICHELLE

Operator's Middle Initial M

License Number 26417

Card Serial Number 126417

Effective Date 10/01/2012

Expiration Date 10/01/2014

Subject's Last Name INSTRUMENT

Subject's First Name TEST

Subject's Middle Initial

Subject's Date of Birth 00/00/0000

Subject's Sex Male

Driver's License Number

Driver's License Expiration 00/00/0000

Driver's License State

Court Name DFS

End Date 23 May 2014 End Time 11:32

Result Time 11:30

Result Date 23 May 2014 Result 0.00

Data Type DIAG

Sample Value Pass

Sample Time 11:24

Data Type BLK

Sample Value 0.000

Sample Time 11:25

Data Type CHK

Sample Value 0.102

Sample Time 11:26

Data Type BLK

Sample Value 0.000

Sample Time 11:27

Data Type SUBJ

Sample Value 0.000

Sample Time 11:28

Data Type BLK

Sample Value 0.000

Sample Time 11:29

Data Type SUBJ

Sample Value 0.000

Sample Time 11:30

Data Type BLK

Sample Value 0.000

Sample Time 11:31

Standard Type Dry Gas Std

Standard Value 0.100

Standard Lot Number AG320502-10

Standard Expiration Date 07/24/2015

Tank Pressure 835

Barometric Pressure 760 mmHg

Blow Sample Number 1 Blow Duration 3.68 sec

Blow Volume 1716 cc End-of-Blow Time 11:28

Blow Sample Number 2 Blow Duration 3.72 sec

Blow Volume 1665 cc End-of-Blow Time 11:30

Tamper Evident Stamp 65c11f6d

Test Status Code 0

Test Status Success

Intox EC/IR-II: Accuracy Check

MmB

Department of Forensic Science

Serial Number: 010539 Test Number: 344

Test Date: 05/23/2014 Test Time: 11:33 EDT

Dry Gas Target: 0.100

Lot Number: AG320502-10 Exp Date: 07/24/2015

Tank Pressure: 830 psi Barometric Pressure: 760 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	11:34
CHK	0.102	11:34
BLK	0.000	11:36
CHK	0.102	11:37
BLK	0.000	11:38
CHK	0.102	11:39

Test Status: *Success*

Calibration CRC: 03678BCC

Intox EC/IR-II: Accuracy Check*Department of Forensic Science*

Serial Number: 010539 Test Number: 345

Test Date: 05/23/2014 Test Time: 11:42 EDT

Dry Gas Target: 0.300

Lot Number: AG321804-04 Exp Date: 08/06/2015

Tank Pressure: 105 psi Barometric Pressure: 760 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	11:43
CHK	0.305	11:44
BLK	0.000	11:45
CHK	0.305	11:46
BLK	0.000	11:48
CHK	0.305	11:49

Test Status: *Success*

Calibration CRC: 03678BCC