

Department of Forensic Science
Breath Alcohol Section
Instrument Location History
1/1/1990 To 4/2/2014

AMS
TLW
4/22/14

Instrument Serial Number: 010556

As of 02-Apr-14

Start Date	End Date	Agency Name
01-Apr-14		DFS Central Lab
20-Apr-09	01-Apr-14	Vienna PD
23-Jun-08	20-Apr-09	DFS Central Lab

AMS
TLW

INTOX EC/IR II Quality Assurance Worksheet

Instrument Serial Number **010556** Worksheet Start Date **4/1/2014**
 Location **Vienna Police Department**
 Address **215 S. Center St., Vienna, VA 22180**
 DFS Technician **Heather Stanton** License No. **18910**

Laboratory On-Site

Site Specification: No detrimental environmental conditions exist.

Instrument Barometer (mm HG) **756** Reference Barometer (mm HG) **756**
 Reference Barometer(RB)Serial # **009113** RB Calibration Due **5/8/2014**

Measurement Assurance Check

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.302
0.300	0.298	0.289	0.306	Sample 2	0.302
Precision		sample min	sample max	Sample 3	0.302
0		0.302	0.302		

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.101
0.100	0.099	0.096	0.102	Sample 2	0.101
Precision		sample min	sample max	Sample 3	0.101
0		0.101	0.101		


Dry gas standard Lot No. (with tank no.) **AG227501-27**

Replaced dry gas standard (+O-ring)

Installed at Location

Removed to DFS-Central

Supplies
 Mouthpieces
 Certificates of Analysis
 Operator Worksheet
 Other:



Notes:

[Large shaded area for notes]

Instrument Serial Number 010556

Certification Date

Calibrated

Certified

Measurement Assurance Check

Instrument Test

Troubleshooting/Maintenance Worksheet Completed

(Successful completion denotes satisfactory condition of the item.)

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.250		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.150		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.080		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.020		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

All measurements are in g/210L

Estimation of Uncertainty of Measurement and traceability records are located within the Breath Alcohol Section

DFS Technician Heath M Stark Date 4/2/14

Issuing Analyst MSK Date 4/3/14
4/22/14

HMS
TLW



COMMONWEALTH OF VIRGINIA
DEPARTMENT OF FORENSIC SCIENCE

CERTIFICATE OF BLOOD ALCOHOL ANALYSIS
AS DETERMINED BY A CHEMICAL TEST OF THE ACCUSED'S BREATH

NAME OF ACCUSED INSTRUMENT, TEST,	NAME OF COURT DFS
---	-----------------------------

BREATH ANALYSIS

SAMPLE EXAMINED AND TEST CONDUCTED BY STANTON, HEATHER, M		AGENCY DFS Central Lab
DFS LICENSE NUMBER 18910	LICENSE EXPIRES 10/01/2014	DATE TEST CONDUCTED 04/01/2014
TEST EQUIPMENT NUMBER 010556		

RESULTS: TIME SAMPLE TAKEN 11:04 EDT

SAMPLE'S ALCOHOL CONTENT 0.00 GRAMS PER 210 LITERS OF BREATH

ATTEST:

I CERTIFY THAT THE ABOVE IS AN ACCURATE RECORD OF THE TEST CONDUCTED; THAT THE TEST WAS CONDUCTED WITH THE TYPE OF EQUIPMENT AND IN ACCORDANCE WITH THE METHODS APPROVED BY THE DEPARTMENT OF FORENSIC SCIENCE; THAT THE TEST WAS CONDUCTED IN ACCORDANCE WITH THE DEPARTMENT'S SPECIFICATIONS; THAT PRIOR TO ADMINISTRATION OF THE TEST THE ACCUSED WAS ADVISED OF HIS RIGHT TO OBSERVE THE PROCESS AND SEE THE BLOOD ALCOHOL READING ON THE EQUIPMENT USED TO PERFORM THE BREATH TEST, AND THAT I POSSESS A VALID LICENSE TO CONDUCT SUCH TEST, GIVEN UNDER MY HAND THIS _____ DAY OF _____, 20_____.

BREATH TEST OPERATOR

I HAVE RECEIVED A COPY OF THIS CERTIFICATE OF ANALYSIS _____ SUBJECT'S SIGNATURE

SUBJECT REFUSED TO SIGN FOR COPY OF CERTIFICATE OF ANALYSIS _____ OPERATOR'S SIGNATURE



IntoxNet MIS Report
Report Generated 02 Apr 2014 at 08:23

HMS
TLN

Test Results

Instrument Serial Number 010556

Test # 000513 Subject Test

Test Location 1 Department of
Test Date 01 Apr 2014

Test Location 2 Forensic Science
Test Time 10:57
Remote/Local Local

Test Location 3
System Check Passed

Operator's Last Name STANTON
Agency DFS Central Lab

Operator's First Name HEATHER

Operator's Middle Initial M

Card Serial Number 118910

License Number 18910

Effective Date 10/01/2012

Expiration Date 10/01/2014

Subject's Last Name INSTRUMENT

Subject's First Name TEST

Subject's Middle Initial

Subject's Date of Birth 00/00/0000

Subject's Sex Male

Driver's License Number

Driver's License Expiration 00/00/0000

Driver's License State

Court Name DFS

End Date 01 Apr 2014 End Time 11:05

Result Time 11:04

Result Date 01 Apr 2014 Result 0.00

Data Type DIAG

Sample Value Pass

Sample Time 10:58

Data Type BLK

Sample Value 0.000

Sample Time 10:58

Data Type CHK

Sample Value 0.101

Sample Time 10:59

Data Type BLK

Sample Value 0.000

Sample Time 11:01

Data Type SUBJ

Sample Value 0.000

Sample Time 11:01

Data Type BLK

Sample Value 0.000

Sample Time 11:02

Data Type SUBJ

Sample Value 0.000

Sample Time 11:04

Data Type BLK

Sample Value 0.000

Sample Time 11:05

Standard Type Dry Gas Std

Standard Value 0.099

Standard Lot Number AG227501-27

Standard Expiration Date 10/01/2014

Tank Pressure 580

Barometric Pressure 755 mmHg

Blow Sample Number 1 Blow Duration 3.61 sec

Blow Volume 1733 cc End-of-Blow Time 11:01

Blow Sample Number 2 Blow Duration 3.47 sec

Blow Volume 1727 cc End-of-Blow Time 11:04

Tamper Evident Stamp fdebb0f5

Test Status Code 0

Test Status Success

LMS
TLN

Intox EC/IR-II: Accuracy Check

Department of Forensic Science

Serial Number: 010556 Test Number: 514

Test Date: 04/01/2014 Test Time: 11:08 EDT

Dry Gas Target: 0.099

Lot Number: AG227501-27 Exp Date: 10/01/2014

Tank Pressure: 575 psi Barometric Pressure: 755 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	11:09
CHK	0.101	11:09
BLK	0.000	11:11
CHK	0.101	11:12
BLK	0.000	11:13
CHK	0.101	11:14

Test Status: *Success*

Calibration CRC: 81E670F0

HMS
TLN

Intox EC/IR-II: Accuracy Check

Department of Forensic Science

Serial Number: 010556 Test Number: 515

Test Date: 04/01/2014 Test Time: 11:19 EDT

Dry Gas Target: 0.298

Lot Number: AG321804-03 Exp Date: 08/06/2015

Tank Pressure: 306 psi Barometric Pressure: 754 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	11:20
CHK	0.302	11:21
BLK	0.000	11:22
CHK	0.302	11:23
BLK	0.000	11:25
CHK	0.302	11:26

Test Status: *Success*

Calibration CRC: 81E670F0