

MMB

**Department of Forensic Science
Breath Alcohol Section
Instrument Maintenance History
1/1/1990 To 5/30/2014**

Instrument Serial Number: 010599 **As of 30-May-14**

Maintenance Date	Certification Date	Technician Initials	Remarks
29-May-14	06-Mar-14	MMB (26417)	REPLACED KEYBOARD.
06-Mar-14	06-Mar-14	DBH (21250)	CERTIFIED.
15-Jan-14	15-Jan-14	TLN (25387)	CERTIFIED.
30-Sep-13	30-Sep-13	HMS (18910)	FIRMWARE/SOFTWARE UPGRADE, CALIBRATED AND CERTIFIED.
05-Mar-13	05-Mar-13	SED (24178)	CALIBRATED AND CERTIFIED. REPLACED DRY GAS STANDARD.
18-Sep-12	18-Sep-12	DBH (21250)	CERTIFIED.
28-Mar-12	28-Mar-12	DBH (21250)	CERTIFIED. REPLACED DRY GAS STANDARD.
14-Dec-11	14-Dec-11	DBH (21250)	MOVED INSTRUMENT TO NEW LOCATION, CERTIFIED
01-Nov-11	01-Nov-11	TLN (25387)	CERTIFIED. REPLACED DRY GAS STANDARD.
23-May-11	23-May-11	DBH (21250)	CERTIFIED
07-Dec-10	07-Dec-10	DBH (21250)	REPLACED DRY GAS STANDARD, CERTIFIED
23-Jun-10	23-Jun-10	DBH (21250)	CERTIFIED
05-Jan-10	05-Jan-10	DBH (21250)	REPLACED DRY GAS STANDARD, CERTIFIED
28-Jul-09	28-Jul-09	DBH (21250)	CERTIFIED.
16-Apr-09	16-Apr-09	NMJ (22392)	CALIBRATED, CERTIFIED AND ACCEPTED.

mmB

INTOX EC/IR II Quality Assurance Worksheet

Instrument Serial Number **010599** Worksheet Start Date **5/29/2014**

Location **Portsmouth PD**

Address **711 Crawford St Portsmouth VA 23704**

DFS Technician **Michelle Billeter** License No. **26417**

Laboratory On-Site

Site Specification: No detrimental environmental conditions exist.

Instrument Barometer (mm HG) **762** Reference Barometer (mm HG) **765**

Reference Barometer(RB)Serial # **09111** RB Calibration Due **5/19/2015**

Measurement Assurance Check


Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.300		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.100		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Dry gas standard Lot No. (with tank no.) **AG226201-27**

- Replaced dry gas standard (+O-ring)
- Installed at Location
- Removed to DFS-Central

- Supplies
- Mouthpieces
- Certificates of Analysis
- Operator Worksheet
- Other:



Notes:

Replaced keyboard.

Instrument Serial Number

010599

Certification Date

Calibrated

Certified

Measurement Assurance Check

Instrument Test

Troubleshooting/Maintenance Worksheet Completed

(Successful completion denotes satisfactory condition of the item.)

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.250		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.150		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.080		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.020		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

All measurements are in g/210L

Estimation of Uncertainty of Measurement and traceability records are located within the Breath Alcohol Section

DFS Technician

Michelle Bellis

Date

3/29/14

Issuing Analyst

Michelle Bellis

Date

5/30/14

WKE

AMB



COMMONWEALTH OF VIRGINIA
DEPARTMENT OF FORENSIC SCIENCE

CERTIFICATE OF BLOOD ALCOHOL ANALYSIS
AS DETERMINED BY A CHEMICAL TEST OF THE ACCUSED'S BREATH

NAME OF ACCUSED INSTRUMENT, TEST.	NAME OF COURT DFS
--	--------------------------

BREATH ANALYSIS

SAMPLE EXAMINED AND TEST CONDUCTED BY RILLETTER, MICHELLE, M		AGENCY DFS Central Lab
DFS LICENSE NUMBER 26417	LICENSE EXPIRES 10/01/2014	DATE TEST CONDUCTED 05/29/2014
TEST EQUIPMENT NUMBER 010599		

RESULTS: TIME SAMPLE TAKEN 11:04 EDT

SAMPLE'S ALCOHOL CONTENT 0.00 GRAMS PER 210 LITERS OF BREATH

ATTEST:

I CERTIFY THAT THE ABOVE IS AN ACCURATE RECORD OF THE TEST CONDUCTED; THAT THE TEST WAS CONDUCTED WITH THE TYPE OF EQUIPMENT AND IN ACCORDANCE WITH THE METHODS APPROVED BY THE DEPARTMENT OF FORENSIC SCIENCE; THAT THE TEST WAS CONDUCTED IN ACCORDANCE WITH THE DEPARTMENT'S SPECIFICATIONS; THAT PRIOR TO ADMINISTRATION OF THE TEST THE ACCUSED WAS ADVISED OF HIS RIGHT TO OBSERVE THE PROCESS AND SEE THE BLOOD ALCOHOL READING ON THE EQUIPMENT USED TO PERFORM THE BREATH TEST, AND THAT I POSSESS A VALID LICENSE TO CONDUCT SUCH TEST, GIVEN UNDER MY HAND THIS _____ DAY OF _____, 20_____.

BREATH TEST OPERATOR

I HAVE RECEIVED A COPY OF THIS CERTIFICATE OF ANALYSIS _____ SUBJECT'S SIGNATURE

SUBJECT REFUSED TO SIGN FOR COPY OF CERTIFICATE OF ANALYSIS _____ OPERATOR'S SIGNATURE

MMB



IntoxNet MIS Report

Report Generated 30 May 2014 at 09:06

Test Results

Instrument Serial Number 010599

Test # 000401 Subject Test

Test Location 1 Department of
Test Date 29 May 2014

Test Location 2 Forensic Science
Test Time 10:57
Remote/Local Local

Test Location 3
System Check Passed

Operator's Last Name BILLETER
Agency DFS Central Lab

Operator's First Name MICHELLE

Operator's Middle Initial M

Card Serial Number 126417

License Number 26417

Subject's Last Name INSTRUMENT

Effective Date 10/01/2012

Expiration Date 10/01/2014

Subject's Middle Initial

Subject's Date of Birth 00/00/0000

Subject's First Name TEST

Subject's Sex Male

Driver's License Number

Driver's License Expiration 00/00/0000

Driver's License State

Court Name DFS

End Date 29 May 2014 End Time 11:05

Result Time 11:04

Result Date 29 May 2014 Result 0.00

Data Type DIAG

Sample Value Pass

Sample Time 10:57

Data Type BLK

Sample Value 0.000

Sample Time 10:58

Data Type CHK

Sample Value 0.102

Sample Time 10:59

Data Type BLK

Sample Value 0.000

Sample Time 11:00

Data Type SUBJ

Sample Value 0.000

Sample Time 11:01

Data Type BLK

Sample Value 0.000

Sample Time 11:02

Data Type SUBJ

Sample Value 0.000

Sample Time 11:04

Data Type BLK

Sample Value 0.000

Sample Time 11:05

Standard Type Dry Gas Std

Standard Value 0.100

Standard Lot Number AG226201-27

Standard Expiration Date 09/18/2014

Tank Pressure 409

Barometric Pressure 762 mmHg

Blow Sample Number 1 Blow Duration 3.40 sec

Blow Volume 1624 cc End-of-Blow Time 11:01

Blow Sample Number 2 Blow Duration 3.78 sec

Blow Volume 1664 cc End-of-Blow Time 11:04

Tamper Evident Stamp 080f9bc8

Test Status Code 0

Test Status Success

MMB

Department of Forensic Science
Breath Alcohol Section
Trouble Call Log
1/1/1990 To 5/29/2014

Instrument Serial Number: 010599

29-May-14

Date: 03-Aug-09 **Time:** 7:23 AM **Remote:** False **Tech:** DBH (21250) **Location:** Caroline County VSP Area 44
Indication: DISPLAY BLANK / FROZEN / INCORRECT
Problem: DISPLAY FROZEN
Technician Response: CONDUCTED SITE VISIT. CYCLED POWER TO THE INSTRUMENT. PROBLEM RESOLVED.
Resolution: RETURNED TO SERVICE

Date: 27-Feb-14 **Time:** 10:17 AM **Remote:** True **Tech:** MMB (26417) **Location:** Danville PD
Indication: KEYBOARD / CARD READER NOT RESPONDING
Problem: CARD READER NOT RESPONDING
Technician Response: SITE VISIT CONDUCTED. ISSUE RESOLVED.
Resolution: NO FURTHER ACTION REQUIRED

Date: 29-May-14 **Time:** 8:22 AM **Remote:** True **Tech:** MMB (26417) **Location:** Portsmouth PD
Indication: KEYBOARD / CARD READER NOT RESPONDING
Problem: KEYBOARD NOT RESPONDING
Technician Response: SITE VISIT CONDUCTED. ISSUE RESOLVED.
Resolution: NO FURTHER ACTION REQUIRED
