

**Department of Forensic Science
Breath Alcohol Section
Instrument Location History
1/1/1990 To 6/12/2014**

Instrument Serial Number: 010615

As of 12-Jun-14

| Start Date | End Date | Agency Name |
|-------------------|-----------------|--------------------------------|
| 12-Jun-14 | | Charles City County SO |
| 05-Feb-13 | 12-Jun-14 | DFS Central Lab |
| 23-Jan-13 | 05-Feb-13 | New River Valley Regional Jail |
| 09-Aug-12 | 23-Jan-13 | DFS Central Lab |
| 01-Apr-09 | 09-Aug-12 | Carroll County SO |
| 06-Mar-09 | 01-Apr-09 | DFS Central Lab |
| 14-Jan-09 | 06-Mar-09 | Colonial Beach PD |
| 15-Aug-08 | 14-Jan-09 | DFS Central Lab |

TLN

INTOX EC/IR II Quality Assurance Worksheet

Instrument Serial Number **010615** Worksheet Start Date **6/12/2014**
 Location **Charles City Co SO**
 Address **10600 Courthouse Rd Charles City VA 23030**
 DFS Technician **Timmy Neece, Jr** License No. **25387**

Laboratory On-Site

Site Specification: No detrimental environmental conditions exist.

Instrument Barometer (mm HG) **757** Reference Barometer (mm HG) **760**
 Reference Barometer(RB)Serial # **009110** RB Calibration Due **5/28/2015**

Measurement Assurance Check

| Standard (sea level) | PA Target | minimum | maximum | Sample 1 | Sample 2 | Sample 3 |
|----------------------|-----------|------------|------------|----------|----------|----------|
| 0.300 | 0.299 | 0.290 | 0.307 | 0.306 | 0.305 | 0.304 |
| Precision | | sample min | sample max | | | |
| 0.002 | | 0.304 | 0.306 | | | |

| Standard (sea level) | PA Target | minimum | maximum | Sample 1 | Sample 2 | Sample 3 |
|----------------------|-----------|------------|------------|----------|----------|----------|
| 0.100 | 0.100 | 0.097 | 0.103 | 0.101 | 0.101 | 0.101 |
| Precision | | sample min | sample max | | | |
| 0 | | 0.101 | 0.101 | | | |

Dry gas standard Lot No. (with tank no.) **AG227501-10**

- Replaced dry gas standard (+O-ring)
- Installed at Location
- Removed to DFS-Central

Supplies
 Mouthpieces
 Certificates of Analysis
 Operator Worksheet
 Other:



Notes:

[Large shaded area for notes]

Instrument Serial Number

010615

Certification Date

Calibrated

Certified

Measurement Assurance Check

Instrument Test

Troubleshooting/Maintenance Worksheet Completed

(Successful completion denotes satisfactory condition of the item.)

| | | | | |
|----------------------|-----------|------------|------------|----------|
| Standard (sea level) | PA Target | minimum | maximum | Sample 1 |
| 0.250 | | 0.000 | 0.000 | Sample 2 |
| Precision | | sample min | sample max | Sample 3 |
| 0 | | 0.000 | 0.000 | |

| | | | | |
|----------------------|-----------|------------|------------|----------|
| Standard (sea level) | PA Target | minimum | maximum | Sample 1 |
| 0.150 | | 0.000 | 0.000 | Sample 2 |
| Precision | | sample min | sample max | Sample 3 |
| 0 | | 0.000 | 0.000 | |


| | | | | |
|----------------------|-----------|------------|------------|----------|
| Standard (sea level) | PA Target | minimum | maximum | Sample 1 |
| 0.080 | | -0.003 | 0.003 | Sample 2 |
| Precision | | sample min | sample max | Sample 3 |
| 0 | | 0.000 | 0.000 | |

| | | | | |
|----------------------|-----------|------------|------------|----------|
| Standard (sea level) | PA Target | minimum | maximum | Sample 1 |
| 0.020 | | -0.003 | 0.003 | Sample 2 |
| Precision | | sample min | sample max | Sample 3 |
| 0 | | 0.000 | 0.000 | |

All measurements are in g/210L

Estimation of Uncertainty of Measurement and traceability records are located within the Breath Alcohol Section


DFS Technician



Date

6/12/14

Issuing Analyst



Date

6/12/14



Intox EC/IR-II: Accuracy Check

Department of Forensic Science

Serial Number: 010615 Test Number: 706

Test Date: 06/12/2014 Test Time: 09:03 EDT

Dry Gas Target: 0.299

Lot Number: AG334603-07 Exp Date: 12/12/2015

Tank Pressure: 281 psi Barometric Pressure: 757 mmHg

System Check: *Passed*

| Test | g/210L | Time |
|------|--------|-------|
| BLK | 0.000 | 09:04 |
| CHK | 0.306 | 09:05 |
| BLK | 0.000 | 09:06 |
| CHK | 0.305 | 09:07 |
| BLK | 0.000 | 09:08 |
| CHK | 0.304 | 09:09 |

Test Status: *Success*

Calibration CRC: 727F0A58

Intox EC/IR-II: Accuracy Check

Department of Forensic Science

Serial Number: 010615 Test Number: 707

Test Date: 06/12/2014 Test Time: 09:12 EDT

Dry Gas Target: 0.100

Lot Number: AG227501-10 Exp Date: 10/01/2014

Tank Pressure: 257 psi Barometric Pressure: 757 mmHg

System Check: *Passed*

| Test | g/210L | Time |
|------|--------|-------|
| BLK | 0.000 | 09:13 |
| CHK | 0.101 | 09:13 |
| BLK | 0.000 | 09:15 |
| CHK | 0.101 | 09:16 |
| BLK | 0.000 | 09:17 |
| CHK | 0.101 | 09:18 |

Test Status: *Success*

Calibration CRC: 727F0A58

TZV



COMMONWEALTH OF VIRGINIA
DEPARTMENT OF FORENSIC SCIENCE

CERTIFICATE OF BLOOD ALCOHOL ANALYSIS
AS DETERMINED BY A CHEMICAL TEST OF THE ACCUSED'S BREATH

| | |
|---|-----------------------------|
| NAME OF ACCUSED INSTRUMENT, TEST, | NAME OF COURT DFS |
|---|-----------------------------|

BREATH ANALYSIS

| | | |
|--|--------------------------------------|--|
| SAMPLE EXAMINED AND TEST CONDUCTED BY NEECE, JR., TIMMY, L | | AGENCY DFS Central Lab |
| DFS LICENSE NUMBER 25387 | LICENSE EXPIRES 10/01/2014 | DATE TEST CONDUCTED 06/12/2014 |
| TEST EQUIPMENT NUMBER 010615 | | |

RESULTS: TIME SAMPLE TAKEN 09:26 EDT

SAMPLE'S ALCOHOL CONTENT 0.00 GRAMS PER 210 LITERS OF BREATH

ATTEST:

I CERTIFY THAT THE ABOVE IS AN ACCURATE RECORD OF THE TEST CONDUCTED; THAT THE TEST WAS CONDUCTED WITH THE TYPE OF EQUIPMENT AND IN ACCORDANCE WITH THE METHODS APPROVED BY THE DEPARTMENT OF FORENSIC SCIENCE; THAT THE TEST WAS CONDUCTED IN ACCORDANCE WITH THE DEPARTMENT'S SPECIFICATIONS; THAT PRIOR TO ADMINISTRATION OF THE TEST THE ACCUSED WAS ADVISED OF HIS RIGHT TO OBSERVE THE PROCESS AND SEE THE BLOOD ALCOHOL READING ON THE EQUIPMENT USED TO PERFORM THE BREATH TEST, AND THAT I POSSESS A VALID LICENSE TO CONDUCT SUCH TEST, GIVEN UNDER MY HAND THIS _____ DAY OF _____, 20_____.

| |
|----------------------|
| BREATH TEST OPERATOR |
|----------------------|

I HAVE RECEIVED A COPY OF THIS CERTIFICATE OF ANALYSIS _____ SUBJECT'S SIGNATURE

SUBJECT REFUSED TO SIGN FOR COPY OF CERTIFICATE OF ANALYSIS _____ OPERATOR'S SIGNATURE



IntoxNet MIS Report

Report Generated 12 Jun 2014 at 13:44

Test Results

Instrument Serial Number 010615

Test # 000708 Subject Test

Test Location 1 Department of.
Test Date 12 Jun 2014

Test Location 2 Forensic Science
Test Time 09:20
Remote/Local Local

Test Location 3
System Check Passed

Operator's Last Name NEECE,JR.
Agency DFS Central Lab

Operator's First Name TIMMY

Operator's Middle Initial L

Card Serial Number 125387

License Number 25387

Effective Date 10/01/2012

Expiration Date 10/01/2014

Subject's Last Name INSTRUMENT

Subject's First Name TEST

Subject's Middle Initial

Subject's Date of Birth 00/00/0000

Subject's Sex Male

Driver's License Number

Driver's License Expiration 00/00/0000

Driver's License State

Court Name DFS

End Date 12 Jun 2014 End Time 09:27

Result Time 09:26

Result Date 12 Jun 2014 Result 0.00

Data Type DIAG

Sample Value Pass

Sample Time 09:20

Data Type BLK

Sample Value 0.000

Sample Time 09:21

Data Type CHK

Sample Value 0.101

Sample Time 09:22

Data Type BLK

Sample Value 0.000

Sample Time 09:23

Data Type SUBJ

Sample Value 0.000

Sample Time 09:24

Data Type BLK

Sample Value 0.000

Sample Time 09:25

Data Type SUBJ

Sample Value 0.000

Sample Time 09:26

Data Type BLK

Sample Value 0.000

Sample Time 09:27

Standard Type Dry Gas Std

Standard Value 0.100

Standard Lot Number AG227501-10

Standard Expiration Date 10/01/2014

Tank Pressure 247

Barometric Pressure 757 mmHg

Blow Sample Number 1 Blow Duration 3.80 sec

Blow Volume 1775 cc End-of-Blow Time 09:24

Blow Sample Number 2 Blow Duration 3.69 sec

Blow Volume 1878 cc End-of-Blow Time 09:26

Tamper Evident Stamp 46b05e8d

Test Status Code 0

Test Status Success