

Department of Forensic Science  
Breath Alcohol Section  
Instrument Location History  
1/1/1990 To 6/2/2014

DBH  
mms

---

Instrument Serial Number: 010622 As of 02-Jun-14

---

Start Date	End Date	Agency Name
29-May-14		DFS Central Lab
12-Dec-08	29-May-14	Campbell County SO
04-Aug-08	12-Dec-08	DFS Central Lab

DBIT  
mmB

# INTOX EC/IR II Quality Assurance Worksheet

Instrument Serial Number : 010622      Worksheet Start Date : 5/29/2014

Location : Campbell CO SO

Address : 87 Courthouse Lane Rustburg VA 24588

DFS Technician : Donald Hall      License No. : 21250

Laboratory       On-Site

Site Specification: No detrimental environmental conditions exist.

Instrument Barometer (mm HG) : 740      Reference Barometer (mm HG) : 742

Reference Barometer(RB)Serial # : 009109      RB Calibration Due : 8/1/2014

### Measurement Assurance Check

<b>Standard (sea level)</b>	<b>PA Target</b>	<b>minimum</b>	<b>maximum</b>	<b>Sample 1</b>	0.297
0.300	0.292	0.283	0.300	<b>Sample 2</b>	0.297
<b>Precision</b>		<b>sample min</b>	<b>sample max</b>	<b>Sample 3</b>	0.297
0		0.297	0.297		

<b>Standard (sea level)</b>	<b>PA Target</b>	<b>minimum</b>	<b>maximum</b>	<b>Sample 1</b>	0.099
0.100	0.097	0.094	0.100	<b>Sample 2</b>	0.099
<b>Precision</b>		<b>sample min</b>	<b>sample max</b>	<b>Sample 3</b>	0.099
0		0.099	0.099		

Dry gas standard Lot No. (with tank no.) : AG320502-15

Replaced dry gas standard (+O-ring)

Installed at Location

Removed to DFS-Central

Supplies  
Mouthpieces  
Certificates of Analysis  
Operator Worksheet  
Other:



Notes:

[Large shaded area for notes]

Instrument Serial Number

010622

Certification Date

Calibrated

Certified

Measurement Assurance Check

Instrument Test

Troubleshooting/Maintenance Worksheet Completed

(Successful completion denotes satisfactory condition of the item.)

<b>Standard (sea level)</b>	<b>PA Target</b>	<b>minimum</b>	<b>maximum</b>	<b>Sample 1</b>
0.250		0.000	0.000	<b>Sample 2</b>
<b>Precision</b>		<b>sample min</b>	<b>sample max</b>	<b>Sample 3</b>
0		0.000	0.000	

<b>Standard (sea level)</b>	<b>PA Target</b>	<b>minimum</b>	<b>maximum</b>	<b>Sample 1</b>
0.150		0.000	0.000	<b>Sample 2</b>
<b>Precision</b>		<b>sample min</b>	<b>sample max</b>	<b>Sample 3</b>
0		0.000	0.000	

<b>Standard (sea level)</b>	<b>PA Target</b>	<b>minimum</b>	<b>maximum</b>	<b>Sample 1</b>
0.080		-0.003	0.003	<b>Sample 2</b>
<b>Precision</b>		<b>sample min</b>	<b>sample max</b>	<b>Sample 3</b>
0		0.000	0.000	

<b>Standard (sea level)</b>	<b>PA Target</b>	<b>minimum</b>	<b>maximum</b>	<b>Sample 1</b>
0.020		-0.003	0.003	<b>Sample 2</b>
<b>Precision</b>		<b>sample min</b>	<b>sample max</b>	<b>Sample 3</b>
0		0.000	0.000	

All measurements are in g/210L

Estimation of Uncertainty of Measurement and traceability records are located within the Breath Alcohol Section

DFS Technician

*Donald B Hall*

Date

*6/2/14*

Issuing Analyst

*Michelle Buller*

Date

*6/2/14*

*note*

DBH  
MMB



COMMONWEALTH OF VIRGINIA  
DEPARTMENT OF FORENSIC SCIENCE

CERTIFICATE OF BLOOD ALCOHOL ANALYSIS  
AS DETERMINED BY A CHEMICAL TEST OF THE ACCUSED'S BREATH

NAME OF ACCUSED <b>INSTRUMENT, TEST,</b>	NAME OF COURT <b>DFS</b>
---	-----------------------------

**BREATH ANALYSIS**

SAMPLE EXAMINED AND TEST CONDUCTED BY <b>HALL, DONALD, B</b>		AGENCY <b>DFS Central Lab</b>
DFS LICENSE NUMBER <b>21250</b>	LICENSE EXPIRES <b>10/01/2014</b>	DATE TEST CONDUCTED <b>05/29/2014</b>
TEST EQUIPMENT NUMBER <b>010622</b>		

RESULTS: TIME SAMPLE TAKEN 13:11 EDT

SAMPLE'S ALCOHOL CONTENT 0.00 GRAMS PER 210 LITERS OF BREATH

**ATTEST:**

I CERTIFY THAT THE ABOVE IS AN ACCURATE RECORD OF THE TEST CONDUCTED; THAT THE TEST WAS CONDUCTED WITH THE TYPE OF EQUIPMENT AND IN ACCORDANCE WITH THE METHODS APPROVED BY THE DEPARTMENT OF FORENSIC SCIENCE; THAT THE TEST WAS CONDUCTED IN ACCORDANCE WITH THE DEPARTMENT'S SPECIFICATIONS; THAT PRIOR TO ADMINISTRATION OF THE TEST THE ACCUSED WAS ADVISED OF HIS RIGHT TO OBSERVE THE PROCESS AND SEE THE BLOOD ALCOHOL READING ON THE EQUIPMENT USED TO PERFORM THE BREATH TEST, AND THAT I POSSESS A VALID LICENSE TO CONDUCT SUCH TEST, GIVEN UNDER MY HAND THIS \_\_\_\_\_ DAY OF \_\_\_\_\_, 20 \_\_\_\_\_.

BREATH TEST OPERATOR
----------------------

I HAVE RECEIVED A COPY OF THIS CERTIFICATE OF ANALYSIS \_\_\_\_\_ SUBJECT'S SIGNATURE

SUBJECT REFUSED TO SIGN FOR COPY OF CERTIFICATE OF ANALYSIS \_\_\_\_\_ OPERATOR'S SIGNATURE



# IntoxNet MIS Report

Report Generated 02 Jun 2014 at 10:54

DSH  
mmB

## Test Results

**Instrument Serial Number 010622**

**Test # 001179**

**Subject Test**

Test Location 1 Department of  
Test Date 29 May 2014

Test Location 2 Forensic Science  
Test Time 13:05  
Remote/Local Local

Test Location 3  
System Check Passed

Operator's Last Name HALL  
Agency DFS Central Lab

Operator's First Name DONALD

Operator's Middle Initial B

Card Serial Number 121250

License Number 21250

Effective Date 10/01/2012

Expiration Date 10/01/2014

Subject's Last Name INSTRUMENT

Subject's First Name TEST

Subject's Middle Initial

Subject's Date of Birth 00/00/0000

Subject's Sex Male

Driver's License Number

Driver's License Expiration 00/00/0000

Driver's License State

Court Name DFS

End Date 29 May 2014 End Time 13:12

Result Time 13:11

Result Date 29 May 2014 Result 0.00

Data Type DIAG

Sample Value Pass

Sample Time 13:05

Data Type BLK

Sample Value 0.000

Sample Time 13:06

Data Type CHK

Sample Value 0.099

Sample Time 13:07

Data Type BLK

Sample Value 0.000

Sample Time 13:08

Data Type SUBJ

Sample Value 0.000

Sample Time 13:09

Data Type BLK

Sample Value 0.000

Sample Time 13:10

Data Type SUBJ

Sample Value 0.000

Sample Time 13:11

Data Type BLK

Sample Value 0.000

Sample Time 13:12

Standard Type Dry Gas Std

Standard Value 0.097

Standard Lot Number AG320502-15

Standard Expiration Date 07/24/2015

Tank Pressure 772

Barometric Pressure 740 mmHg

Blow Sample Number 1 Blow Duration 4.03 sec

Blow Volume 2064 cc End-of-Blow Time 13:09

Blow Sample Number 2 Blow Duration 4.23 sec

Blow Volume 2042 cc End-of-Blow Time 13:11

Tamper Evident Stamp cc3b697b

Test Status Code 0

Test Status Success

DBH  
MMB

**Intox EC/IR-II: Accuracy Check**

*Department of Forensic Science*

Serial Number: 010622      Test Number: 1180

Test Date: 05/29/2014      Test Time: 13:14 EDT

Dry Gas Target: 0.097

Lot Number: AG320502-15      Exp Date: 07/24/2015

Tank Pressure: 767 psi      Barometric Pressure: 740 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	13:14
CHK	0.099	13:15
BLK	0.000	13:17
CHK	0.099	13:18
BLK	0.000	13:19
CHK	0.099	13:20

Test Status: *Success*

Calibration CRC: BF7A46B5

D611  
mmB

**Intox EC/1R-11: Accuracy Check**

*Department of Forensic Science*

Serial Number: 010622      Test Number: 1181

Test Date: 05/29/2014      Test Time: 13:22 EDT

Dry Gas Target: 0.292

Lot Number: AG322702-09      Exp Date: 08/15/2015

Tank Pressure: 66 psi      Barometric Pressure: 740 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	13:23
CHK	0.297	13:24
BLK	0.000	13:26
CHK	0.297	13:27
BLK	0.000	13:28
CHK	0.297	13:29

Test Status: *Success*

Calibration CRC: BF7A46B5