

DBH
TW

INTOX EC/IR II Quality Assurance Worksheet

Instrument Serial Number **010514** Worksheet Start Date **6/30/2014**
 Location **Albemarle CO Regional Jail**
 Address **160 Peregory Ln Charlottesville VA 22902**
 DFS Technician **Donald Hall** License No. **21250**

Laboratory On-Site

Site Specification: No detrimental environmental conditions exist.

Instrument Barometer (mm HG) **747** Reference Barometer (mm HG) **748**
 Reference Barometer(RB)Serial # **009109** RB Calibration Due **6/30/2014**

Measurement Assurance Check

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.300		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.100		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Dry gas standard Lot No. (with tank no.) **AG230501-35**

- Replaced dry gas standard (+O-ring)
- Installed at Location
- Removed to DFS-Central

Supplies
 Mouthpieces
 Certificates of Analysis
 Operator Worksheet
 Other:



Notes:

[Large shaded area for notes]

Instrument Serial Number

010514

Certification Date

Calibrated

Certified

Measurement Assurance Check

Instrument Test

Troubleshooting/Maintenance Worksheet Completed

(Successful completion denotes satisfactory condition of the item.)

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.250		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.150		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.080		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.020		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

All measurements are in g/210L

Estimation of Uncertainty of Measurement and traceability records are located within the Breath Alcohol Section

DFS Technician

Donald B Hall

Date

7/14/14

Issuing Analyst

J-VZ

Date

7/14/14

ML

DBH
Tlw



COMMONWEALTH OF VIRGINIA
DEPARTMENT OF FORENSIC SCIENCE

CERTIFICATE OF BLOOD ALCOHOL ANALYSIS
AS DETERMINED BY A CHEMICAL TEST OF THE ACCUSED'S BREATH

NAME OF ACCUSED INSTRUMENT, TEST,	NAME OF COURT DFS
---	-----------------------------

BREATH ANALYSIS

SAMPLE EXAMINED AND TEST CONDUCTED BY HALL, DONALD, B		AGENCY DFS Central Lab
DFS LICENSE NUMBER 21250	LICENSE EXPIRES 10/01/2014	DATE TEST CONDUCTED 06/30/2014
TEST EQUIPMENT NUMBER 010514		

RESULTS: TIME SAMPLE TAKEN 14:24 EDT

SAMPLE'S ALCOHOL CONTENT 0.00 GRAMS PER 210 LITERS OF BREATH

ATTEST:

I CERTIFY THAT THE ABOVE IS AN ACCURATE RECORD OF THE TEST CONDUCTED; THAT THE TEST WAS CONDUCTED WITH THE TYPE OF EQUIPMENT AND IN ACCORDANCE WITH THE METHODS APPROVED BY THE DEPARTMENT OF FORENSIC SCIENCE; THAT THE TEST WAS CONDUCTED IN ACCORDANCE WITH THE DEPARTMENT'S SPECIFICATIONS; THAT PRIOR TO ADMINISTRATION OF THE TEST THE ACCUSED WAS ADVISED OF HIS RIGHT TO OBSERVE THE PROCESS AND SEE THE BLOOD ALCOHOL READING ON THE EQUIPMENT USED TO PERFORM THE BREATH TEST, AND THAT I POSSESS A VALID LICENSE TO CONDUCT SUCH TEST, GIVEN UNDER MY HAND THIS _____ DAY OF _____, 20_____.

BREATH TEST OPERATOR

I HAVE RECEIVED A COPY OF THIS CERTIFICATE OF ANALYSIS _____ SUBJECT'S SIGNATURE

SUBJECT REFUSED TO SIGN FOR COPY OF CERTIFICATE OF ANALYSIS _____ OPERATOR'S SIGNATURE



IntoxNet MIS Report
Report Generated 14 Jul 2014 at 09:32

DBH
TLW

Test Results

Instrument Serial Number 010514

Test # 000288 Subject Test

Test Location 1 Department of
Test Date 30 Jun 2014

Test Location 2 Forensic Science
Test Time 14:18
Remote/Local Local

Test Location 3
System Check Passed

Operator's Last Name HALL
Agency DFS Central Lab

Operator's First Name DONALD

Operator's Middle Initial B

Card Serial Number 121250

License Number 21250

Effective Date 10/01/2012

Expiration Date 10/01/2014

Subject's Last Name INSTRUMENT

Subject's First Name TEST

Subject's Middle Initial

Subject's Date of Birth 00/00/0000

Subject's Sex Male

Driver's License Number

Driver's License Expiration 00/00/0000

Driver's License State

Court Name DFS

End Date 30 Jun 2014 End Time 14:25

Result Time 14:24

Result Date 30 Jun 2014 Result 0.00

Data Type DIAG

Sample Value Pass

Sample Time 14:18

Data Type BLK

Sample Value 0.000

Sample Time 14:19

Data Type CHK

Sample Value 0.100

Sample Time 14:20

Data Type BLK

Sample Value 0.000

Sample Time 14:21

Data Type SUBJ

Sample Value 0.000

Sample Time 14:22

Data Type BLK

Sample Value 0.000

Sample Time 14:23

Data Type SUBJ

Sample Value 0.000

Sample Time 14:24

Data Type BLK

Sample Value 0.000

Sample Time 14:25

Standard Type Dry Gas Std

Standard Value 0.098

Standard Lot Number AG230501-35

Standard Expiration Date 10/31/2014

Tank Pressure 610

Barometric Pressure 747 mmHg

Blow Sample Number 1 Blow Duration 3.66 sec

Blow Volume 1960 cc End-of-Blow Time 14:22

Blow Sample Number 2 Blow Duration 3.55 sec

Blow Volume 1821 cc End-of-Blow Time 14:24

Tamper Evident Stamp ad528c98

Test Status Code 0

Test Status Success