

Department of Forensic Science
Breath Alcohol Section
Instrument Location History
1/1/1990 To 8/13/2014

~~AMS~~
TLN

Instrument Serial Number: 010541 As of 13-Aug-14

Start Date	End Date	Agency Name
12-Aug-14		DFS Central Lab
06-Mar-14	12-Aug-14	Arlington County PD
27-Feb-13	06-Mar-14	DFS Central Lab
21-Oct-08	27-Feb-13	Scott County Sheriff's Office
16-Jun-08	21-Oct-08	DFS Central Lab

LMS
TLW

INTOX EC/IR II Quality Assurance Worksheet

Instrument Serial Number: 010541 Worksheet Start Date: 8/12/2014

Location: Arlington Police Department

Address: 1425 N. Courthouse Rd., Arlington, VA 22201

DFS Technician: Heather Stanton License No.: 18910

Laboratory On-Site

Site Specification: No detrimental environmental conditions exist.

Instrument Barometer (mm HG): 750 Reference Barometer (mm HG): 753

Reference Barometer(RB)Serial #: 009113 RB Calibration Due: 4/28/2015

Measurement Assurance Check

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.300
0.300	0.296	0.287	0.304	Sample 2	0.300
Precision		sample min	sample max	Sample 3	0.300
0		0.300	0.300		

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.100
0.100	0.099	0.096	0.102	Sample 2	0.100
Precision		sample min	sample max	Sample 3	0.099
0.001		0.099	0.100		

Dry gas standard Lot No. (with tank no.): AG332301-14

- Replaced dry gas standard (+O-ring)
- Installed at Location
- Removed to DFS-Central

Supplies
Mouthpieces
Certificates of Analysis
Operator Worksheet
Other:



Notes:

[Redacted area]

Instrument Serial Number

010541

Certification Date

Calibrated

Certified

Measurement Assurance Check

Instrument Test

Troubleshooting/Maintenance Worksheet Completed

(Successful completion denotes satisfactory condition of the item.)

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.250		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.150		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.080		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.020		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

All measurements are in g/210L

Estimation of Uncertainty of Measurement and traceability records are located within the Breath Alcohol Section

DFS Technician Heather Stark Date 8/13/14

Issuing Analyst ASK Date 8/13/14

HMS
TLN



COMMONWEALTH OF VIRGINIA
DEPARTMENT OF FORENSIC SCIENCE

CERTIFICATE OF BLOOD ALCOHOL ANALYSIS
AS DETERMINED BY A CHEMICAL TEST OF THE ACCUSED'S BREATH

NAME OF ACCUSED INSTRUMENT, TEST,	NAME OF COURT DFS
--	--------------------------

BREATH ANALYSIS

SAMPLE EXAMINED AND TEST CONDUCTED BY STANTON, HEATHER, M		AGENCY DFS Central Lab
DFS LICENSE NUMBER 18910	LICENSE EXPIRES 10/01/2014	DATE TEST CONDUCTED 08/12/2014
TEST EQUIPMENT NUMBER 010541	[REDACTED]	

RESULTS: TIME SAMPLE TAKEN 10:49 EDT

SAMPLE'S ALCOHOL CONTENT 0.00 GRAMS PER 210 LITERS OF BREATH

ATTEST:

I CERTIFY THAT THE ABOVE IS AN ACCURATE RECORD OF THE TEST CONDUCTED; THAT THE TEST WAS CONDUCTED WITH THE TYPE OF EQUIPMENT AND IN ACCORDANCE WITH THE METHODS APPROVED BY THE DEPARTMENT OF FORENSIC SCIENCE; THAT THE TEST WAS CONDUCTED IN ACCORDANCE WITH THE DEPARTMENT'S SPECIFICATIONS; THAT PRIOR TO ADMINISTRATION OF THE TEST THE ACCUSED WAS ADVISED OF HIS RIGHT TO OBSERVE THE PROCESS AND SEE THE BLOOD ALCOHOL READING ON THE EQUIPMENT USED TO PERFORM THE BREATH TEST, AND THAT I POSSESS A VALID LICENSE TO CONDUCT SUCH TEST, GIVEN UNDER MY HAND THIS _____ DAY OF _____, 20_____.

BREATH TEST OPERATOR

I HAVE RECEIVED A COPY OF THIS CERTIFICATE OF ANALYSIS _____ SUBJECT'S SIGNATURE

SUBJECT REFUSED TO SIGN FOR COPY OF CERTIFICATE OF ANALYSIS _____ OPERATOR'S SIGNATURE



IntoxNet MIS Report
Report Generated 12 Aug 2014 at 11:18

MMS
JMS
TLN

Test Results

Instrument Serial Number 010541

Test # 000692 Subject Test

Test Location 1 Department of
Test Date 12 Aug 2014

Test Location 2 Forensic Science
Test Time 10:43
Remote/Local Local

Test Location 3
System Check Passed

Operator's Last Name STANTON
Agency DFS Central Lab

Operator's First Name HEATHER

Operator's Middle Initial M

Card Serial Number 118910

Effective Date 10/01/2012

Expiration Date 10/01/2014

Subject's Last Name INSTRUMENT

Subject's First Name TEST

Subject's Middle Initial

Subject's Date of Birth 00/00/0000

Subject's Sex Male

Driver's License Number

Driver's License Expiration 00/00/0000

Driver's License State

Court Name DFS

End Date 12 Aug 2014 End Time 10:50

Result Time 10:49

Result Date 12 Aug 2014 Result 0.00

Data Type DIAG

Sample Value Pass

Sample Time 10:43

Data Type BLK

Sample Value 0.000

Sample Time 10:44

Data Type CHK

Sample Value 0.100

Sample Time 10:45

Data Type BLK

Sample Value 0.000

Sample Time 10:46

Data Type SUBJ

Sample Value 0.000

Sample Time 10:47

Data Type BLK

Sample Value 0.000

Sample Time 10:48

Data Type SUBJ

Sample Value 0.000

Sample Time 10:49

Data Type BLK

Sample Value 0.000

Sample Time 10:50

Standard Type Dry Gas Std

Standard Value 0.099

Standard Lot Number AG332301-14

Standard Expiration Date 11/19/2015

Tank Pressure 801

Barometric Pressure 750 mmHg

Blow Sample Number 1 Blow Duration 3.59 sec

Blow Volume 1970 cc End-of-Blow Time 10:47

Blow Sample Number 2 Blow Duration 3.43 sec

Blow Volume 1756 cc End-of-Blow Time 10:49

Tamper Evident Stamp 3226c4bf

Test Status Code 0

Test Status Success

KMS
TLN

Intox EC/IR-II: Accuracy Check

Department of Forensic Science

Serial Number: 010541 Test Number: 693

Test Date: 08/12/2014 Test Time: 10:51 EDT

Dry Gas Target: 0.099

Lot Number: AG332301-14 Exp Date: 11/19/2015

Tank Pressure: 796 psi Barometric Pressure: 750 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	10:52
CHK	0.100	10:53
BLK	0.000	10:54
CHK	0.100	10:55
BLK	0.000	10:56
CHK	0.099	10:57

Test Status: *Success*

Calibration CRC: CC7D2102

HLMS
TLN

Intox EC/IR-II: Accuracy Check

Department of Forensic Science

Serial Number: 010541 Test Number: 694

Test Date: 08/12/2014 Test Time: 11:01 EDT

Dry Gas Target: 0.296

Lot Number: AG334603-05 Exp Date: 12/12/2015

Tank Pressure: 100 psi Barometric Pressure: 750 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	11:02
CHK	0.300	11:03
BLK	0.000	11:04
CHK	0.300	11:05
BLK	0.000	11:07
CHK	0.300	11:08

Test Status: *Success*

Calibration CRC: CC7D2102