

TLW

Department of Forensic Science
Breath Alcohol Section
Instrument Location History
1/1/1990 To 9/18/2014

Instrument Serial Number: 010465 **As of 18-Sep-14**

Start Date	End Date	Agency Name
18-Sep-14		Meherrin River Regional Jail
11-Sep-14	18-Sep-14	DFS Central Lab
08-Apr-14	11-Sep-14	Galax PD
18-Mar-14	08-Apr-14	DFS Central Lab
24-May-12	18-Mar-14	Virginia Beach PD
24-May-12	24-May-12	DFS Central Lab
23-Feb-10	24-May-12	Virginia Beach PD
23-Sep-09	23-Feb-10	DFS Central Lab
18-Nov-08	23-Sep-09	Charlottesville PD
13-May-08	18-Nov-08	DFS Central Lab

TW

INTOX EC/IR II Quality Assurance Worksheet

Instrument Serial Number **010465** Worksheet Start Date **9/18/2014**
 Location **Meherrin River-Satellite Jail - Mecklenburg Satellite**
Regional
TW 9/19/14
 Address **600 Prison Road Boydton, VA 23917**
 DFS Technician **Timmy Neece, Jr** License No. **25387**

Laboratory On-Site

Site Specification: No detrimental environmental conditions exist.

Instrument Barometer (mm HG) **753** Reference Barometer (mm HG) **754**
 Reference Barometer(RB)Serial # **009111** RB Calibration Due **5/19/2015**

Measurement Assurance Check

Standard (sea level)	PA Target	minimum	maximum	Sample 1	Sample 2	Sample 3
0.300	0.297	0.288	0.305	0.303	0.302	0.301
Precision		sample min	sample max			
0.002		0.301	0.303			

Standard (sea level)	PA Target	minimum	maximum	Sample 1	Sample 2	Sample 3
0.100	0.099	0.096	0.102	0.100	0.100	0.100
Precision		sample min	sample max			
0		0.100	0.100			

Dry gas standard Lot No. (with tank no.) **AG320502-02**

- Replaced dry gas standard (+O-ring)
- Installed at Location
- Removed to DFS-Central

Supplies
 Mouthpieces
 Certificates of Analysis
 Operator Worksheet
 Other:

50

Notes:

[Redacted Notes Area]

Instrument Serial Number

010465

Certification Date

Calibrated

Certified

Measurement Assurance Check

Instrument Test

Troubleshooting/Maintenance Worksheet Completed

(Successful completion denotes satisfactory condition of the item.)

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.250		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.150		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.080		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.020		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

All measurements are in g/210L

Estimation of Uncertainty of Measurement and traceability records are located within the Breath Alcohol Section

DFS Technician

[Handwritten Signature]

Date

9/19/14

Issuing Analyst

[Handwritten Signature]

Date

9/19/14

MSK

TLN

Intox EC/IR-II: Accuracy Check

Department of Forensic Science

Serial Number: 010465 Test Number: 2686

Test Date: 09/18/2014 Test Time: 10:36 EDT

Dry Gas Target: 0.297

Lot Number: AG421201-04 Exp Date: 07/31/2016

Tank Pressure: 438 psi Barometric Pressure: 753 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	10:37
CHK	0.303	10:37
BLK	0.000	10:39
CHK	0.302	10:40
BLK	0.000	10:42
CHK	0.301	10:43

Test Status: *Success*

Calibration CRC: FC317593

TLN

Intox EC/IR-II: Accuracy Check

Department of Forensic Science

Serial Number: 010465 Test Number: 2687

Test Date: 09/18/2014 Test Time: 10:46 EDT

Dry Gas Target: 0.099

Lot Number: AG320502-02 Exp Date: 07/24/2015

Tank Pressure: 625 psi Barometric Pressure: 754 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	10:47
CHK	0.100	10:47
BLK	0.000	10:49
CHK	0.100	10:50
BLK	0.000	10:51
CHK	0.100	10:52

Test Status: *Success*

Calibration CRC: FC317593

TLN



COMMONWEALTH OF VIRGINIA
DEPARTMENT OF FORENSIC SCIENCE

CERTIFICATE OF BLOOD ALCOHOL ANALYSIS
AS DETERMINED BY A CHEMICAL TEST OF THE ACCUSED'S BREATH

NAME OF ACCUSED INSTRUMENT, TEST,	NAME OF COURT DFS
---	-----------------------------

BREATH ANALYSIS

SAMPLE EXAMINED AND TEST CONDUCTED BY NEECE, JR., TIMMY, L.		AGENCY DFS Central Lab
DFS LICENSE NUMBER 25387	LICENSE EXPIRES 10/01/2014	DATE TEST CONDUCTED 09/18/2014
TEST EQUIPMENT NUMBER 010465		

RESULTS: TIME SAMPLE TAKEN 11:00 EDT

SAMPLE'S ALCOHOL CONTENT 0.00 GRAMS PER 210 LITERS OF BREATH

ATTEST:

I CERTIFY THAT THE ABOVE IS AN ACCURATE RECORD OF THE TEST CONDUCTED; THAT THE TEST WAS CONDUCTED WITH THE TYPE OF EQUIPMENT AND IN ACCORDANCE WITH THE METHODS APPROVED BY THE DEPARTMENT OF FORENSIC SCIENCE; THAT THE TEST WAS CONDUCTED IN ACCORDANCE WITH THE DEPARTMENT'S SPECIFICATIONS; THAT PRIOR TO ADMINISTRATION OF THE TEST THE ACCUSED WAS ADVISED OF HIS RIGHT TO OBSERVE THE PROCESS AND SEE THE BLOOD ALCOHOL READING ON THE EQUIPMENT USED TO PERFORM THE BREATH TEST, AND THAT I POSSESS A VALID LICENSE TO CONDUCT SUCH TEST, GIVEN UNDER MY HAND THIS _____ DAY OF _____, 20_____.

BREATH TEST OPERATOR

I HAVE RECEIVED A COPY OF THIS CERTIFICATE OF ANALYSIS _____ SUBJECT'S SIGNATURE

SUBJECT REFUSED TO SIGN FOR COPY OF CERTIFICATE OF ANALYSIS _____ OPERATOR'S SIGNATURE

FOR COURT

TLN



IntoxNet MIS Report

Report Generated 18 Sep 2014 at 15:31

Test Results

Instrument Serial Number 010465

Test # 002688 Subject Test

Test Location 1 Department of
Test Date 18 Sep 2014

Test Location 2 Forensic Science
Test Time 10:54
Remote/Local Local

Test Location 3
System Check Passed

Operator's Last Name NEECE, JR.
Agency DFS Central Lab

Operator's First Name TIMMY

Operator's Middle Initial L

Card Serial Number 125387

Effective Date 10/01/2012

License Number 25387

Expiration Date 10/01/2014

Subject's Last Name INSTRUMENT

Subject's First Name TEST

Subject's Middle Initial

Subject's Date of Birth 00/00/0000

Subject's Sex Male

Driver's License Number

Driver's License Expiration 00/00/0000

Driver's License State

Court Name DFS

End Date 18 Sep 2014 End Time 11:02

Result Time 11:00

Result Date 18 Sep 2014 Result 0.00

Data Type DIAG

Sample Value Pass

Sample Time 10:54

Data Type BLK

Sample Value 0.000

Sample Time 10:55

Data Type CHK

Sample Value 0.100

Sample Time 10:56

Data Type BLK

Sample Value 0.000

Sample Time 10:57

Data Type SUBJ

Sample Value 0.000

Sample Time 10:58

Data Type BLK

Sample Value 0.000

Sample Time 10:59

Data Type SUBJ

Sample Value 0.000

Sample Time 11:00

Data Type BLK

Sample Value 0.000

Sample Time 11:01

Standard Type Dry Gas Std

Standard Value 0.099

Standard Lot Number AG320502-02

Standard Expiration Date 07/24/2015

Tank Pressure 615

Barometric Pressure 754 mmHg

Blow Sample Number 1 Blow Duration 3.59 sec

Blow Volume 1636 cc End-of-Blow Time 10:58

Blow Sample Number 2 Blow Duration 4.16 sec

Blow Volume 2135 cc End-of-Blow Time 11:00

Tamper Evident Stamp 7ebc0e86

Test Status Code 0

Test Status Success