

Department of Forensic Science  
Breath Alcohol Section  
Instrument Location History  
1/1/1990 To 9/8/2014

*LMS*  
*MMB*

Instrument Serial Number: 010514

As of 08-Sep-14

Start Date	End Date	Agency Name
08-Sep-14		DFS Central Lab
30-Apr-14	08-Sep-14	Charlottesville PD
16-Apr-14	30-Apr-14	DFS Central Lab
06-Nov-08	16-Apr-14	Kenbridge PD
16-Jun-08	06-Nov-08	DFS Central Lab

Handwritten initials: HNS, MMB

# INTOX EC/IR II Quality Assurance Worksheet

Instrument Serial Number: 010514      Worksheet Start Date: 9/8/2014

Location: Albemarle County Regional Jail

Address: 160 Peregrory Ln Charlottesville VA 22902

DFS Technician: Heather Stanton      License No.: 18910

Laboratory       On-Site

Site Specification: No detrimental environmental conditions exist.

Instrument Barometer (mm HG): 752      Reference Barometer (mm HG): 753

Reference Barometer(RB)Serial #: 009113      RB Calibration Due: 4/28/2015

### Measurement Assurance Check

<b>Standard (sea level)</b>	<b>PA Target</b>	<b>minimum</b>	<b>maximum</b>	<b>Sample 1</b>
0.300		-0.003	0.003	<b>Sample 2</b>
<b>Precision</b>		<b>sample min</b>	<b>sample max</b>	<b>Sample 3</b>
0		0.000	0.000	

<b>Standard (sea level)</b>	<b>PA Target</b>	<b>minimum</b>	<b>maximum</b>	<b>Sample 1</b>
0.100		-0.003	0.003	<b>Sample 2</b>
<b>Precision</b>		<b>sample min</b>	<b>sample max</b>	<b>Sample 3</b>
0		0.000	0.000	

Dry gas standard Lot No. (with tank no.): AG230501-35

- Replaced dry gas standard (+O-ring)
- Installed at Location
- Removed to DFS-Central

- Supplies
- Mouthpieces
- Certificates of Analysis
- Operator Worksheet
- Other:



### Notes:

Unable to complete instrument test and measurement assurance checks due to instrument being disabled.

Instrument Serial Number

010514

Certification Date

Calibrated

Certified

Measurement Assurance Check

Instrument Test

Troubleshooting/Maintenance Worksheet Completed

(Successful completion denotes satisfactory condition of the item.)

<b>Standard (sea level)</b>	<b>PA Target</b>	<b>minimum</b>	<b>maximum</b>	<b>Sample 1</b>
0.250		0.000	0.000	<b>Sample 2</b>
<b>Precision</b>		<b>sample min</b>	<b>sample max</b>	<b>Sample 3</b>
0		0.000	0.000	

<b>Standard (sea level)</b>	<b>PA Target</b>	<b>minimum</b>	<b>maximum</b>	<b>Sample 1</b>
0.150		0.000	0.000	<b>Sample 2</b>
<b>Precision</b>		<b>sample min</b>	<b>sample max</b>	<b>Sample 3</b>
0		0.000	0.000	

<b>Standard (sea level)</b>	<b>PA Target</b>	<b>minimum</b>	<b>maximum</b>	<b>Sample 1</b>
0.080		-0.003	0.003	<b>Sample 2</b>
<b>Precision</b>		<b>sample min</b>	<b>sample max</b>	<b>Sample 3</b>
0		0.000	0.000	

<b>Standard (sea level)</b>	<b>PA Target</b>	<b>minimum</b>	<b>maximum</b>	<b>Sample 1</b>
0.020		-0.003	0.003	<b>Sample 2</b>
<b>Precision</b>		<b>sample min</b>	<b>sample max</b>	<b>Sample 3</b>
0		0.000	0.000	

All measurements are in g/210L

Estimation of Uncertainty of Measurement and traceability records are located within the Breath Alcohol Section

DFS Technician Heath M. Stark Date 9/8/14

Issuing Analyst Michelle Buller Date 9/11/14

*MSK*

Department of Forensic Science  
Breath Alcohol Section  
Trouble Call Log  
9/8/2014 To 9/8/2014

*LHO*  
*mmB*

**Instrument Serial Number:** 010514

08-Sep-14

---

**Date:** 08-Sep-14    **Time:** 8:05 AM    **Remote:** True    **Tech:** HMS (18910)    **Location:** Charlottesville PD

**Indication:** DISABLED

**Problem:** OPERATOR STATED THAT THE INSTRUMENT WAS DISABLED.

**Technician Response:** SITE VISIT CONDUCTED.

**Resolution:** REMOVED TO DFS-Central

---