

Department of Forensic Science
Breath Alcohol Section
Instrument Location History
1/1/1990 To 9/18/2014

DBH
JL

Instrument Serial Number: 010588

As of 18-Sep-14

Start Date	End Date	Agency Name
17-Sep-14		Dinwiddie County Sheriff's Office
28-Aug-14	17-Sep-14	DFS Central Lab
01-Apr-14	28-Aug-14	Fairfax City PD
13-Mar-14	01-Apr-14	DFS Central Lab
30-Sep-11	13-Mar-14	Smyth County SO
09-Jun-10	30-Sep-11	DFS Central Lab
17-Mar-09	09-Jun-10	Henrico County PD
04-Aug-08	17-Mar-09	DFS Central Lab

JBH
gob

INTOX EC/IR II Quality Assurance Worksheet

Instrument Serial Number **010588** Worksheet Start Date **9/17/2014**
 Location **Dinwiddie CO SO**
 Address **14230 Sycamore Dr Dinwiddie VA 23841**
 DFS Technician **Donald Hall** License No. **21250**

Laboratory On-Site

Site Specification: No detrimental environmental conditions exist.

Instrument Barometer (mm HG) **756** Reference Barometer (mm HG) **755**

Reference Barometer(RB)Serial # **009109** RB Calibration Due **7/28/2015**

Measurement Assurance Check

Standard (sea level)	PA Target	minimum	maximum	Sample 1	Sample 2	Sample 3
0.300	0.298	0.289	0.306	0.304	0.303	0.303
Precision		sample min	sample max			
0.001		0.303	0.304			

Standard (sea level)	PA Target	minimum	maximum	Sample 1	Sample 2	Sample 3
0.100	0.099	0.096	0.102	0.100	0.100	0.101
Precision		sample min	sample max			
0.001		0.100	0.101			


Dry gas standard Lot No. (with tank no.) **AG404902-05**

- Replaced dry gas standard (+O-ring)
- Installed at Location
- Removed to DFS-Central

Supplies
 Mouthpieces
 Certificates of Analysis
 Operator Worksheet
 Other:



Notes:



Instrument Serial Number

010588

Certification Date

Calibrated

Certified

Measurement Assurance Check

Instrument Test

Troubleshooting/Maintenance Worksheet Completed

(Successful completion denotes satisfactory condition of the item.)

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.250		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.150		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.080		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.020		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

All measurements are in g/210L

Estimation of Uncertainty of Measurement and traceability records are located within the Breath Alcohol Section

DFS Technician

Donald B Hall

Date

9/18/14

Issuing Analyst

JM
msk

Date

9/18/14

DBH
jd

Intox EC/IR-11: Accuracy Check

Department of Forensic Science

Serial Number: 010588 Test Number: 1691

Test Date: 09/17/2014 Test Time: 10:55 EDT

Dry Gas Target: 0.298

Lot Number: AG421201-06 Exp Date: 07/31/2016

Tank Pressure: 88 psi Barometric Pressure: 755 mmHg

System Check: Passed

Test	g/210L	Time
BLK	0.000	10:56
CHK	0.304	10:57
BLK	0.000	10:58
CHK	0.303	10:59
BLK	0.000	11:01
CHK	0.303	11:02

Test Status: Success

Calibration CRC: 0FF3A8EE

1

DBH
W

intox EC/1K-11: Accuracy Check

Department of Forensic Science

Serial Number: 010588 Test Number: 1692
Test Date: 09/17/2014 Test Time: 11:05 EDT
Dry Gas Target: 0.099
Lot Number: AG404902-05 Exp Date: 02/18/2016
Tank Pressure: 620 psi Barometric Pressure: 755 mmHg
System Check: Passed

Test	g/210L	Time
BLK	0.000	11:05
CHK	0.100	11:06
BLK	0.000	11:08
CHK	0.100	11:09
BLK	0.000	11:10
CHK	0.101	11:11

Test status: Success

Calibration CRC: 0FF3A8EE

DBH
rd



COMMONWEALTH OF VIRGINIA
DEPARTMENT OF FORENSIC SCIENCE

CERTIFICATE OF BLOOD ALCOHOL ANALYSIS
AS DETERMINED BY A CHEMICAL TEST OF THE ACCUSED'S BREATH

NAME OF ACCUSED INSTRUMENT, TEST,	NAME OF COURT DFS
--	--------------------------

BREATH ANALYSIS

SAMPLE EXAMINED AND TEST CONDUCTED BY HALL, DONALD, B		AGENCY DFS Central Lab
DFS LICENSE NUMBER 21250	LICENSE EXPIRES 10/01/2014	DATE TEST CONDUCTED 09/17/2014
TEST EQUIPMENT NUMBER 010588		

RESULTS: TIME SAMPLE TAKEN 11:21 EDT

SAMPLE'S ALCOHOL CONTENT 0.00 GRAMS PER 210 LITERS OF BREATH

ATTEST:

I CERTIFY THAT THE ABOVE IS AN ACCURATE RECORD OF THE TEST CONDUCTED; THAT THE TEST WAS CONDUCTED WITH THE TYPE OF EQUIPMENT AND IN ACCORDANCE WITH THE METHODS APPROVED BY THE DEPARTMENT OF FORENSIC SCIENCE; THAT THE TEST WAS CONDUCTED IN ACCORDANCE WITH THE DEPARTMENT'S SPECIFICATIONS; THAT PRIOR TO ADMINISTRATION OF THE TEST THE ACCUSED WAS ADVISED OF HIS RIGHT TO OBSERVE THE PROCESS AND SEE THE BLOOD ALCOHOL READING ON THE EQUIPMENT USED TO PERFORM THE BREATH TEST, AND THAT I POSSESS A VALID LICENSE TO CONDUCT SUCH TEST, GIVEN UNDER MY HAND THIS _____ DAY OF _____, 20_____.

BREATH TEST OPERATOR

I HAVE RECEIVED A COPY OF THIS CERTIFICATE OF ANALYSIS _____ SUBJECT'S SIGNATURE

SUBJECT REFUSED TO SIGN FOR COPY OF CERTIFICATE OF ANALYSIS _____ OPERATOR'S SIGNATURE

FOR COURT



IntoxNet MIS Report
Report Generated 18 Sep 2014 at 09:25

DBH
Jrd

Test Results

Instrument Serial Number 010588

Test # 001693 Subject Test

Test Location 1 Department of
Test Date 17 Sep 2014

Test Location 2 Forensic Science
Test Time 11:15
Remote/Local Local

Test Location 3
System Check Passed

Operator's Last Name HALL
Agency DFS Central Lab

Operator's First Name DONALD

Operator's Middle Initial B

Card Serial Number 121250

License Number 21250

Effective Date 10/01/2012

Expiration Date 10/01/2014

Subject's Last Name INSTRUMENT

Subject's First Name TEST

Subject's Middle Initial

Subject's Date of Birth 00/00/0000

Subject's Sex Male

Driver's License Number

Driver's License Expiration 00/00/0000

Driver's License State

Court Name DFS

End Date 17 Sep 2014 End Time 11:22

Result Time 11:21

Result Date 17 Sep 2014 Result 0.00

Data Type DIAG

Sample Value Pass

Sample Time 11:15

Data Type BLK

Sample Value 0.000

Sample Time 11:16

Data Type CHK

Sample Value 0.100

Sample Time 11:17

Data Type BLK

Sample Value 0.000

Sample Time 11:18

Data Type SUBJ

Sample Value 0.000

Sample Time 11:19

Data Type BLK

Sample Value 0.000

Sample Time 11:20

Data Type SUBJ

Sample Value 0.000

Sample Time 11:21

Data Type BLK

Sample Value 0.000

Sample Time 11:22

Standard Type Dry Gas Std

Standard Value 0.099

Standard Lot Number AG404902-05

Standard Expiration Date 02/18/2016

Tank Pressure 610

Barometric Pressure 755 mmHg

Blow Sample Number 1 Blow Duration 3.20 sec

Blow Volume 2106 cc End-of-Blow Time 11:19

Blow Sample Number 2 Blow Duration 4.21 sec

Blow Volume 2341 cc End-of-Blow Time 11:21

Tamper Evident Stamp 24bfe0e0

Test Status Code 0

Test Status Success