

Department of Forensic Science
Breath Alcohol Section
Instrument Location History
1/1/1990 To 10/10/2014

DBH
mmB

Instrument Serial Number: 010526 As of 10-Oct-14

Start Date	End Date	Agency Name
07-Oct-14		Louisa County SO
01-Oct-14	07-Oct-14	DFS Central Lab
30-Apr-14	01-Oct-14	Buchanan County SO
24-Apr-14	30-Apr-14	DFS Central Lab
26-Oct-12	24-Apr-14	Manassas Park PD
08-Jun-12	26-Oct-12	DFS Central Lab
07-May-09	08-Jun-12	Augusta County SO
02-Jun-08	07-May-09	DFS Central Lab

DBH
mmB

INTOX EC/IR II Quality Assurance Worksheet

Instrument Serial Number **010526** Worksheet Start Date **10/7/2014**
 Location **Louisa CO SO**
 Address **1 Woolfolk Ave Louisa VA 23093**
 DFS Technician **Donald Hall** License No. **21250**

Laboratory On-Site

Site Specification: No detrimental environmental conditions exist.

Instrument Barometer (mm HG) **754** Reference Barometer (mm HG) **751**
 Reference Barometer(RB)Serial # **009109** RB Calibration Due **7/28/2015**

Measurement Assurance Check

Standard (sea level)	PA Target	minimum	maximum	Sample 1	Sample 2	Sample 3
0.300	0.297	0.288	0.305	0.297	0.296	0.295
Precision		sample min	sample max			
0.002		0.295	0.297			


Standard (sea level)	PA Target	minimum	maximum	Sample 1	Sample 2	Sample 3
0.100	0.099	0.096	0.102	0.098	0.098	0.098
Precision		sample min	sample max			
0		0.098	0.098			

Dry gas standard Lot No. (with tank no.) **AG320502-28**

- Replaced dry gas standard (+O-ring)
- Installed at Location
- Removed to DFS-Central

Supplies
 Mouthpieces **50**
 Certificates of Analysis
 Operator Worksheet **50**
 Other:

Notes:



Instrument Serial Number

010526

Certification Date

Calibrated

Certified

Measurement Assurance Check

Instrument Test

Troubleshooting/Maintenance Worksheet Completed

(Successful completion denotes satisfactory condition of the item.)

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.250		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.150		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.080		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.020		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

All measurements are in g/210L

Estimation of Uncertainty of Measurement and traceability records are located within the Breath Alcohol Section

DFS Technician

Donald B Hall

Date

10/10/14

Issuing Analyst

Michelle Buller
Jed

Date

10/10/14

DBH
mmB

Intox EC/IR-11: Accuracy Check

Department of Forensic Science

Serial Number: 010526 Test Number: 944

Test Date: 10/07/2014 Test Time: 11:38 EDT

Dry Gas Target: 0.297

Lot Number: AG421201-02 Exp Date: 07/31/2016

Tank Pressure: 365 psi Barometric Pressure: 753 mmHg

System Check: Passed

Test	g/210L	Time
BLK	0.000	11:38
CHK	0.297	11:39
BLK	0.000	11:41
CHK	0.296	11:42
BLK	0.000	11:43
CHK	0.295	11:44

Test Status: Success

Calibration CRC: 317E2DEE

Intox EC/IR-11: Accuracy Check

DBH
mmB

Department of Forensic Science

Serial Number: 010526 Test Number: 945

Test Date: 10/07/2014 Test Time: 11:47 EDT

Dry Gas Target: 0.099

Lot Number: AG320502-28 Exp Date: 07/24/2015

Tank Pressure: 688 psi Barometric Pressure: 753 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	11:47
CHK	0.098	11:48
BLK	0.000	11:50
CHK	0.098	11:50
BLK	0.000	11:52
CHK	0.098	11:52

Test status: *Success*

Calibration CRC: 317E2DEE

DBH
mms



COMMONWEALTH OF VIRGINIA
DEPARTMENT OF FORENSIC SCIENCE

CERTIFICATE OF BLOOD ALCOHOL ANALYSIS
AS DETERMINED BY A CHEMICAL TEST OF THE ACCUSED'S BREATH

NAME OF ACCUSED INSTRUMENT, TEST,	NAME OF COURT DFS
--------------------------------------	----------------------

BREATH ANALYSIS

SAMPLE EXAMINED AND TEST CONDUCTED BY HALL, DONALD, B		AGENCY DFS Central Lab
DFS LICENSE NUMBER 21250	LICENSE EXPIRES 10/01/2016	DATE TEST CONDUCTED 10/07/2014
TEST EQUIPMENT NUMBER 010526		

12:02 EDT

RESULTS: TIME SAMPLE TAKEN _____
0.00

SAMPLE'S ALCOHOL CONTENT _____ GRAMS PER 210 LITERS OF BREATH

ATTEST:

I CERTIFY THAT THE ABOVE IS AN ACCURATE RECORD OF THE TEST CONDUCTED; THAT THE TEST WAS CONDUCTED WITH THE TYPE OF EQUIPMENT AND IN ACCORDANCE WITH THE METHODS APPROVED BY THE DEPARTMENT OF FORENSIC SCIENCE; THAT THE TEST WAS CONDUCTED IN ACCORDANCE WITH THE DEPARTMENT'S SPECIFICATIONS; THAT PRIOR TO ADMINISTRATION OF THE TEST THE ACCUSED WAS ADVISED OF HIS RIGHT TO OBSERVE THE PROCESS AND SEE THE BLOOD ALCOHOL READING ON THE EQUIPMENT USED TO PERFORM THE BREATH TEST, AND THAT I POSSESS A VALID LICENSE TO CONDUCT SUCH TEST, GIVEN UNDER MY HAND THIS _____ DAY OF _____, 20 _____.

BREATH TEST OPERATOR

I HAVE RECEIVED A COPY OF THIS CERTIFICATE OF ANALYSIS _____
SUBJECT'S SIGNATURE

SUBJECT REFUSED TO SIGN FOR COPY OF CERTIFICATE OF ANALYSIS _____
OPERATOR'S SIGNATURE



IntoxNet MIS Report
Report Generated 10 Oct 2014 at 10:35

DBH
mmb

Test Results

Instrument Serial Number 010526

Test # 000946 Subject Test

Test Location 1 Department of
Test Date 07 Oct 2014

Test Location 2 Forensic Science
Test Time 11:55
Remote/Local Local

Test Location 3
System Check Passed

Operator's Last Name HALL
Agency DFS Central Lab

Operator's First Name DONALD

Operator's Middle Initial B

Card Serial Number 121250

License Number 21250

Effective Date 10/01/2014

Expiration Date 10/01/2016

Subject's Last Name INSTRUMENT

Subject's First Name TEST

Subject's Middle Initial

Subject's Date of Birth 00/00/0000

Subject's Sex Male

Driver's License Number

Driver's License Expiration 00/00/0000

Driver's License State

Court Name DFS

End Date 07 Oct 2014 End Time 12:03

Result Time 12:02

Result Date 07 Oct 2014 Result 0.00

Data Type DIAG

Sample Value Pass

Sample Time 11:55

Data Type BLK

Sample Value 0.000

Sample Time 11:56

Data Type CHK

Sample Value 0.098

Sample Time 11:57

Data Type BLK

Sample Value 0.000

Sample Time 11:58

Data Type SUBJ

Sample Value 0.000

Sample Time 11:59

Data Type BLK

Sample Value 0.000

Sample Time 12:00

Data Type SUBJ

Sample Value 0.000

Sample Time 12:02

Data Type BLK

Sample Value 0.000

Sample Time 12:03

Standard Type Dry Gas Std

Standard Value 0.099

Standard Lot Number AG320502-28

Standard Expiration Date 07/24/2015

Tank Pressure 683

Barometric Pressure 753 mmHg

Blow Sample Number 1 Blow Duration 4.16 sec

Blow Volume 2076 cc End-of-Blow Time 11:59

Blow Sample Number 2 Blow Duration 0.07 sec

Blow Volume 17 cc End-of-Blow Time 12:01

Blow Sample Number 3 Blow Duration 3.69 sec

Blow Volume 1921 cc End-of-Blow Time 12:02

Tamper Evident Stamp 0ad8ec6a

Test Status Code 0

Test Status Success