

m m B

Department of Forensic Science
Breath Alcohol Section
Instrument Location History
1/1/1990 To 10/23/2014

Instrument Serial Number: 010601 **As of 23-Oct-14**

Start Date	End Date	Agency Name
22-Oct-14		DFS Central Lab
20-May-14	22-Oct-14	Mathews County SO
08-May-14	20-May-14	DFS Central Lab
02-Oct-09	08-May-14	Page County SO
04-Aug-08	02-Oct-09	DFS Central Lab

mmb

INTOX EC/IR II Quality Assurance Worksheet

Instrument Serial Number **010601** Worksheet Start Date **10/22/2014**

Location **Mathews County Sheriffs Office**

Address **10622 Buckley Hall Rd Mathews VA 23109**

DFS Technician **Michelle Billeter** License No. **26417**

Laboratory On-Site

Site Specification: No detrimental environmental conditions exist.

Instrument Barometer (mm HG) **754** Reference Barometer (mm HG) **758**

Reference Barometer(RB)Serial # **009111** RB Calibration Due **5/19/2015**

Measurement Assurance Check

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.303
0.300	0.298	0.289	0.306	Sample 2	0.302
Precision		sample min	sample max	Sample 3	0.302
0.001		0.302	0.303		

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.101
0.100	0.099	0.096	0.102	Sample 2	0.101
Precision		sample min	sample max	Sample 3	0.101
0		0.101	0.101		

Dry gas standard Lot No. (with tank no.) **AG320502-29**

- Replaced dry gas standard (+O-ring)
- Installed at Location
- Removed to DFS-Central

- Supplies
- Mouthpieces
- Certificates of Analysis
- Operator Worksheet
- Other:



Notes:

Instrument Serial Number

010601

Certification Date

Calibrated

Certified

Measurement Assurance Check

Instrument Test

Troubleshooting/Maintenance Worksheet Completed

(Successful completion denotes satisfactory condition of the item.)

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.250		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.150		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.080		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.020		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

All measurements are in g/210L

Estimation of Uncertainty of Measurement and traceability records are located within the Breath Alcohol Section

DFS Technician

Michelle Bellid

Date

10/23/14

Issuing Analyst

Michelle Bellid

Date

10/23/14

TW

MMB



COMMONWEALTH OF VIRGINIA
DEPARTMENT OF FORENSIC SCIENCE

CERTIFICATE OF BLOOD ALCOHOL ANALYSIS
AS DETERMINED BY A CHEMICAL TEST OF THE ACCUSED'S BREATH

MMB
102214

NAME OF ACCUSED INSTRUMENT, TEST,	NAME OF COURT DES
---	-----------------------------

BREATH ANALYSIS

SAMPLE EXAMINED AND TEST CONDUCTED BY BILLETTER, MICHELLE, M		AGENCY DES Central Lab
DFS LICENSE NUMBER 26417	LICENSE EXPIRES 10/01/2016	DATE TEST CONDUCTED 10/22/2014
TEST EQUIPMENT NUMBER 010601		

RESULTS: TIME SAMPLE TAKEN 10:40 EDT

SAMPLE'S ALCOHOL CONTENT sequence GRAMS PER 210 LITERS OF BREATH

ATTEST:

I CERTIFY THAT THE ABOVE IS AN ACCURATE RECORD OF THE TEST CONDUCTED; THAT THE TEST WAS CONDUCTED WITH THE TYPE OF EQUIPMENT AND IN ACCORDANCE WITH THE METHODS APPROVED BY THE DEPARTMENT OF FORENSIC SCIENCE; THAT THE TEST WAS CONDUCTED IN ACCORDANCE WITH THE DEPARTMENT'S SPECIFICATIONS; THAT PRIOR TO ADMINISTRATION OF THE TEST THE ACCUSED WAS ADVISED OF HIS RIGHT TO OBSERVE THE PROCESS AND SEE THE BLOOD ALCOHOL READING ON THE EQUIPMENT USED TO PERFORM THE BREATH TEST, AND THAT I POSSESS A VALID LICENSE TO CONDUCT SUCH TEST, GIVEN UNDER MY HAND THIS _____ DAY OF _____, 20_____.

BREATH TEST OPERATOR

I HAVE RECEIVED A COPY OF THIS CERTIFICATE OF ANALYSIS _____ SUBJECT'S SIGNATURE

SUBJECT REFUSED TO SIGN FOR COPY OF CERTIFICATE OF ANALYSIS _____ OPERATOR'S SIGNATURE

MMB



COMMONWEALTH OF VIRGINIA
DEPARTMENT OF FORENSIC SCIENCE

CERTIFICATE OF BLOOD ALCOHOL ANALYSIS
AS DETERMINED BY A CHEMICAL TEST OF THE ACCUSED'S BREATH

NAME OF ACCUSED INSTRUMENT, TEST,	NAME OF COURT DFS
---	-----------------------------

BREATH ANALYSIS

SAMPLE EXAMINED AND TEST CONDUCTED BY BILLETER, MICHELLE, M		AGENCY DFS Central Lab
DFS LICENSE NUMBER 26417	LICENSE EXPIRES 10/01/2016	DATE TEST CONDUCTED 10/22/2014
TEST EQUIPMENT NUMBER 010601		

RESULTS: TIME SAMPLE TAKEN ~~10:40 EDT~~

SAMPLE'S ALCOHOL CONTENT ~~0.00~~ GRAMS PER 210 LITERS OF BREATH

ATTEST:

I CERTIFY THAT THE ABOVE IS AN ACCURATE RECORD OF THE TEST CONDUCTED; THAT THE TEST WAS CONDUCTED WITH THE TYPE OF EQUIPMENT AND IN ACCORDANCE WITH THE METHODS APPROVED BY THE DEPARTMENT OF FORENSIC SCIENCE; THAT THE TEST WAS CONDUCTED IN ACCORDANCE WITH THE DEPARTMENT'S SPECIFICATIONS; THAT PRIOR TO ADMINISTRATION OF THE TEST THE ACCUSED WAS ADVISED OF HIS RIGHT TO OBSERVE THE PROCESS AND SEE THE BLOOD ALCOHOL READING ON THE EQUIPMENT USED TO PERFORM THE BREATH TEST, AND THAT I POSSESS A VALID LICENSE TO CONDUCT SUCH TEST, GIVEN UNDER MY HAND THIS _____ DAY OF _____, 20_____.

BREATH TEST OPERATOR

I HAVE RECEIVED A COPY OF THIS CERTIFICATE OF ANALYSIS _____ SUBJECT'S SIGNATURE

SUBJECT REFUSED TO SIGN FOR COPY OF CERTIFICATE OF ANALYSIS _____ OPERATOR'S SIGNATURE

mmb



IntoxNet MIS Report

Report Generated 22 Oct 2014 at 11:15

Test Results

Instrument Serial Number 010601

Test # 000712 Subject Test

Test Location 1 Department of
Test Date 22 Oct 2014

Test Location 2 Forensic Science
Test Time 10:42
Remote/Local Local

Test Location 3
System Check Passed

Operator's Last Name BILLETER
Agency DFS Central Lab

Operator's First Name MICHELLE

Operator's Middle Initial M

Card Serial Number 126417

License Number 26417

Effective Date 10/01/2014

Expiration Date 10/01/2016

Subject's Last Name INSTRUMENT

Subject's First Name TEST

Subject's Middle Initial

Subject's Date of Birth 00/00/0000

Subject's Sex Male

Driver's License Number

Driver's License Expiration 00/00/0000

Driver's License State

Court Name DFS

End Date 22 Oct 2014 End Time 10:49

Result Time 10:48

Result Date 22 Oct 2014 Result 0.00

Data Type DIAG

Sample Value Pass

Sample Time 10:42

Data Type BLK

Sample Value 0.000

Sample Time 10:43

Data Type CHK

Sample Value 0.102

Sample Time 10:44

Data Type BLK

Sample Value 0.000

Sample Time 10:45

Data Type SUBJ

Sample Value 0.000

Sample Time 10:46

Data Type BLK

Sample Value 0.000

Sample Time 10:47

Data Type SUBJ

Sample Value 0.000

Sample Time 10:48

Data Type BLK

Sample Value 0.000

Sample Time 10:49

Standard Type Dry Gas Std

Standard Value 0.099

Standard Lot Number AG320502-29

Standard Expiration Date 07/24/2015

Tank Pressure 1022

Barometric Pressure 754 mmHg

Blow Sample Number 1 Blow Duration 3.75 sec

Blow Volume 1848 cc End-of-Blow Time 10:46

Blow Sample Number 2 Blow Duration 3.65 sec

Blow Volume 1831 cc End-of-Blow Time 10:48

Tamper Evident Stamp 8183ba8b

Test Status Code 0

Test Status Success

Intox EC/IR-II: Accuracy Check

mms

Department of Forensic Science

Serial Number: 010601 Test Number: 713

Test Date: 10/22/2014 Test Time: 10:54 EDT

Dry Gas Target: 0.099

Lot Number: AG320502-29 Exp Date: 07/24/2015

Tank Pressure: 1017 psi Barometric Pressure: 754 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	10:55
CHK	0.101	10:55
BLK	0.000	10:57
CHK	0.101	10:58
BLK	0.000	10:59
CHK	0.101	11:00

Test Status: *Success*

Calibration CRC: 6577EBA4

Intox EC/IR-II: Accuracy Check*Department of Forensic Science*

Serial Number: 010601 Test Number: 714

Test Date: 10/22/2014 Test Time: 11:02 EDT

Dry Gas Target: 0.298

Lot Number: AG421201-03 Exp Date: 07/31/2016

Tank Pressure: 404 psi Barometric Pressure: 754 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	11:03
CHK	0.303	11:04
BLK	0.000	11:06
CHK	0.302	11:06
BLK	0.000	11:08
CHK	0.302	11:09

Test Status: *Success*

Calibration CRC: 6577EBA4