

Department of Forensic Science
Breath Alcohol Section
Instrument Location History
1/1/1990 To 10/10/2014

DBH
mmB

Instrument Serial Number: 010620 As of 10-Oct-14

Start Date	End Date	Agency Name
09-Oct-14		DFS Central Lab
08-May-14	09-Oct-14	Page County SO
30-Apr-14	08-May-14	DFS Central Lab
08-Jan-09	30-Apr-14	Louisa County SO
15-Aug-08	08-Jan-09	DFS Central Lab

DBH
mmB

INTOX EC/IR II Quality Assurance Worksheet

Instrument Serial Number **010620** Worksheet Start Date **10/9/2014**
 Location **Page CO SO**
 Address **108 S Court St Luray VA 22835**
 DFS Technician **Donald Hall** License No. **21250**

Laboratory On-Site

Site Specification: No detrimental environmental conditions exist.

Instrument Barometer (mm HG) **740** Reference Barometer (mm HG) **740**
 Reference Barometer(RB)Serial # **009109** RB Calibration Due **7/28/2015**

Measurement Assurance Check

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.293
0.300	0.292	0.283	0.300	Sample 2	0.293
Precision		sample min	sample max	Sample 3	0.292
0.001		0.292	0.293		

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.098
0.100	0.097	0.094	0.100	Sample 2	0.098
Precision		sample min	sample max	Sample 3	0.098
0		0.098	0.098		

Dry gas standard Lot No. (with tank no.) **AG320502-23**

- Replaced dry gas standard (+O-ring)
- Installed at Location
- Removed to DFS-Central

Supplies
 Mouthpieces
 Certificates of Analysis
 Operator Worksheet
 Other:



Notes:

Instrument Serial Number

010620

Certification Date

Calibrated

Certified

Measurement Assurance Check

Instrument Test

Troubleshooting/Maintenance Worksheet Completed

(Successful completion denotes satisfactory condition of the item.)

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.250		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.150		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.080		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.020		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

All measurements are in g/210L

Estimation of Uncertainty of Measurement and traceability records are located within the Breath Alcohol Section

DFS Technician Donald B Hall Date 10/10/14

Issuing Analyst Muzelle Belliter Date 10/14/14
TLW

DBH
mmB



COMMONWEALTH OF VIRGINIA
DEPARTMENT OF FORENSIC SCIENCE

CERTIFICATE OF BLOOD ALCOHOL ANALYSIS
AS DETERMINED BY A CHEMICAL TEST OF THE ACCUSED'S BREATH

NAME OF ACCUSED INSTRUMENT, TEST,	NAME OF COURT DFS
------------------------------------------	--------------------------

BREATH ANALYSIS

SAMPLE EXAMINED AND TEST CONDUCTED BY HALL, DONALD, B		AGENCY DFS Central Lab
DFS LICENSE NUMBER 21250	LICENSE EXPIRES 10/01/2016	DATE TEST CONDUCTED 10/09/2014
TEST EQUIPMENT NUMBER 010620		

RESULTS: TIME SAMPLE TAKEN 12:46 EDT

SAMPLE'S ALCOHOL CONTENT 0.00 GRAMS PER 210 LITERS OF BREATH

ATTEST:

I CERTIFY THAT THE ABOVE IS AN ACCURATE RECORD OF THE TEST CONDUCTED; THAT THE TEST WAS CONDUCTED WITH THE TYPE OF EQUIPMENT AND IN ACCORDANCE WITH THE METHODS APPROVED BY THE DEPARTMENT OF FORENSIC SCIENCE; THAT THE TEST WAS CONDUCTED IN ACCORDANCE WITH THE DEPARTMENT'S SPECIFICATIONS; THAT PRIOR TO ADMINISTRATION OF THE TEST THE ACCUSED WAS ADVISED OF HIS RIGHT TO OBSERVE THE PROCESS AND SEE THE BLOOD ALCOHOL READING ON THE EQUIPMENT USED TO PERFORM THE BREATH TEST, AND THAT I POSSESS A VALID LICENSE TO CONDUCT SUCH TEST, GIVEN UNDER MY HAND THIS _____ DAY OF _____, 20_____.

BREATH TEST OPERATOR

I HAVE RECEIVED A COPY OF THIS CERTIFICATE OF ANALYSIS _____ SUBJECT'S SIGNATURE

SUBJECT REFUSED TO SIGN FOR COPY OF CERTIFICATE OF ANALYSIS _____ OPERATOR'S SIGNATURE



IntoxNet MIS Report

Report Generated 10 Oct 2014 at 15:57

DBH
MNB

Test Results

Instrument Serial Number 010620

Test # 000871 Subject Test

Test Location 1 Department of
Test Date 09 Oct 2014

Test Location 2 Forensic Science
Test Time 12:39
Remote/Local Local

Test Location 3
System Check Passed

Operator's Last Name HALL
Agency DFS Central Lab

Operator's First Name DONALD

Operator's Middle Initial B

License Number 21250

Card Serial Number 121250

Effective Date 10/01/2014

Expiration Date 10/01/2016

Subject's Last Name INSTRUMENT

Subject's First Name TEST

Subject's Middle Initial

Subject's Date of Birth 00/00/0000

Subject's Sex Male

Driver's License Number

Driver's License Expiration 00/00/0000

Driver's License State

Court Name DFS

End Date 09 Oct 2014 End Time 12:47

Result Time 12:46

Result Date 09 Oct 2014 Result 0.00

Data Type DIAG Sample Value Pass Sample Time 12:39

Data Type BLK Sample Value 0.000 Sample Time 12:40

Data Type CHK Sample Value 0.099 Sample Time 12:41

Data Type BLK Sample Value 0.000 Sample Time 12:42

Data Type SUBJ Sample Value 0.000 Sample Time 12:43

Data Type BLK Sample Value 0.000 Sample Time 12:44

Data Type SUBJ Sample Value 0.000 Sample Time 12:46

Data Type BLK Sample Value 0.000 Sample Time 12:47

Standard Type Dry Gas Std Standard Value 0.097

Standard Lot Number AG320502-23 Standard Expiration Date 07/24/2015

Tank Pressure 821 Barometric Pressure 740 mmHg

Blow Sample Number 1 Blow Duration 5.46 sec Blow Volume 2999 cc End-of-Blow Time 12:43

Blow Sample Number 2 Blow Duration 3.77 sec Blow Volume 2109 cc End-of-Blow Time 12:46

Tamper Evident Stamp 729daaba Test Status Code 0

Test Status Success

DBH
mmB

Intox EC/IR-II: Accuracy Check

Department of Forensic Science

Serial Number: 010620 Test Number: 872

Test Date: 10/09/2014 Test Time: 12:48 EDT

Dry Gas Target: 0.097

Lot Number: AG320502-23 Exp Date: 07/24/2015

Tank Pressure: 816 psi Barometric Pressure: 740 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	12:49
CHK	0.098	12:50
BLK	0.000	12:52
CHK	0.098	12:52
BLK	0.000	12:54
CHK	0.098	12:54

Test Status: *Success*

Calibration CRC: 0C2FEE1E

JBH
mmb

Intox EC/IR-II: Accuracy Check

Department of Forensic Science

Serial Number: 010620 Test Number: 873

Test Date: 10/09/2014 Test Time: 12:57 EDT

Dry Gas Target: 0.292

Lot Number: AG421201-02 Exp Date: 07/31/2016

Tank Pressure: 272 psi Barometric Pressure: 740 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	12:57
CHK	0.293	12:58
BLK	0.000	13:00
CHK	0.293	13:00
BLK	0.000	13:02
CHK	0.292	13:03

Test Status: *Success*

Calibration CRC: 0C2FEE1E