

Department of Forensic Science
Breath Alcohol Section
Instrument Location History
1/1/1990 To 11/14/2014

DBH
mmb

Instrument Serial Number: 010589

As of 14-Nov-14

Start Date	End Date	Agency Name
13-Nov-14		New River Valley Regional Jail
22-Oct-14	13-Nov-14	DFS Central Lab
20-May-14	22-Oct-14	Gloucester County SO
06-May-14	20-May-14	DFS Central Lab
06-Jan-09	06-May-14	Waynesboro PD
04-Aug-08	06-Jan-09	DFS Central Lab

DBH
mmB

INTOX EC/IR II Quality Assurance Worksheet

Instrument Serial Number **010589** Worksheet Start Date **11/13/2014**

Location **New River Valley Regional Jail**

Address **108 Baker Rd Dublin VA 24060**

DFS Technician **Donald Hall** License No. **21250**

Laboratory On-Site

Site Specification: No detrimental environmental conditions exist.

Instrument Barometer (mm HG) **706** Reference Barometer (mm HG) **707**

Reference Barometer(RB)Serial # **009109** RB Calibration Due **7/28/2015**

Measurement Assurance Check

Standard (sea level)	PA Target	minimum	maximum	Sample 1	Sample 2	Sample 3
0.300	0.279	0.270	0.287	0.285	0.284	0.283
Precision		sample min	sample max			
0.002		0.283	0.285			

Standard (sea level)	PA Target	minimum	maximum	Sample 1	Sample 2	Sample 3
0.100	0.093	0.090	0.096	0.094	0.094	0.093
Precision		sample min	sample max			
0.001		0.093	0.094			

Dry gas standard Lot No. (with tank no.) **AG406301-29**

- Replaced dry gas standard (+O-ring)
- Installed at Location
- Removed to DFS-Central

Supplies
Mouthpieces
Certificates of Analysis
Operator Worksheet
Other:

Notes:

[Redacted Notes Area]

Instrument Serial Number

010589

Certification Date

Calibrated

Certified

Measurement Assurance Check

Instrument Test

Troubleshooting/Maintenance Worksheet Completed

(Successful completion denotes satisfactory condition of the item.)

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.250		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.150		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.080		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.020		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

All measurements are in g/210L

Estimation of Uncertainty of Measurement and traceability records are located within the Breath Alcohol Section

DFS Technician Donald B Hall Date 11/14/14

Issuing Analyst Michelle Buller Date 11/17/14

JPD

DBh
mmB

Intox EC/IR-II: Accuracy Check

Department of Forensic Science

Serial Number: 010589 Test Number: 889
Test Date: 11/13/2014 Test Time: 16:10 EST
Dry Gas Target: 0.279

Lot Number: AG421201-03 Exp Date: 07/31/2016
Tank Pressure: 193 psi Barometric Pressure: 706 mmHg
System Check: Passed

Test	g/210L	Time
BLK	0.000	16:10
CHK	0.285	16:11
BLK	0.000	16:13
CHK	0.284	16:14
BLK	0.000	16:15
CHK	0.283	16:16

Test Status: Success

Calibration CRC: A54237C4

DBH
mmB

Intox EC/IR-11: Accuracy Check

Department of Forensic Science

Serial Number: 010589 Test Number: 890

Test Date: 11/13/2014 Test Time: 16:19 EST

Dry Gas Target: 0.093

Lot Number: AG406301-29 Exp Date: 03/04/2016

Tank Pressure: 1071 psi Barometric Pressure: 706 mmHg

System Check: Passed

Test	g/210L	Time
BLK	0.000	16:20
CHK	0.094	16:20
BLK	0.000	16:22
CHK	0.094	16:23
BLK	0.000	16:24
CHK	0.093	16:25

Test Status: Success

Calibration CRC: A54237C4

DSH
mmB



COMMONWEALTH OF VIRGINIA
DEPARTMENT OF FORENSIC SCIENCE

CERTIFICATE OF BLOOD ALCOHOL ANALYSIS
AS DETERMINED BY A CHEMICAL TEST OF THE ACCUSED'S BREATH

NAME OF ACCUSED INSTRUMENT, TEST,	NAME OF COURT DFS
--------------------------------------	----------------------

BREATH ANALYSIS

SAMPLE EXAMINED AND TEST CONDUCTED BY HALL, DONALD, B		AGENCY DFS Central Lab
DFS LICENSE NUMBER 21250	LICENSE EXPIRES 10/01/2016	DATE TEST CONDUCTED 11/13/2014
TEST EQUIPMENT NUMBER 010589		

RESULTS: TIME SAMPLE TAKEN 16:34 EST

SAMPLE'S ALCOHOL CONTENT 0.00 GRAMS PER 210 LITERS OF BREATH

ATTEST:

I CERTIFY THAT THE ABOVE IS AN ACCURATE RECORD OF THE TEST CONDUCTED; THAT THE TEST WAS CONDUCTED WITH THE TYPE OF EQUIPMENT AND IN ACCORDANCE WITH THE METHODS APPROVED BY THE DEPARTMENT OF FORENSIC SCIENCE; THAT THE TEST WAS CONDUCTED IN ACCORDANCE WITH THE DEPARTMENT'S SPECIFICATIONS; THAT PRIOR TO ADMINISTRATION OF THE TEST THE ACCUSED WAS ADVISED OF HIS RIGHT TO OBSERVE THE PROCESS AND SEE THE BLOOD ALCOHOL READING ON THE EQUIPMENT USED TO PERFORM THE BREATH TEST, AND THAT I POSSESS A VALID LICENSE TO CONDUCT SUCH TEST, GIVEN UNDER MY HAND THIS _____ DAY OF _____, 20_____.

BREATH TEST OPERATOR

I HAVE RECEIVED A COPY OF THIS CERTIFICATE OF ANALYSIS _____ SUBJECT'S SIGNATURE

SUBJECT REFUSED TO SIGN FOR COPY OF CERTIFICATE OF ANALYSIS _____ OPERATOR'S SIGNATURE



IntoxNet MIS Report

Report Generated 14 Nov 2014 at 09:36

DBH
mmB

Test Results

Instrument Serial Number 010589

Test # 000891 Subject Test

Test Location 1 Department of
Test Date 13 Nov 2014

Test Location 2 Forensic Science
Test Time 16:27
Remote/Local Local

Test Location 3
System Check Passed

Operator's Last Name HALL
Agency DFS Central Lab

Operator's First Name DONALD

Operator's Middle Initial B

Card Serial Number 121250

License Number 21250

Effective Date 10/01/2014

Expiration Date 10/01/2016

Subject's Last Name INSTRUMENT

Subject's First Name TEST

Subject's Middle Initial

Subject's Date of Birth 00/00/0000

Subject's Sex Male

Driver's License Number

Driver's License Expiration 00/00/0000

Driver's License State

Court Name DFS

End Date 13 Nov 2014 End Time 16:35

Result Time 16:34

Result Date 13 Nov 2014 Result 0.00

Data Type DIAG

Sample Value Pass

Sample Time 16:28

Data Type BLK

Sample Value 0.000

Sample Time 16:29

Data Type CHK

Sample Value 0.093

Sample Time 16:29

Data Type BLK

Sample Value 0.000

Sample Time 16:31

Data Type SUBJ

Sample Value 0.000

Sample Time 16:31

Data Type BLK

Sample Value 0.000

Sample Time 16:32

Data Type SUBJ

Sample Value 0.000

Sample Time 16:34

Data Type BLK

Sample Value 0.000

Sample Time 16:35

Standard Type Dry Gas Std

Standard Value 0.093

Standard Lot Number AG406301-29

Standard Expiration Date 03/04/2016

Tank Pressure 1061

Barometric Pressure 706 mmHg

Blow Sample Number 1 Blow Duration 3.61 sec

Blow Volume 1924 cc End-of-Blow Time 16:31

Blow Sample Number 2 Blow Duration 4.07 sec

Blow Volume 2041 cc End-of-Blow Time 16:34

Tamper Evident Stamp 2f0705b4

Test Status Code 0

Test Status Success