

AMS
TLW

INTOX EC/IR II Quality Assurance Worksheet

Instrument Serial Number **010448** Worksheet Start Date **12/12/2014**

Location **Virginia Beach Jail**

Address **2501 James Madison Blvd., Virginia Beach, VA 23456**

DFS Technician **Heather Stanton** License No. **18910**

Laboratory On-Site

Site Specification: No detrimental environmental conditions exist.

Instrument Barometer (mm HG) **763** Reference Barometer (mm HG) **765**

Reference Barometer(RB)Serial # **009113** RB Calibration Due **4/28/2015**

Measurement Assurance Check

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.300		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.100		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Dry gas standard Lot No. (with tank no.) **AG419702-33**

- Replaced dry gas standard (+O-ring)
- Installed at Location
- Removed to DFS-Central

- Supplies
- Mouthpieces
- Certificates of Analysis
- Operator Worksheet
- Other:



Notes:

Instrument Serial Number

010448

Certification Date

Calibrated

Certified

Measurement Assurance Check

Instrument Test

Troubleshooting/Maintenance Worksheet Completed

(Successful completion denotes satisfactory condition of the item.)

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.250		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.150		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.080		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.020		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

All measurements are in g/210L

Estimation of Uncertainty of Measurement and traceability records are located within the Breath Alcohol Section

DFS Technician

Heath W. Stork

Date

12/19/14

Issuing Analyst

J.P.D.

Date

12/19/14

Handwritten initials/signature in the top right corner.



COMMONWEALTH OF VIRGINIA
DEPARTMENT OF FORENSIC SCIENCE

CERTIFICATE OF BLOOD ALCOHOL ANALYSIS
AS DETERMINED BY A CHEMICAL TEST OF THE ACCUSED'S BREATH

NAME OF ACCUSED INSTRUMENT, TEST,	NAME OF COURT DFS
--	--------------------------

BREATH ANALYSIS

SAMPLE EXAMINED AND TEST CONDUCTED BY STANTON, HEATHER, M		AGENCY DFS Central Lab
DFS LICENSE NUMBER 18910	LICENSE EXPIRES 10/01/2016	DATE TEST CONDUCTED 12/12/2014
TEST EQUIPMENT NUMBER 010448		

RESULTS: TIME SAMPLE TAKEN 10:49 EST

SAMPLE'S ALCOHOL CONTENT 0.00 GRAMS PER 210 LITERS OF BREATH

ATTEST:

I CERTIFY THAT THE ABOVE IS AN ACCURATE RECORD OF THE TEST CONDUCTED; THAT THE TEST WAS CONDUCTED WITH THE TYPE OF EQUIPMENT AND IN ACCORDANCE WITH THE METHODS APPROVED BY THE DEPARTMENT OF FORENSIC SCIENCE; THAT THE TEST WAS CONDUCTED IN ACCORDANCE WITH THE DEPARTMENT'S SPECIFICATIONS; THAT PRIOR TO ADMINISTRATION OF THE TEST THE ACCUSED WAS ADVISED OF HIS RIGHT TO OBSERVE THE PROCESS AND SEE THE BLOOD ALCOHOL READING ON THE EQUIPMENT USED TO PERFORM THE BREATH TEST, AND THAT I POSSESS A VALID LICENSE TO CONDUCT SUCH TEST, GIVEN UNDER MY HAND THIS _____ DAY OF _____, 20_____.

BREATH TEST OPERATOR

I HAVE RECEIVED A COPY OF THIS CERTIFICATE OF ANALYSIS _____ SUBJECT'S SIGNATURE

SUBJECT REFUSED TO SIGN FOR COPY OF CERTIFICATE OF ANALYSIS _____ OPERATOR'S SIGNATURE



LMS
Thd

Test Results

Instrument Serial Number 010448

Test # 001529 Subject Test

Test Location 1 Department of
Test Date 12 Dec 2014

Test Location 2 Forensic Science
Test Time 10:43
Remote/Local Local

Test Location 3
System Check Passed

Operator's Last Name STANTON
Agency DFS Central Lab

Operator's First Name HEATHER

Operator's Middle Initial M

Card Serial Number 118910

License Number 18910

Effective Date 10/01/2014

Expiration Date 10/01/2016

Subject's Last Name INSTRUMENT

Subject's First Name TEST

Subject's Middle Initial

Subject's Date of Birth 00/00/0000

Subject's Sex Male

Driver's License Number

Driver's License Expiration 00/00/0000

Driver's License State

Court Name DFS

End Date 12 Dec 2014 End Time 10:50

Result Time 10:49

Result Date 12 Dec 2014 Result 0.00

Data Type DIAG

Sample Value Pass

Sample Time 10:43

Data Type BLK

Sample Value 0.000

Sample Time 10:44

Data Type CHK

Sample Value 0.100

Sample Time 10:45

Data Type BLK

Sample Value 0.000

Sample Time 10:46

Data Type SUBJ

Sample Value 0.000

Sample Time 10:47

Data Type BLK

Sample Value 0.000

Sample Time 10:48

Data Type SUBJ

Sample Value 0.000

Sample Time 10:49

Data Type BLK

Sample Value 0.000

Sample Time 10:50

Standard Type Dry Gas Std

Standard Value 0.100

Standard Lot Number AG419702-33

Standard Expiration Date 07/16/2016

Tank Pressure 438

Barometric Pressure 763 mmHg

Blow Sample Number 1 Blow Duration 3.44 sec

Blow Volume 1760 cc End-of-Blow Time 10:47

Blow Sample Number 2 Blow Duration 3.37 sec

Blow Volume 1760 cc End-of-Blow Time 10:49

Tamper Evident Stamp f6ac2ca2

Test Status Code 0

Test Status Success