

Department of Forensic Science  
Breath Alcohol Section  
Instrument Maintenance History  
2/5/2014 To 2/5/2015

MMB

---

**Instrument Serial Number:** 010588

---

As of 05-Feb-15

<b>Maintenance Date</b>	<b>Certification Date</b>	<b>Technician Initials</b>	<b>Remarks</b>
05-Feb-15	09-Sep-14	MMB (26417)	REPLACED KEYBOARD.
09-Sep-14	09-Sep-14	DBH (21250)	CERTIFIED.
24-Mar-14	24-Mar-14	HMS (18910)	CERTIFIED

mmB

# INTOX EC/IR II Quality Assurance Worksheet

Instrument Serial Number **010588** Worksheet Start Date **2/5/2015**  
 Location **Dinwiddie Co SO**  
 Address **14230 Sycamore Dr Dinwiddie VA 23841**  
 DFS Technician **Michelle Billeter** License No. **26417**

Laboratory  On-Site

Site Specification: No detrimental environmental conditions exist.

Instrument Barometer (mm HG) **751** Reference Barometer (mm HG) **753**

Reference Barometer(RB)Serial # **009111** RB Calibration Due **5/19/2015**


### Measurement Assurance Check

<b>Standard (sea level)</b>	<b>PA Target</b>	<b>minimum</b>	<b>maximum</b>	<b>Sample 1</b>
0.300		-0.003	0.003	<b>Sample 2</b>
<b>Precision</b>		<b>sample min</b>	<b>sample max</b>	<b>Sample 3</b>
0		0.000	0.000	

<b>Standard (sea level)</b>	<b>PA Target</b>	<b>minimum</b>	<b>maximum</b>	<b>Sample 1</b>
0.100		-0.003	0.003	<b>Sample 2</b>
<b>Precision</b>		<b>sample min</b>	<b>sample max</b>	<b>Sample 3</b>
0		0.000	0.000	

Dry gas standard Lot No. (with tank no.) **AG404902-05**

- Replaced dry gas standard (+O-ring)
  - Installed at Location
  - Removed to DFS-Central
- Supplies  
 Mouthpieces  
 Certificates of Analysis  
 Operator Worksheet  
 Other:



### Notes:

Replaced keyboard.

Instrument Serial Number

010588

Certification Date

Calibrated

Certified

Measurement Assurance Check

Instrument Test

Troubleshooting/Maintenance Worksheet Completed

(Successful completion denotes satisfactory condition of the item.)

<b>Standard (sea level)</b>	<b>PA Target</b>	<b>minimum</b>	<b>maximum</b>	<b>Sample 1</b>
0.250		0.000	0.000	<b>Sample 2</b>
<b>Precision</b>		<b>sample min</b>	<b>sample max</b>	<b>Sample 3</b>
0		0.000	0.000	

<b>Standard (sea level)</b>	<b>PA Target</b>	<b>minimum</b>	<b>maximum</b>	<b>Sample 1</b>
0.150		0.000	0.000	<b>Sample 2</b>
<b>Precision</b>		<b>sample min</b>	<b>sample max</b>	<b>Sample 3</b>
0		0.000	0.000	

<b>Standard (sea level)</b>	<b>PA Target</b>	<b>minimum</b>	<b>maximum</b>	<b>Sample 1</b>
0.080		-0.003	0.003	<b>Sample 2</b>
<b>Precision</b>		<b>sample min</b>	<b>sample max</b>	<b>Sample 3</b>
0		0.000	0.000	

<b>Standard (sea level)</b>	<b>PA Target</b>	<b>minimum</b>	<b>maximum</b>	<b>Sample 1</b>
0.020		-0.003	0.003	<b>Sample 2</b>
<b>Precision</b>		<b>sample min</b>	<b>sample max</b>	<b>Sample 3</b>
0		0.000	0.000	

All measurements are in g/210L

Estimation of Uncertainty of Measurement and traceability records are located within the Breath Alcohol Section

DFS Technician

*Michelle Ballin*

Date

2/5/15

Issuing Analyst

*Michelle Ballin*

Date

2/5/15

*JPB*

mmb



COMMONWEALTH OF VIRGINIA  
DEPARTMENT OF FORENSIC SCIENCE

CERTIFICATE OF BLOOD ALCOHOL ANALYSIS  
AS DETERMINED BY A CHEMICAL TEST OF THE ACCUSED'S BREATH

NAME OF ACCUSED  INSTRUMENT, TEST,	NAME OF COURT  DFS
--	--------------------------

BREATH ANALYSIS

SAMPLE EXAMINED AND TEST CONDUCTED BY  BILLETER, MICHELLE, M		AGENCY  DFS Central Lab
DFS LICENSE NUMBER  26417	LICENSE EXPIRES  10/01/2016	DATE TEST CONDUCTED  02/05/2015
TEST EQUIPMENT NUMBER  010588		

RESULTS: TIME SAMPLE TAKEN 08:25 EST

SAMPLE'S ALCOHOL CONTENT 0.00 GRAMS PER 210 LITERS OF BREATH

ATTEST:

I CERTIFY THAT THE ABOVE IS AN ACCURATE RECORD OF THE TEST CONDUCTED; THAT THE TEST WAS CONDUCTED WITH THE TYPE OF EQUIPMENT AND IN ACCORDANCE WITH THE METHODS APPROVED BY THE DEPARTMENT OF FORENSIC SCIENCE; THAT THE TEST WAS CONDUCTED IN ACCORDANCE WITH THE DEPARTMENT'S SPECIFICATIONS; THAT PRIOR TO ADMINISTRATION OF THE TEST THE ACCUSED WAS ADVISED OF HIS RIGHT TO OBSERVE THE PROCESS AND SEE THE BLOOD ALCOHOL READING ON THE EQUIPMENT USED TO PERFORM THE BREATH TEST, AND THAT I POSSESS A VALID LICENSE TO CONDUCT SUCH TEST, GIVEN UNDER MY HAND THIS \_\_\_\_\_ DAY OF \_\_\_\_\_, 20\_\_\_\_\_.

BREATH TEST OPERATOR
----------------------

I HAVE RECEIVED A COPY OF THIS CERTIFICATE OF ANALYSIS \_\_\_\_\_ SUBJECT'S SIGNATURE

SUBJECT REFUSED TO SIGN FOR COPY OF CERTIFICATE OF ANALYSIS \_\_\_\_\_ OPERATOR'S SIGNATURE

FOR COURT



# IntoxNet MIS Report

Report Generated 05 Feb 2015 at 10:46

## Test Results

**Instrument Serial Number 010588**

**Test # 001726**      Subject Test

Test Location 1    Department of  
Test Date    05 Feb 2015

Test Location 2    Forensic Science  
Test Time    08:19  
Remote/Local    Local

Test Location 3  
System Check    Passed

Operator's Last Name    BILLETTER  
Agency    DFS Central Lab

Operator's First Name    MICHELLE

Operator's Middle Initial    M

Card Serial Number    126417

Effective Date    10/01/2014

Expiration Date    10/01/2016

Subject's Last Name    INSTRUMENT

Subject's First Name    TEST

Subject's Middle Initial

Subject's Date of Birth    00/00/0000

Subject's Sex    Male

Driver's License Number

Driver's License Expiration    00/00/0000

Driver's License State

Court Name    DFS

End Date    05 Feb 2015      End Time    08:27

Result Time    08:25

Result Date    05 Feb 2015      Result    0.00

Data Type    DIAG

Sample Value    Pass

Sample Time    08:19

Data Type    BLK

Sample Value    0.000

Sample Time    08:20

Data Type    CHK

Sample Value    0.101

Sample Time    08:21

Data Type    BLK

Sample Value    0.000

Sample Time    08:22

Data Type    SUBJ

Sample Value    0.000

Sample Time    08:23

Data Type    BLK

Sample Value    0.000

Sample Time    08:24

Data Type    SUBJ

Sample Value    0.000

Sample Time    08:25

Data Type    BLK

Sample Value    0.000

Sample Time    08:26

Standard Type    Dry Gas Std

Standard Value    0.099

Standard Lot Number    AG404902-05

Standard Expiration Date    02/18/2016

Tank Pressure    409

Barometric Pressure    751 mmHg

Blow Sample Number    1      Blow Duration    4.03 sec

Blow Volume    1853 cc      End-of-Blow Time    08:23

Blow Sample Number    2      Blow Duration    3.83 sec

Blow Volume    1808 cc      End-of-Blow Time    08:25

Tamper Evident Stamp    64fa6aae

Test Status Code    0

Test Status    Success

Department of Forensic Science  
Breath Alcohol Section  
Trouble Call Log  
2/5/2014 To 2/5/2015

TLN  
mmB

Instrument Serial Number: 010588

05-Feb-15

---

**Date:** 04-Feb-15    **Time:** 3:15 PM    **Remote:** True    **Tech:** TLN (25387)    **Location:** Dinwiddie County Sheriff's Office

**Indication:** KEYBOARD / CARD READER NOT RESPONDING

**Problem:** KEYBOARD / CARD READER NOT RESPONDING

**Technician Response:** SITE VISIT CONDUCTED. ISSUE RESOLVED.

**Resolution:** NO FURTHER ACTION REQUIRED

---