

Department of Forensic Science
Breath Alcohol Section
Instrument Location History
1/1/1990 To 3/20/2015

DBH
mms

Instrument Serial Number: 010465

As of 20-Mar-15

Start Date	End Date	Agency Name
17-Mar-15		Fluvanna County Sheriff's Office
24-Feb-15	17-Mar-15	DFS Central Lab
18-Sep-14	24-Feb-15	Meherrin River Regional Jail
11-Sep-14	18-Sep-14	DFS Central Lab
08-Apr-14	11-Sep-14	Galax PD
18-Mar-14	08-Apr-14	DFS Central Lab
24-May-12	18-Mar-14	Virginia Beach PD
24-May-12	24-May-12	DFS Central Lab
23-Feb-10	24-May-12	Virginia Beach PD
23-Sep-09	23-Feb-10	DFS Central Lab
18-Nov-08	23-Sep-09	Charlottesville PD
13-May-08	18-Nov-08	DFS Central Lab

DBH

mmB

INTOX EC/IR II Quality Assurance Worksheet

Instrument Serial Number: 010465 Worksheet Start Date: 3/17/2015

Location: Fluvanna CO SO

Address: 160 Commons Blvd Palmyra VA 22963

DFS Technician: Donald Hall License No.: 21250

Laboratory On-Site

Site Specification: No detrimental environmental conditions exist.

Instrument Barometer (mm HG): 746 Reference Barometer (mm HG): 746

Reference Barometer(RB)Serial #: 009109 RB Calibration Due: 7/28/2015

Measurement Assurance Check

Standard (sea level)	PA Target	minimum	maximum	Sample 1	Sample 2	Sample 3
0.300	0.295	0.286	0.303	0.302	0.302	0.301
Precision		sample min	sample max			
0.001		0.301	0.302			

Standard (sea level)	PA Target	minimum	maximum	Sample 1	Sample 2	Sample 3
0.100	0.098	0.095	0.101	0.099	0.099	0.099
Precision		sample min	sample max			
0		0.099	0.099			

Dry gas standard Lot No. (with tank no.): AG419702-51

- Replaced dry gas standard (+O-ring)
 - Installed at Location
 - Removed to DFS-Central
- Supplies: 100
- Mouthpieces
- Certificates of Analysis
- Operator Worksheet
- Other:

Notes:

[Redacted Notes Area]

Instrument Serial Number

010465

Certification Date

Calibrated

Certified

Measurement Assurance Check

Instrument Test

Troubleshooting/Maintenance Worksheet Completed

(Successful completion denotes satisfactory condition of the item.)

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.250		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.150		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.080		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.020		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

All measurements are in g/210L

Estimation of Uncertainty of Measurement and traceability records are located within the Breath Alcohol Section

DFS Technician

Donald B Hill

Date

3/20/15

Issuing Analyst

Michelle Buller
TLW

Date

3/20/15

DBH
mmB

Intox EC/IR-II: Accuracy Check

Department of Forensic Science

Serial Number: 010465 Test Number: 2722

Test Date: 03/17/2015 Test Time: 14:06 EDT

Dry Gas Target: 0.295

Lot Number: AG428003-08 Exp Date: 10/07/2016

Tank Pressure: 419 psi Barometric Pressure: 746 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	14:07
CHK	0.302	14:08
BLK	0.000	14:10
CHK	0.302	14:10
BLK	0.000	14:12
CHK	0.301	14:13

Test Status: *Success*

Calibration CRC: FC317593

Intox EC/IR-11: Accuracy Check

DB
mmB

Department of Forensic Science

Serial Number: 010465 Test Number: 2723

Test Date: 03/17/2015 Test Time: 14:17 EDT

Dry Gas Target: 0.098

Lot Number: AG419702-51 Exp Date: 07/16/2016

Tank Pressure: 541 psi Barometric Pressure: 746 mmHg

System Check: Passed

Test	q/210L	Time
BLK	0.000	14:17
CHK	0.099	14:18
BLK	0.000	14:20
CHK	0.099	14:20
BLK	0.000	14:22
CHK	0.099	14:23

Test Status: Success

Calibration CRC: FC317593

DBH
mmB



COMMONWEALTH OF VIRGINIA
DEPARTMENT OF FORENSIC SCIENCE

CERTIFICATE OF BLOOD ALCOHOL ANALYSIS
AS DETERMINED BY A CHEMICAL TEST OF THE ACCUSED'S BREATH

NAME OF ACCUSED INSTRUMENT, TEST,	NAME OF COURT DFS
---	-----------------------------

BREATH ANALYSIS

SAMPLE EXAMINED AND TEST CONDUCTED BY HALL, DONALD, B		AGENCY DFS Central Lab
DFS LICENSE NUMBER 21250	LICENSE EXPIRES 10/01/2016	DATE TEST CONDUCTED 03/17/2015
TEST EQUIPMENT NUMBER 010465		

14:32 EDT

RESULTS: TIME SAMPLE TAKEN _____
0.00

SAMPLE'S ALCOHOL CONTENT _____ GRAMS PER 210 LITERS OF BREATH

ATTEST:

I CERTIFY THAT THE ABOVE IS AN ACCURATE RECORD OF THE TEST CONDUCTED; THAT THE TEST WAS CONDUCTED WITH THE TYPE OF EQUIPMENT AND IN ACCORDANCE WITH THE METHODS APPROVED BY THE DEPARTMENT OF FORENSIC SCIENCE; THAT THE TEST WAS CONDUCTED IN ACCORDANCE WITH THE DEPARTMENT'S SPECIFICATIONS; THAT PRIOR TO ADMINISTRATION OF THE TEST THE ACCUSED WAS ADVISED OF HIS RIGHT TO OBSERVE THE PROCESS AND SEE THE BLOOD ALCOHOL READING ON THE EQUIPMENT USED TO PERFORM THE BREATH TEST, AND THAT I POSSESS A VALID LICENSE TO CONDUCT SUCH TEST, GIVEN UNDER MY HAND THIS _____ DAY OF _____, 20_____.

BREATH TEST OPERATOR

I HAVE RECEIVED A COPY OF THIS CERTIFICATE OF ANALYSIS _____
SUBJECT'S SIGNATURE

SUBJECT REFUSED TO SIGN FOR COPY OF CERTIFICATE OF ANALYSIS _____
OPERATOR'S SIGNATURE



IntoxNet MIS Report
Report Generated 19 Mar 2015 at 15:50

DBH
mmB

Test Results

Instrument Serial Number 010465

Test # 002724 Subject Test

Test Location 1 Department of
Test Date 17 Mar 2015

Test Location 2 Forensic Science
Test Time 14:26
Remote/Local Local

Test Location 3
System Check Passed

Operator's Last Name HALL
Agency DFS Central Lab

Operator's First Name DONALD

Operator's Middle Initial B

Card Serial Number 121250

License Number 21250

Effective Date 10/01/2014

Expiration Date 10/01/2016

Subject's Last Name INSTRUMENT

Subject's First Name TEST

Subject's Middle Initial

Subject's Date of Birth 00/00/0000

Subject's Sex Male

Driver's License Number

Driver's License Expiration 00/00/0000

Driver's License State

Court Name DFS

End Date 17 Mar 2015 End Time 14:33

Result Time 14:32

Result Date 17 Mar 2015 Result 0.00

Data Type DIAG

Sample Value Pass

Sample Time 14:26

Data Type BLK

Sample Value 0.000

Sample Time 14:27

Data Type CHK

Sample Value 0.098

Sample Time 14:28

Data Type BLK

Sample Value 0.000

Sample Time 14:29

Data Type SUBJ

Sample Value 0.000

Sample Time 14:30

Data Type BLK

Sample Value 0.000

Sample Time 14:31

Data Type SUBJ

Sample Value 0.000

Sample Time 14:32

Data Type BLK

Sample Value 0.000

Sample Time 14:33

Standard Type Dry Gas Std

Standard Value 0.098

Standard Lot Number AG419702-51

Standard Expiration Date 07/16/2016

Tank Pressure 531

Barometric Pressure 746 mmHg

Blow Sample Number 1 Blow Duration 3.66 sec

Blow Volume 1999 cc End-of-Blow Time 14:30

Blow Sample Number 2 Blow Duration 3.94 sec

Blow Volume 1970 cc End-of-Blow Time 14:32

Tamper Evident Stamp abd5e1cb

Test Status Code 0

Test Status Success