

DBH
pd

**Department of Forensic Science
Breath Alcohol Section
Instrument Location History
1/1/1990 To 3/27/2015**

Instrument Serial Number: 010503

As of 27-Mar-15

Start Date	End Date	Agency Name
26-Mar-15		DFS Central Lab
15-Oct-14	26-Mar-15	Rockingham County SO
08-Oct-14	15-Oct-14	DFS Central Lab
08-May-14	08-Oct-14	Madison County Sheriff's Office
30-Apr-14	08-May-14	DFS Central Lab
13-Dec-12	30-Apr-14	Charlottesville PD
12-Dec-12	13-Dec-12	DFS Central Lab
18-Nov-08	12-Dec-12	Charlottesville PD
19-May-08	18-Nov-08	DFS Central Lab

DB
JW

INTOX EC/IR II Quality Assurance Worksheet

Instrument Serial Number **010503** Worksheet Start Date **3/26/2015**
 Location **Rockingham County SO**
 Address **25 S Liberty St Harrisonburg VA 22801**
 DFS Technician **Donald Hall** License No. **21250**

Laboratory On-Site

Site Specification: No detrimental environmental conditions exist.


Instrument Barometer (mm HG) **719** Reference Barometer (mm HG) **722**
 Reference Barometer(RB)Serial # **009109** RB Calibration Due **7/28/2015**

Measurement Assurance Check

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.289
0.300	0.283	0.274	0.291	Sample 2	0.288
Precision		sample min	sample max	Sample 3	0.288
0.001		0.288	0.289		

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.094
0.100	0.095	0.092	0.098	Sample 2	0.094
Precision		sample min	sample max	Sample 3	0.095
0.001		0.094	0.095		

Dry gas standard Lot No. (with tank no.) **AG419702-48**

Replaced dry gas standard (+O-ring) Supplies
 Installed at Location Mouthpieces
 Removed to DFS-Central Certificates of Analysis
 Operator Worksheet
 Other: 

Notes:

[Redacted Notes Area]

Instrument Serial Number

010503

Certification Date

Calibrated

Certified

Measurement Assurance Check

Instrument Test

Troubleshooting/Maintenance Worksheet Completed

(Successful completion denotes satisfactory condition of the item.)

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.250		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.150		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.080		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.020		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

All measurements are in g/210L

Estimation of Uncertainty of Measurement and traceability records are located within the Breath Alcohol Section

DFS Technician

Daniel B Hall

Date

3/27/15

Issuing Analyst

Jed

Date

3/27/15

mmb

Intox EC/IR-II: Accuracy Check

DE
FD

Department of Forensic Science

Serial Number: 010503 Test Number: 1386

Test Date: 03/26/2015 Test Time: 11:15 EDT

Dry Gas Target: 0.095

Lot Number: AG419702-48 Exp Date: 07/16/2016

Tank Pressure: 683 psi Barometric Pressure: 718 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	11:16
CHK	0.094	11:16
BLK	0.000	11:18
CHK	0.094	11:19
BLK	0.000	11:20
CHK	0.095	11:21

Test Status: *Success*

Calibration CRC: 350E2E5B

Intox EC/IR-II: Accuracy Check

Department of Forensic Science

Serial Number: 010503 Test Number: 1387

Test Date: 03/26/2015 Test Time: 11:26 EDT

Dry Gas Target: 0.283

Lot Number: AG428003-05 Exp Date: 10/07/2016

Tank Pressure: 546 psi Barometric Pressure: 718 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	11:27
CHK	0.289	11:27
BLK	0.000	11:29
CHK	0.288	11:30
BLK	0.000	11:32
CHK	0.288	11:33

Test Status: *Success*

Calibration CRC: 350E2E5B

1
JL

DB
Jrd



COMMONWEALTH OF VIRGINIA
DEPARTMENT OF FORENSIC SCIENCE

CERTIFICATE OF BLOOD ALCOHOL ANALYSIS
AS DETERMINED BY A CHEMICAL TEST OF THE ACCUSED'S BREATH

NAME OF ACCUSED INSTRUMENT, TEST,	NAME OF COURT DFS
---	-----------------------------

BREATH ANALYSIS

SAMPLE EXAMINED AND TEST CONDUCTED BY HALL, DONALD, B		AGENCY DFS Central Lab
DFS LICENSE NUMBER 21250	LICENSE EXPIRES 10/01/2016	DATE TEST CONDUCTED 03/26/2015
TEST EQUIPMENT NUMBER 010503		

RESULTS: TIME SAMPLE TAKEN 11:12 EDT

SAMPLE'S ALCOHOL CONTENT 0.00 GRAMS PER 210 LITERS OF BREATH

ATTEST:

I CERTIFY THAT THE ABOVE IS AN ACCURATE RECORD OF THE TEST CONDUCTED; THAT THE TEST WAS CONDUCTED WITH THE TYPE OF EQUIPMENT AND IN ACCORDANCE WITH THE METHODS APPROVED BY THE DEPARTMENT OF FORENSIC SCIENCE; THAT THE TEST WAS CONDUCTED IN ACCORDANCE WITH THE DEPARTMENT'S SPECIFICATIONS; THAT PRIOR TO ADMINISTRATION OF THE TEST THE ACCUSED WAS ADVISED OF HIS RIGHT TO OBSERVE THE PROCESS AND SEE THE BLOOD ALCOHOL READING ON THE EQUIPMENT USED TO PERFORM THE BREATH TEST, AND THAT I POSSESS A VALID LICENSE TO CONDUCT SUCH TEST, GIVEN UNDER MY HAND THIS _____ DAY OF _____, 20_____.

BREATH TEST OPERATOR

I HAVE RECEIVED A COPY OF THIS CERTIFICATE OF ANALYSIS _____ SUBJECT'S SIGNATURE

SUBJECT REFUSED TO SIGN FOR COPY OF CERTIFICATE OF ANALYSIS _____ OPERATOR'S SIGNATURE



IntoxNet MIS Report
Report Generated 26 Mar 2015 at 11:40

LMS *DBH*
ind

Test Results

Instrument Serial Number 010503

Test # 001385 Subject Test

Test Location 1 Department of
Test Date 26 Mar 2015

Test Location 2 Forensic Science
Test Time 11:06
Remote/Local Local

Test Location 3
System Check Passed

Operator's Last Name HALL
Agency DFS Central Lab

Operator's First Name DONALD

Operator's Middle Initial B

Card Serial Number 121250

License Number 21250

Effective Date 10/01/2014

Expiration Date 10/01/2016

Subject's Last Name INSTRUMENT

Subject's First Name TEST

Subject's Middle Initial

Subject's Date of Birth 00/00/0000

Subject's Sex Male

Driver's License Number

Driver's License Expiration 00/00/0000

Driver's License State

Court Name DFS

End Date 26 Mar 2015 End Time 11:14

Result Time 11:12

Result Date 26 Mar 2015 Result 0.00

Data Type DIAG

Sample Value Pass

Sample Time 11:06

Data Type BLK

Sample Value 0.000

Sample Time 11:07

Data Type CHK

Sample Value 0.095

Sample Time 11:08

Data Type BLK

Sample Value 0.000

Sample Time 11:09

Data Type SUBJ

Sample Value 0.000

Sample Time 11:10

Data Type BLK

Sample Value 0.000

Sample Time 11:11

Data Type SUBJ

Sample Value 0.000

Sample Time 11:12

Data Type BLK

Sample Value 0.000

Sample Time 11:13

Standard Type Dry Gas Std

Standard Value 0.095

Standard Lot Number AG419702-48

Standard Expiration Date 07/16/2016

Tank Pressure 688

Barometric Pressure 719 mmHg

Blow Sample Number 1 Blow Duration 3.55 sec

Blow Volume 1966 cc End-of-Blow Time 11:10

Blow Sample Number 2 Blow Duration 4.04 sec

Blow Volume 2174 cc End-of-Blow Time 11:12

Tamper Evident Stamp c4581f6e

Test Status Code 0

Test Status Success