DBH
Sled

Department of Forensic Science
Breath Alcohol Section
Instrument Location History
1/1/1990 To 3/27/2015

Instrument Serial Number: 010606 **As of 27-Mar-15**

Start Date	End Date	Agency Name
26-Mar-15		DFS Central Lab
09-Oct-14	26-Mar-15	Page County SO
01-Oct-14	09-Oct-14	DFS Central Lab
11-Feb-11	01-Oct-14	Dickenson County Sheriff's Office
22-Jan-09	11-Feb-11	DFS Central Lab
10-Dec-08	22-Jan-09	Accomack County SO
15-Aug-08	10-Dec-08	DFS Central Lab

DBH

INTOX EC/IR II
Quality Assurance Worksheet

Instrument Serial Number	010606	Worksheet Start Date	3/26/2015
Location	Page CO SO		
Address	108 S Court St Luray VA 22835		
DFS Technician	Donald Hall	License No.	21250

☐ Laboratory ☑ On-Site

Site Specification: No detrimental environmental conditions exist. ☑

Instrument Barometer (mm HG)	728	Reference Barometer (mm HG)	731
Reference Barometer(RB)Serial #	009109	RB Calibration Due	7/28/2015

Measurement Assurance Check

Standard (sea level)	PA Target	minimum	maximum		
0.300	0.287	0.278	0.295	Sample 1	0.294
Precision		sample min	sample max	Sample 2	0.294
0		0.294	0.294	Sample 3	0.294

Standard (sea level)	PA Target	minimum	maximum		
0.100	0.096	0.093	0.099	Sample 1	0.096
Precision		sample min	sample max	Sample 2	0.097
0.001		0.096	0.097	Sample 3	0.097

Dry gas standard Lot No. (with tank no.) AG320502-23

☐ Replaced dry gas standard (+O-ring)

☐ Installed at Location

☑ Removed to DFS-Central

Supplies
Mouthpieces
Certificates of Analysis
Operator Worksheet
Other:

Notes:

Instrument Serial Number 010606 Certification Date

☐ Calibrated ☐ Certified ☑ Measurement Assurance Check

☑ Instrument Test ☐ Troubleshooting/Maintenance Worksheet Completed
(Successful completion denotes satisfactory condition of the item.)

Standard (sea level)	PA Target	minimum	maximum	
0.250		0.000	0.000	Sample 1
Precision		sample min	sample max	Sample 2
0		0.000	0.000	Sample 3

Standard (sea level)	PA Target	minimum	maximum	
0.150		0.000	0.000	Sample 1
Precision		sample min	sample max	Sample 2
0		0.000	0.000	Sample 3

Standard (sea level)	PA Target	minimum	maximum	
0.080		-0.003	0.003	Sample 1
Precision		sample min	sample max	Sample 2
0		0.000	0.000	Sample 3

Standard (sea level)	PA Target	minimum	maximum	
0.020		-0.003	0.003	Sample 1
Precision		sample min	sample max	Sample 2
0		0.000	0.000	Sample 3

All measurements are in g/210L
Estimation of Uncertainty of Measurement and traceability records are located within the Breath Alcohol Section

DFS Technician Donald B Hall Date 3/27/15

Issuing Analyst Date 3/27/15

DBH
hd

COMMONWEALTH OF VIRGINIA
DEPARTMENT OF FORENSIC SCIENCE

CERTIFICATE OF BLOOD ALCOHOL ANALYSIS

AS DETERMINED BY A CHEMICAL TEST OF THE ACCUSED'S BREATH

NAME OF ACCUSED	NAME OF COURT
INSTRUMENT, TEST,	DFS

BREATH ANALYSIS

SAMPLE EXAMINED AND TEST CONDUCTED BY		AGENCY
HALL, DONALD, B		DFS Central Lab
DFS LICENSE NUMBER	LICENSE EXPIRES	DATE TEST CONDUCTED
21250	10/01/2016	03/26/2015
TEST EQUIPMENT NUMBER		
010606		

RESULTS: TIME SAMPLE TAKEN 13:29 EDT

SAMPLE'S ALCOHOL CONTENT 0.00 GRAMS PER 210 LITERS OF BREATH

ATTEST:

I CERTIFY THAT THE ABOVE IS AN ACCURATE RECORD OF THE TEST CONDUCTED; THAT THE TEST WAS CONDUCTED WITH THE TYPE OF EQUIPMENT AND IN ACCORDANCE WITH THE METHODS APPROVED BY THE DEPARTMENT OF FORENSIC SCIENCE; THAT THE TEST WAS CONDUCTED IN ACCORDANCE WITH THE DEPARTMENT'S SPECIFICATIONS; THAT PRIOR TO ADMINISTRATION OF THE TEST THE ACCUSED WAS ADVISED OF HIS RIGHT TO OBSERVE THE PROCESS AND SEE THE BLOOD ALCOHOL READING ON THE EQUIPMENT USED TO PERFORM THE BREATH TEST, AND THAT I POSSESS A VALID LICENSE TO CONDUCT SUCH TEST, GIVEN UNDER MY HAND THIS _____ DAY OF _____, 20 _____.

BREATH TEST OPERATOR

☐ I HAVE RECEIVED A COPY OF THIS CERTIFICATE OF ANALYSIS _____
SUBJECT'S SIGNATURE

☐ SUBJECT REFUSED TO SIGN FOR COPY OF CERTIFICATE OF ANALYSIS _____
OPERATOR'S SIGNATURE

FOR COURT

DSH

HMS

ned

IntoxNet MIS Report

Report Generated 26 Mar 2015 at 13:55

Test Results

Instrument Serial Number 010606

Test # 000363 Subject Test

Test Location 1 Department of **Test Location 2** Forensic Science **Test Location 3**

Test Date 26 Mar 2015 **Test Time** 13:23 **System Check** Passed

Remote/Local Local

Operator's Last Name HALL **Operator's First Name** DONALD **Operator's Middle Initial** B

Agency DFS Central Lab **License Number** 21250

Card Serial Number 121250 **Effective Date** 10/01/2014 **Expiration Date** 10/01/2016

Subject's Last Name INSTRUMENT **Subject's First Name** TEST

Subject's Middle Initial **Subject's Date of Birth** 00/00/0000 **Subject's Sex** Male

Driver's License Number **Driver's License Expiration** 00/00/0000

Driver's License State **Court Name** DFS

End Date 26 Mar 2015 **End Time** 13:31 **Result Time** 13:29 **Result Date** 26 Mar 2015 **Result** 0.00

Data Type	DIAG	**Sample Value**	Pass	**Sample Time**	13:23
Data Type	BLK	**Sample Value**	0.000	**Sample Time**	13:24
Data Type	CHK	**Sample Value**	0.097	**Sample Time**	13:25
Data Type	BLK	**Sample Value**	0.000	**Sample Time**	13:26
Data Type	SUBJ	**Sample Value**	0.000	**Sample Time**	13:27
Data Type	BLK	**Sample Value**	0.000	**Sample Time**	13:28
Data Type	SUBJ	**Sample Value**	0.000	**Sample Time**	13:29
Data Type	BLK	**Sample Value**	0.000	**Sample Time**	13:30

Standard Type Dry Gas Std **Standard Value** 0.096

Standard Lot Number AG320502-23 **Standard Expiration Date** 07/24/2015

Tank Pressure 703 **Barometric Pressure** 728 mmHg

Blow Sample Number	1	**Blow Duration**	3.81 sec	**Blow Volume**	1999 cc	**End-of-Blow Time**	13:27
Blow Sample Number	2	**Blow Duration**	4.04 sec	**Blow Volume**	2140 cc	**End-of-Blow Time**	13:29

Tamper Evident Stamp 0dff5fcd **Test Status Code** 0

Test Status Success

DSH
Sel

Intox EC/IR-II: Accuracy Check

Department of Forensic Science
Serial Number: *010606* Test Number: *364*
Test Date: *03/26/2015* Test Time: *13:32 EDT*
Dry Gas Target: *0.096*
Lot Number: *AG320502-23* Exp Date: *07/24/2015*
Tank Pressure: *698 psi* Barometric Pressure: *727 mmHg*
System Check: *Passed*

Test	g/210L	Time
BLK	0.000	13:33
CHK	0.096	13:33
BLK	0.000	13:35
CHK	0.097	13:36
BLK	0.000	13:37
CHK	0.097	13:38

Test Status: *Success*

Calibration CRC: 35EC39BD

DBH
gud

Intox EC/IR-II: Accuracy Check

Department of Forensic Science
Serial Number: *010606* Test Number: *365*
Test Date: *03/26/2015* Test Time: *13:41 EDT*
Dry Gas Target: *0.287*
Lot Number: *AG428003-05* Exp Date: *10/07/2016*
Tank Pressure: *526 psi* Barometric Pressure: *727 mmHg*
System Check: *Passed*

Test	g/210L	Time
BLK	0.000	13:42
CHK	0.294	13:42
BLK	0.000	13:44
CHK	0.294	13:45
BLK	0.000	13:47
CHK	0.294	13:48

Test Status: *Success*

Calibration CRC: 35EC39BD