

Department of Forensic Science  
Breath Alcohol Section  
Instrument Location History  
1/1/1990 To 4/17/2015

HMS  
TLN

Instrument Serial Number: 010446

As of 17-Apr-15

Start Date	End Date	Agency Name
17-Apr-15		DFS Central Lab
03-Mar-15	17-Apr-15	Fairfax County PD
30-Jun-14	03-Mar-15	DFS Central Lab
30-Apr-14	30-Jun-14	Charlottesville PD
18-Apr-14	30-Apr-14	DFS Central Lab
15-Nov-12	18-Apr-14	Meherrin River Regional Jail
19-Dec-11	15-Nov-12	DFS Central Lab
09-Dec-11	19-Dec-11	Chesterfield County PD
08-Dec-11	09-Dec-11	DFS Central Lab
25-Jul-11	08-Dec-11	Chesterfield County PD
15-Oct-10	25-Jul-11	DFS Central Lab
09-Oct-09	15-Oct-10	Chesterfield County SO
30-Apr-09	09-Oct-09	DFS Central Lab
13-Nov-08	30-Apr-09	Nelson County Sheriff's Office
19-May-08	13-Nov-08	DFS Central Lab

*KMS*  
*TKD*

# INTOX EC/IR II Quality Assurance Worksheet

Instrument Serial Number: 010446      Worksheet Start Date: 4/17/2015  
Location: Fairfax County Mount Vernon  
Address: 2511 Parkers Lane, Alexandria, VA 22306  
DFS Technician: Heather Stanton      License No.: 18910

Laboratory       On-Site

Site Specification: No detrimental environmental conditions exist.

Instrument Barometer (mm HG): 761      Reference Barometer (mm HG): 763

Reference Barometer(RB)Serial #: 009111      RB Calibration Due: 5/19/2015

### Measurement Assurance Check


Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.300		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.100		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Dry gas standard Lot No. (with tank no.)      AG332301-02

- Replaced dry gas standard (+O-ring)
- Installed at Location
- Removed to DFS-Central

- Supplies
- Mouthpieces
- Certificates of Analysis
- Operator Worksheet
- Other:



### Notes:

Unable to perform instrument test and measurement assurance checks due to instrument being disabled.

Instrument Serial Number

010446

Certification Date

Calibrated

Certified

Measurement Assurance Check

Instrument Test

Troubleshooting/Maintenance Worksheet Completed

(Successful completion denotes satisfactory condition of the item.)

<b>Standard (sea level)</b>	<b>PA Target</b>	<b>minimum</b>	<b>maximum</b>	<b>Sample 1</b>
0.250		0.000	0.000	<b>Sample 2</b>
<b>Precision</b>		<b>sample min</b>	<b>sample max</b>	<b>Sample 3</b>
0		0.000	0.000	

<b>Standard (sea level)</b>	<b>PA Target</b>	<b>minimum</b>	<b>maximum</b>	<b>Sample 1</b>
0.150		0.000	0.000	<b>Sample 2</b>
<b>Precision</b>		<b>sample min</b>	<b>sample max</b>	<b>Sample 3</b>
0		0.000	0.000	

<b>Standard (sea level)</b>	<b>PA Target</b>	<b>minimum</b>	<b>maximum</b>	<b>Sample 1</b>
0.080		-0.003	0.003	<b>Sample 2</b>
<b>Precision</b>		<b>sample min</b>	<b>sample max</b>	<b>Sample 3</b>
0		0.000	0.000	

<b>Standard (sea level)</b>	<b>PA Target</b>	<b>minimum</b>	<b>maximum</b>	<b>Sample 1</b>
0.020		-0.003	0.003	<b>Sample 2</b>
<b>Precision</b>		<b>sample min</b>	<b>sample max</b>	<b>Sample 3</b>
0		0.000	0.000	

All measurements are in g/210L

Estimation of Uncertainty of Measurement and traceability records are located within the Breath Alcohol Section

DFS Technician

*Heath W. Stach*

Date

*4/17/15*

Issuing Analyst

*[Signature]*

Date

*4/17/15*

*JPD*

Department of Forensic Science  
Breath Alcohol Section  
Trouble Call Log  
4/16/2015 To 4/16/2015

MMB  
KMS  
TLN

Instrument Serial Number: 010446

16-Apr-15

---

Date: 16-Apr-15 Time: 2:03 PM Remote: True Tech: MMB (26417) Location: Fairfax County PD

Indication: DISABLED

Problem: DISABLED. RECEIVED TWO OUT OF TOLERANCE MESSAGES.

Technician Response: SITE VISIT CONDUCTED.

Resolution REMOVED TO DFS-Central

---