

Department of Forensic Science
Breath Alcohol Section
Instrument Location History
1/1/1990 To 4/23/2015

AMS
mmb

Instrument Serial Number: 010470

As of 23-Apr-15

Start Date	End Date	Agency Name
22-Apr-15		DFS Central Lab
18-Nov-14	22-Apr-15	Chincoteague PD
28-Oct-14	18-Nov-14	DFS Central Lab
29-May-14	28-Oct-14	Appomattox County SO
20-May-14	29-May-14	DFS Central Lab
22-Apr-09	20-May-14	Rockbridge County SO
02-Apr-09	22-Apr-09	DFS Central Lab
29-Oct-08	02-Apr-09	Grayson County Sheriff's Office
13-May-08	29-Oct-08	DFS Central Lab

INTOX EC/IR II Quality Assurance Worksheet

LMS
mmB

Instrument Serial Number: 010470 Worksheet Start Date: 4/22/2015

Location: Chincoteague Police Department

Address: 6150 Community Dr., Chincoteague, VA 23336

DFS Technician: Heather Stanton License No.: 18910

Laboratory On-Site

Site Specification: No detrimental environmental conditions exist.

Instrument Barometer (mm HG): 760 Reference Barometer (mm HG): 758

Reference Barometer(RB)Serial #: 009111 RB Calibration Due: 5/19/2015

Measurement Assurance Check

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.306
0.300	0.300	0.291	0.309	Sample 2	0.305
Precision		sample min	sample max	Sample 3	0.305
0.001		0.305	0.306		

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.102
0.100	0.100	0.097	0.103	Sample 2	0.101
Precision		sample min	sample max	Sample 3	0.101
0.001		0.101	0.102		


Dry gas standard Lot No. (with tank no.) AG406301-21

Replaced dry gas standard (+O-ring)

Installed at Location

Removed to DFS-Central

Supplies
Mouthpieces
Certificates of Analysis
Operator Worksheet
Other:



Notes:

Instrument Serial Number 010470

Certification Date

Calibrated

Certified

Measurement Assurance Check

Instrument Test

Troubleshooting/Maintenance Worksheet Completed

(Successful completion denotes satisfactory condition of the item.)

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.250		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.150		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.080		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.020		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

All measurements are in g/210L

Estimation of Uncertainty of Measurement and traceability records are located within the Breath Alcohol Section

DFS Technician

Heath W. Stark

Date

4/23/15

Issuing Analyst

Michelle Bellini

Date

4/24/15

jit

HMS
mmB



COMMONWEALTH OF VIRGINIA
DEPARTMENT OF FORENSIC SCIENCE

CERTIFICATE OF BLOOD ALCOHOL ANALYSIS
AS DETERMINED BY A CHEMICAL TEST OF THE ACCUSED'S BREATH

NAME OF ACCUSED INSTRUMENT, TEST,	NAME OF COURT DFS
---	-----------------------------

BREATH ANALYSIS

SAMPLE EXAMINED AND TEST CONDUCTED BY STANTON, HEATHER, M		AGENCY DFS Central Lab
DFS LICENSE NUMBER 18910	LICENSE EXPIRES 10/01/2016	DATE TEST CONDUCTED 04/22/2015
TEST EQUIPMENT NUMBER 010470		

RESULTS: TIME SAMPLE TAKEN **07:47 EDT**

SAMPLE'S ALCOHOL CONTENT **0.00** GRAMS PER 210 LITERS OF BREATH

ATTEST:

I CERTIFY THAT THE ABOVE IS AN ACCURATE RECORD OF THE TEST CONDUCTED; THAT THE TEST WAS CONDUCTED WITH THE TYPE OF EQUIPMENT AND IN ACCORDANCE WITH THE METHODS APPROVED BY THE DEPARTMENT OF FORENSIC SCIENCE; THAT THE TEST WAS CONDUCTED IN ACCORDANCE WITH THE DEPARTMENT'S SPECIFICATIONS; THAT PRIOR TO ADMINISTRATION OF THE TEST THE ACCUSED WAS ADVISED OF HIS RIGHT TO OBSERVE THE PROCESS AND SEE THE BLOOD ALCOHOL READING ON THE EQUIPMENT USED TO PERFORM THE BREATH TEST, AND THAT I POSSESS A VALID LICENSE TO CONDUCT SUCH TEST, GIVEN UNDER MY HAND THIS _____ DAY OF _____, 20_____.

BREATH TEST OPERATOR

I HAVE RECEIVED A COPY OF THIS CERTIFICATE OF ANALYSIS _____ SUBJECT'S SIGNATURE

SUBJECT REFUSED TO SIGN FOR COPY OF CERTIFICATE OF ANALYSIS _____ OPERATOR'S SIGNATURE

FOR COURT



IntoxNet MIS Report
Report Generated 22 Apr 2015 at 08:31

HMS
mmb

Test Results

Instrument Serial Number 010470

Test # 000856 Subject Test

Test Location 1 Department of
Test Date 22 Apr 2015

Test Location 2 Forensic Science
Test Time 07:41
Remote/Local Local

Test Location 3
System Check Passed

Operator's Last Name STANTON
Agency DFS Central Lab
Card Serial Number 118910
Subject's Last Name INSTRUMENT
Subject's Middle Initial
Driver's License Number
Driver's License State

Operator's First Name HEATHER
Effective Date 10/01/2014
Subject's Date of Birth 00/00/0000

Operator's Middle Initial M
License Number 18910
Expiration Date 10/01/2016
Subject's First Name TEST
Subject's Sex Male
Driver's License Expiration 00/00/0000
Court Name DFS

End Date	22 Apr 2015	End Time	07:49	Result Time	07:47	Result Date	22 Apr 2015	Result	0.00
Data Type	DIAG	Sample Value	Pass	Sample Time	07:41				
Data Type	BLK	Sample Value	0.000	Sample Time	07:42				
Data Type	CHK	Sample Value	0.102	Sample Time	07:43				
Data Type	BLK	Sample Value	0.000	Sample Time	07:44				
Data Type	SUBJ	Sample Value	0.000	Sample Time	07:45				
Data Type	BLK	Sample Value	0.000	Sample Time	07:46				
Data Type	SUBJ	Sample Value	0.000	Sample Time	07:47				
Data Type	BLK	Sample Value	0.000	Sample Time	07:48				

Standard Type Dry Gas Std
Standard Lot Number AG406301-21
Tank Pressure 1002

Standard Value 0.100
Standard Expiration Date 03/04/2016
Barometric Pressure 760 mmHg

Blow Sample Number	1	Blow Duration	3.31 sec	Blow Volume	1800 cc	End-of-Blow Time	07:45
Blow Sample Number	2	Blow Duration	3.41 sec	Blow Volume	1746 cc	End-of-Blow Time	07:47
Tamper Evident Stamp 48335703				Test Status Code 0			
Test Status Success							

HMS

MMB

Intox EC/IR-II: Accuracy Check

Department of Forensic Science

Serial Number: 010470 Test Number: 857

Test Date: 04/22/2015 Test Time: 07:51 EDT

Dry Gas Target: 0.100

Lot Number: AG406301-21 Exp Date: 03/04/2016

Tank Pressure: 997 psi Barometric Pressure: 760 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	07:52
CHK	0.102	07:53
BLK	0.000	07:54
CHK	0.101	07:55
BLK	0.000	07:56
CHK	0.101	07:57

Test Status: *Success*

Calibration CRC: E166735F

LMS

mmB

Intox EC/IR-II: Accuracy Check

Department of Forensic Science

Serial Number: 010470 Test Number: 858

Test Date: 04/22/2015 Test Time: 08:01 EDT

Dry Gas Target: 0.300

Lot Number: AG506501-09 Exp Date: 03/06/2017

Tank Pressure: 193 psi Barometric Pressure: 760 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	08:02
CHK	0.306	08:03
BLK	0.000	08:04
CHK	0.305	08:05
BLK	0.000	08:07
CHK	0.305	08:08

Test Status: *Success*

Calibration CRC: E166735F