

Department of Forensic Science
Breath Alcohol Section
Instrument Location History
1/1/1990 To 4/16/2015

AMS
mmb

Instrument Serial Number: 010605

As of 16-Apr-15

Start Date	End Date	Agency Name
14-Apr-15		DFS Central Lab
05-Nov-14	14-Apr-15	Craig County SO
21-Oct-14	05-Nov-14	DFS Central Lab
20-May-14	21-Oct-14	Clifton Forge PD
06-May-14	20-May-14	DFS Central Lab
05-May-09	06-May-14	Staunton PD
04-Aug-08	05-May-09	DFS Central Lab

JMS
mmb

INTOX EC/IR II Quality Assurance Worksheet

Instrument Serial Number 010605 Worksheet Start Date 4/14/2015

Location Craig County

Address 182 Main St., Suite 1, New Castle, VA 24127

DFS Technician Heather Stanton License No. 18910

Laboratory On-Site

Site Specification: No detrimental environmental conditions exist.

Instrument Barometer (mm HG) 726 Reference Barometer (mm HG) 729

Reference Barometer(RB)Serial # 009111 RB Calibration Due 5/19/2015

Measurement Assurance Check


Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.291
0.300	0.286	0.277	0.294	Sample 2	0.291
Precision		sample min	sample max	Sample 3	0.290
0.001		0.290	0.291		

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.096
0.100	0.095	0.092	0.098	Sample 2	0.096
Precision		sample min	sample max	Sample 3	0.096
0		0.096	0.096		

Dry gas standard Lot No. (with tank no.) AG406301-23

- Replaced dry gas standard (+O-ring)
- Installed at Location
- Removed to DFS-Central

Supplies
Mouthpieces
Certificates of Analysis
Operator Worksheet
Other:



Notes:

Instrument Serial Number

010605

Certification Date

04/16/15

Calibrated

Certified

Measurement Assurance Check

Instrument Test

Troubleshooting/Maintenance Worksheet Completed

(Successful completion denotes satisfactory condition of the item.)

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.250		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.150		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.080		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.020		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

All measurements are in g/210L

Estimation of Uncertainty of Measurement and traceability records are located within the Breath Alcohol Section

DFS Technician

Heath m Stark

Date

4/16/15

Issuing Analyst

Michelle Bellin

Date

4/16/15

JPD

HMS
mmb



COMMONWEALTH OF VIRGINIA
DEPARTMENT OF FORENSIC SCIENCE

CERTIFICATE OF BLOOD ALCOHOL ANALYSIS
AS DETERMINED BY A CHEMICAL TEST OF THE ACCUSED'S BREATH

NAME OF ACCUSED INSTRUMENT, TEST,	NAME OF COURT DFS
---	-----------------------------

BREATH ANALYSIS

SAMPLE EXAMINED AND TEST CONDUCTED BY STANTON, HEATHER, M		AGENCY DFS Central Lab
DFS LICENSE NUMBER 18910	LICENSE EXPIRES 10/01/2016	DATE TEST CONDUCTED 04/14/2015
TEST EQUIPMENT NUMBER 010605		

RESULTS: TIME SAMPLE TAKEN 13:03 EDT

SAMPLE'S ALCOHOL CONTENT 0.00 GRAMS PER 210 LITERS OF BREATH

ATTEST:

I CERTIFY THAT THE ABOVE IS AN ACCURATE RECORD OF THE TEST CONDUCTED; THAT THE TEST WAS CONDUCTED WITH THE TYPE OF EQUIPMENT AND IN ACCORDANCE WITH THE METHODS APPROVED BY THE DEPARTMENT OF FORENSIC SCIENCE; THAT THE TEST WAS CONDUCTED IN ACCORDANCE WITH THE DEPARTMENT'S SPECIFICATIONS; THAT PRIOR TO ADMINISTRATION OF THE TEST THE ACCUSED WAS ADVISED OF HIS RIGHT TO OBSERVE THE PROCESS AND SEE THE BLOOD ALCOHOL READING ON THE EQUIPMENT USED TO PERFORM THE BREATH TEST, AND THAT I POSSESS A VALID LICENSE TO CONDUCT SUCH TEST, GIVEN UNDER MY HAND THIS _____ DAY OF _____, 20_____.

BREATH TEST OPERATOR

I HAVE RECEIVED A COPY OF THIS CERTIFICATE OF ANALYSIS _____ SUBJECT'S SIGNATURE

SUBJECT REFUSED TO SIGN FOR COPY OF CERTIFICATE OF ANALYSIS _____ OPERATOR'S SIGNATURE

FOR COURT



IntoxNet MIS Report
Report Generated 16 Apr 2015 at 08:33

HMS
MMB

Test Results

Instrument Serial Number 010605

Test # 000889 Subject Test

Test Location 1 Department of
Test Date 14 Apr 2015

Test Location 2 Forensic Science
Test Time 12:57
Remote/Local Local

Test Location 3
System Check Passed

Operator's Last Name STANTON
Agency DFS Central Lab

Operator's First Name HEATHER

Operator's Middle Initial M

Card Serial Number 118910

License Number 18910

Effective Date 10/01/2014

Expiration Date 10/01/2016

Subject's Last Name INSTRUMENT

Subject's First Name TEST

Subject's Middle Initial

Subject's Date of Birth 00/00/0000

Subject's Sex Male

Driver's License Number

Driver's License Expiration 00/00/0000

Driver's License State

Court Name DFS

End Date 14 Apr 2015 End Time 13:04

Result Time 13:03

Result Date 14 Apr 2015

Result 0.00

Data Type DIAG

Sample Value Pass

Sample Time 12:57

Data Type BLK

Sample Value 0.000

Sample Time 12:58

Data Type CHK

Sample Value 0.097

Sample Time 12:59

Data Type BLK

Sample Value 0.000

Sample Time 13:00

Data Type SUBJ

Sample Value 0.000

Sample Time 13:01

Data Type BLK

Sample Value 0.000

Sample Time 13:02

Data Type SUBJ

Sample Value 0.000

Sample Time 13:03

Data Type BLK

Sample Value 0.000

Sample Time 13:04

Standard Type Dry Gas Std

Standard Value 0.095

Standard Lot Number AG406301-23

Standard Expiration Date 03/04/2016

Tank Pressure 1115

Barometric Pressure 725 mmHg

Blow Sample Number 1 Blow Duration 3.36 sec

Blow Volume 1746 cc End-of-Blow Time 13:01

Blow Sample Number 2 Blow Duration 3.49 sec

Blow Volume 1750 cc End-of-Blow Time 13:03

Tamper Evident Stamp b5ac6859

Test Status Code 0

Test Status Success

AMS
mmB

Intox EC/IR-II: Accuracy Check

Department of Forensic Science

Serial Number: 010605 Test Number: 890

Test Date: 04/14/2015 Test Time: 13:08 EDT

Dry Gas Target: 0.095

Lot Number: AG406301-23 Exp Date: 03/04/2016

Tank Pressure: 1115 psi Barometric Pressure: 725 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	13:08
CHK	0.096	13:09
BLK	0.000	13:11
CHK	0.096	13:12
BLK	0.000	13:13
CHK	0.096	13:14

Test Status: *Success*

Calibration CRC: 9D97808E

HMS

mmb

Intox EC/IR-II: Accuracy Check

Department of Forensic Science

Serial Number: 010605 Test Number: 891

Test Date: 04/14/2015 Test Time: 13:18 EDT

Dry Gas Target: 0.286

Lot Number: AG506501-09 Exp Date: 03/06/2017

Tank Pressure: 321 psi Barometric Pressure: 725 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	13:19
CHK	0.291	13:20
BLK	0.000	13:22
CHK	0.291	13:23
BLK	0.000	13:24
CHK	0.290	13:25

Test Status: *Success*

Calibration CRC: 9D97808E