

Department of Forensic Science
Breath Alcohol Section
Instrument Location History
1/1/1990 To 4/23/2015

DBH
MMB

Instrument Serial Number: 010627

As of 23-Apr-15

Start Date	End Date	Agency Name
22-Apr-15		DFS Central Lab
13-Nov-14	22-Apr-15	Blacksburg PD
28-Oct-14	13-Nov-14	DFS Central Lab
11-Jul-14	28-Oct-14	Amherst County SO
24-Jun-14	11-Jul-14	DFS Central Lab
06-Jan-09	24-Jun-14	Westmoreland County Sheriff's Office
15-Aug-08	06-Jan-09	DFS Central Lab

Instrument Serial Number

010627

Certification Date

Calibrated

Certified

Measurement Assurance Check

Instrument Test

Troubleshooting/Maintenance Worksheet Completed

(Successful completion denotes satisfactory condition of the item.)

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.250		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.150		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.080		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.020		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

All measurements are in g/210L

Estimation of Uncertainty of Measurement and traceability records are located within the Breath Alcohol Section

DFS Technician Donald B Hall Date 4/23/15

Issuing Analyst Michele Ballin Date 4/24/15

SU

DB/H
mmB



COMMONWEALTH OF VIRGINIA
DEPARTMENT OF FORENSIC SCIENCE

CERTIFICATE OF BLOOD ALCOHOL ANALYSIS
AS DETERMINED BY A CHEMICAL TEST OF THE ACCUSED'S BREATH

NAME OF ACCUSED INSTRUMENT, TEST,	NAME OF COURT DPS
--------------------------------------	----------------------

BREATH ANALYSIS

SAMPLE EXAMINED AND TEST CONDUCTED BY HALL, DONALD, B		AGENCY DPS Central Lab
DFS LICENSE NUMBER 21290	LICENSE EXPIRES 10/01/2016	DATE TEST CONDUCTED 04/22/2015
TEST EQUIPMENT NUMBER 010627		

RESULTS: TIME SAMPLE TAKEN 11:04 EDT

SAMPLE'S ALCOHOL CONTENT 0.00 GRAMS PER 210 LITERS OF BREATH

ATTEST:

I CERTIFY THAT THE ABOVE IS AN ACCURATE RECORD OF THE TEST CONDUCTED; THAT THE TEST WAS CONDUCTED WITH THE TYPE OF EQUIPMENT AND IN ACCORDANCE WITH THE METHODS APPROVED BY THE DEPARTMENT OF FORENSIC SCIENCE; THAT THE TEST WAS CONDUCTED IN ACCORDANCE WITH THE DEPARTMENT'S SPECIFICATIONS; THAT PRIOR TO ADMINISTRATION OF THE TEST THE ACCUSED WAS ADVISED OF HIS RIGHT TO OBSERVE THE PROCESS AND SEE THE BLOOD ALCOHOL READING ON THE EQUIPMENT USED TO PERFORM THE BREATH TEST, AND THAT I POSSESS A VALID LICENSE TO CONDUCT SUCH TEST, GIVEN UNDER MY HAND THIS _____ DAY OF _____, 20_____.

BREATH TEST OPERATOR

I HAVE RECEIVED A COPY OF THIS CERTIFICATE OF ANALYSIS _____ SUBJECT'S SIGNATURE

SUBJECT REFUSED TO SIGN FOR COPY OF CERTIFICATE OF ANALYSIS _____ OPERATOR'S SIGNATURE

DBT
mmf

Intox EC/IR-II: Accuracy Check

Department of Forensic Science

Serial Number: 010627 Test Number: 709

Test Date: 04/22/2015 Test Time: 11:08 EDT

Dry Gas Target: 0.092

Lot Number: AG406301-25 Exp Date: 03/04/2016

Tank Pressure: 693 psi Barometric Pressure: 699 mmHg

System Check: Passed

Test	g/210L	Time
BLK	0.000	11:08
CHK	0.093	11:09
BLK	0.000	11:11
CHK	0.093	11:11
BLK	0.000	11:13
CHK	0.093	11:14

Test Status: Success

Calibration CRC: 3D7490EF

Intox EC/IR-II: Accuracy Check

DBH
mmB

Department of Forensic Science

Serial Number: 010627 Test Number: 710

Test Date: 04/22/2015 Test Time: 11:19 EDT

Dry Gas Target: 0.276

Lot Number: AG428003-05 Exp Date: 10/07/2016

Tank Pressure: 75 psi Barometric Pressure: 699 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	11:19
CHK	0.283	11:20
BLK	0.000	11:22
CHK	0.283	11:23
BLK	0.000	11:24
CHK	0.283	11:25

Test Status: *Success*

Calibration CRC: 3D7490EF