

Department of Forensic Science
Breath Alcohol Section
Instrument Location History
1/1/1990 To 5/12/2015

Instrument Serial Number: 010425

As of 12-May-15

Start Date	End Date	Agency Name
12-May-15		DFS Central Lab
10-Dec-14	12-May-15	Poquoson PD
21-Oct-14	10-Dec-14	DFS Central Lab
21-May-14	21-Oct-14	Alleghany County SO
01-May-14	21-May-14	DFS Central Lab
30-Sep-09	01-May-14	Orange County SO
08-May-08	30-Sep-09	DFS Central Lab

TLW

INTOX EC/IR II Quality Assurance Worksheet

Instrument Serial Number 010425 Worksheet Start Date 5/12/2015
 Location Poquoson PD
 Address 774 Poquoson Ave Poquoson VA 23662
 DFS Technician Timmy Neece, Jr License No. 25387
 Laboratory On-Site

Site Specification: No detrimental environmental conditions exist.

Instrument Barometer (mm HG) 760 Reference Barometer (mm HG) 761
 Reference Barometer(RB)Serial # 009109 RB Calibration Due 7/28/2015

Measurement Assurance Check

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.306
0.300	0.300	0.291	0.309	Sample 2	0.306
Precision		sample min	sample max	Sample 3	0.306
0		0.306	0.306		

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.101
0.100	0.100	0.097	0.103	Sample 2	0.100
Precision		sample min	sample max	Sample 3	0.100
0.001		0.100	0.101		


Dry gas standard Lot No. (with tank no.) AG316301-06

Replaced dry gas standard (+O-ring)

Installed at Location

Removed to DFS-Central

Supplies
 Mouthpieces
 Certificates of Analysis
 Operator Worksheet
 Other:



Notes:

Instrument Serial Number

010425

Certification Date

Calibrated

Certified

Measurement Assurance Check

Instrument Test

Troubleshooting/Maintenance Worksheet Completed

(Successful completion denotes satisfactory condition of the item.)

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.250		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.150		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	


Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.080		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.020		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

All measurements are in g/210L

Estimation of Uncertainty of Measurement and traceability records are located within the Breath Alcohol Section


DFS Technician



Date

5/12/15

Issuing Analyst



Date

5/12/15

JPD

TLJ



COMMONWEALTH OF VIRGINIA
DEPARTMENT OF FORENSIC SCIENCE

CERTIFICATE OF BLOOD ALCOHOL ANALYSIS
AS DETERMINED BY A CHEMICAL TEST OF THE ACCUSED'S BREATH

NAME OF ACCUSED INSTRUMENT, TEST,	NAME OF COURT DFS
---	-----------------------------

BREATH ANALYSIS

SAMPLE EXAMINED AND TEST CONDUCTED BY NEECE, JR., TIMMY, L		AGENCY DFS Central Lab
DFS LICENSE NUMBER 25387	LICENSE EXPIRES 10/01/2016	DATE TEST CONDUCTED 05/12/2015
TEST EQUIPMENT NUMBER 010425		

RESULTS: TIME SAMPLE TAKEN 10:28 EDT

SAMPLE'S ALCOHOL CONTENT 0.00 GRAMS PER 210 LITERS OF BREATH

ATTEST:

I CERTIFY THAT THE ABOVE IS AN ACCURATE RECORD OF THE TEST CONDUCTED; THAT THE TEST WAS CONDUCTED WITH THE TYPE OF EQUIPMENT AND IN ACCORDANCE WITH THE METHODS APPROVED BY THE DEPARTMENT OF FORENSIC SCIENCE; THAT THE TEST WAS CONDUCTED IN ACCORDANCE WITH THE DEPARTMENT'S SPECIFICATIONS; THAT PRIOR TO ADMINISTRATION OF THE TEST THE ACCUSED WAS ADVISED OF HIS RIGHT TO OBSERVE THE PROCESS AND SEE THE BLOOD ALCOHOL READING ON THE EQUIPMENT USED TO PERFORM THE BREATH TEST, AND THAT I POSSESS A VALID LICENSE TO CONDUCT SUCH TEST, GIVEN UNDER MY HAND THIS _____ DAY OF _____, 20 _____.

BREATH TEST OPERATOR

I HAVE RECEIVED A COPY OF THIS CERTIFICATE OF ANALYSIS _____ SUBJECT'S SIGNATURE

SUBJECT REFUSED TO SIGN FOR COPY OF CERTIFICATE OF ANALYSIS _____ OPERATOR'S SIGNATURE

TLN



IntoxNet MIS Report

Report Generated 12 May 2015 at 11:01

Test Results

Instrument Serial Number 010425

Test # 001063 **Subject Test**

Test Location 1 Department of
Test Date 12 May 2015

Test Location 2 Forensic Science
Test Time 10:22
Remote/Local Local

Test Location 3
System Check Passed

Operator's Last Name NEECE, JR.
Agency DFS Central Lab

Operator's First Name TIMMY

Operator's Middle Initial L

Card Serial Number 125387
Subject's Last Name INSTRUMENT
Subject's Middle Initial

License Number 25387
Effective Date 10/01/2014 Expiration Date 10/01/2016

Subject's First Name TEST

Subject's Date of Birth 00/00/0000 Subject's Sex Male

Driver's License Number
Driver's License State

Driver's License Expiration 00/00/0000

Court Name DFS

End Date 12 May 2015 End Time 10:29 Result Time 10:28 Result Date 12 May 2015 Result 0.00

Data Type	DIAG	Sample Value	Pass	Sample Time	10:22
Data Type	BLK	Sample Value	0.000	Sample Time	10:23
Data Type	CHK	Sample Value	0.101	Sample Time	10:24
Data Type	BLK	Sample Value	0.000	Sample Time	10:25
Data Type	SUBJ	Sample Value	0.000	Sample Time	10:26
Data Type	BLK	Sample Value	0.000	Sample Time	10:27
Data Type	SUBJ	Sample Value	0.000	Sample Time	10:28
Data Type	BLK	Sample Value	0.000	Sample Time	10:29

Standard Type Dry Gas Std

Standard Value 0.100

Standard Lot Number AG316301-06

Standard Expiration Date 06/12/2015

Tank Pressure 223

Barometric Pressure 761 mmHg

Blow Sample Number 1 Blow Duration 3.47 sec

Blow Volume 1805 cc End-of-Blow Time 10:26

Blow Sample Number 2 Blow Duration 4.04 sec

Blow Volume 2018 cc End-of-Blow Time 10:28

Tamper Evident Stamp b003c246

Test Status Code 0

Test Status Success

TLN

Intox EC/IR-II: Accuracy Check

Department of Forensic Science

Serial Number: 010425 Test Number: 1064

Test Date: 05/12/2015 Test Time: 10:32 EDT

Dry Gas Target: 0.100

Lot Number: AG316301-06 Exp Date: 06/12/2015

Tank Pressure: 223 psi Barometric Pressure: 760 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	10:32
CHK	0.101	10:33
BLK	0.000	10:35
CHK	0.100	10:36
BLK	0.000	10:37
CHK	0.100	10:38

Test Status: *Success*

Calibration CRC: C094246F

TLN

Intox EC/IR-II: Accuracy Check

Department of Forensic Science

Serial Number: 010425 Test Number: 1065

Test Date: 05/12/2015 Test Time: 10:42 EDT

Dry Gas Target: 0.300

Lot Number: AG428003-02 Exp Date: 10/07/2016

Tank Pressure: 281 psi Barometric Pressure: 760 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	10:43
CHK	0.306	10:44
BLK	0.000	10:45
CHK	0.306	10:46
BLK	0.000	10:48
CHK	0.306	10:49

Test Status: *Success*

Calibration CRC: C094246F