

Department of Forensic Science
Breath Alcohol Section
Instrument Location History
1/1/1990 To 4/29/2015

LMS
TW

Instrument Serial Number: 010452

As of 29-Apr-15

Start Date	End Date	Agency Name
29-Apr-15		Richmond PD
13-Apr-15	29-Apr-15	DFS Central Lab
28-Oct-14	13-Apr-15	Amherst County SO
15-Oct-14	28-Oct-14	DFS Central Lab
06-May-14	15-Oct-14	Waynesboro PD
01-May-14	06-May-14	DFS Central Lab
25-Mar-09	01-May-14	Russell County SO
12-Mar-09	25-Mar-09	DFS Central Lab
22-Oct-08	12-Mar-09	Norton PD
13-May-08	22-Oct-08	DFS Central Lab

LMS
TLN

INTOX EC/IR II Quality Assurance Worksheet

Instrument Serial Number 010452 Worksheet Start Date 4/29/2015

Location Richmond City Justice Center

Address 1701 Fairfield Way, Richmond, VA 23223

DFS Technician Heather Stanton License No. 18910

Laboratory On-Site

Site Specification: No detrimental environmental conditions exist.

Instrument Barometer (mm HG) 755 Reference Barometer (mm HG) 755

Reference Barometer(RB)Serial # 009113 RB Calibration Due 4/16/2016

Measurement Assurance Check

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.306
0.300	0.298	0.289	0.306	Sample 2	0.306
Precision		sample min	sample max	Sample 3	0.305
0.001		0.305	0.306		

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.100
0.100	0.099	0.096	0.102	Sample 2	0.100
Precision		sample min	sample max	Sample 3	0.100
0		0.100	0.100		

Dry gas standard Lot No. (with tank no.) AG406301-17

Replaced dry gas standard (+O-ring)

Installed at Location

Removed to DFS-Central

Supplies
Mouthpieces
Certificates of Analysis
Operator Worksheet
Other:

100

Notes:

Instrument Serial Number

010452

Certification Date

Calibrated

Certified

Measurement Assurance Check

Instrument Test

Troubleshooting/Maintenance Worksheet Completed

(Successful completion denotes satisfactory condition of the item.)

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.250		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.150		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.080		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.020		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

All measurements are in g/210L

Estimation of Uncertainty of Measurement and traceability records are located within the Breath Alcohol Section

DFS Technician

Heath W. Starch

Date

4/29/15

Issuing Analyst

J. [Signature]

Date

4/30/15

Handwritten initials/signature
TLW

Intox EC/IR-II: Accuracy Check

Department of Forensic Science

Serial Number: 010452 Test Number: 767
Test Date: 04/29/2015 Test Time: 10:36 EDT
Dry Gas Target: 0.298

Lot Number: AG506501-09 Exp Date: 03/06/2017
Tank Pressure: 105 psi Barometric Pressure: 755 mmHg
System Check: *Passed*

Test	g/210L	Time
BLK	0.000	10:37
CHK	0.306	10:38
BLK	0.000	10:39
CHK	0.306	10:40
BLK	0.000	10:42
CHK	0.305	10:43

Test Status: *Success*

Calibration CRC: 839A89E7

HMS
TLN

Intox EC/IR-II: Accuracy Check

Department of Forensic Science

Serial Number: 010452 Test Number: 768

Test Date: 04/29/2015 Test Time: 10:46 EDT

Dry Gas Target: 0.099

Lot Number: AG406301-17 Exp Date: 03/04/2016

Tank Pressure: 1090 psi Barometric Pressure: 755 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	10:47
CHK	0.100	10:48
BLK	0.000	10:50
CHK	0.100	10:50
BLK	0.000	10:52
CHK	0.100	10:52

Test Status: *Success*

Calibration CRC: 839A89E7

AMS
xhw



COMMONWEALTH OF VIRGINIA
DEPARTMENT OF FORENSIC SCIENCE

CERTIFICATE OF BLOOD ALCOHOL ANALYSIS
AS DETERMINED BY A CHEMICAL TEST OF THE ACCUSED'S BREATH

NAME OF ACCUSED INSTRUMENT, TEST,	NAME OF COURT DFS
--	--------------------------

BREATH ANALYSIS

SAMPLE EXAMINED AND TEST CONDUCTED BY STANTON, HEATHER, M		AGENCY DFS Central Lab
DFS LICENSE NUMBER 18910	LICENSE EXPIRES 10/01/2016	DATE TEST CONDUCTED 04/29/2015
TEST EQUIPMENT NUMBER 010452		

RESULTS: TIME SAMPLE TAKEN 11:00 EDT

SAMPLE'S ALCOHOL CONTENT 0.00 GRAMS PER 210 LITERS OF BREATH

ATTEST:

I CERTIFY THAT THE ABOVE IS AN ACCURATE RECORD OF THE TEST CONDUCTED; THAT THE TEST WAS CONDUCTED WITH THE TYPE OF EQUIPMENT AND IN ACCORDANCE WITH THE METHODS APPROVED BY THE DEPARTMENT OF FORENSIC SCIENCE; THAT THE TEST WAS CONDUCTED IN ACCORDANCE WITH THE DEPARTMENT'S SPECIFICATIONS; THAT PRIOR TO ADMINISTRATION OF THE TEST THE ACCUSED WAS ADVISED OF HIS RIGHT TO OBSERVE THE PROCESS AND SEE THE BLOOD ALCOHOL READING ON THE EQUIPMENT USED TO PERFORM THE BREATH TEST, AND THAT I POSSESS A VALID LICENSE TO CONDUCT SUCH TEST, GIVEN UNDER MY HAND THIS _____ DAY OF _____, 20_____.

BREATH TEST OPERATOR

I HAVE RECEIVED A COPY OF THIS CERTIFICATE OF ANALYSIS _____ SUBJECT'S SIGNATURE

SUBJECT REFUSED TO SIGN FOR COPY OF CERTIFICATE OF ANALYSIS _____ OPERATOR'S SIGNATURE

FOR COURT



IntoxNet MIS Report
Report Generated 29 Apr 2015 at 12:55

Handwritten initials:
HMS
TLW

Test Results

Instrument Serial Number 010452

Test # 000769 Subject Test

Test Location 1 Department of
Test Date 29 Apr 2015

Test Location 2 Forensic Science
Test Time 10:54
Remote/Local Local

Test Location 3
System Check Passed

Operator's Last Name STANTON
Agency DFS Central Lab

Operator's First Name HEATHER

Operator's Middle Initial M

Card Serial Number 118910

License Number 18910

Effective Date 10/01/2014

Expiration Date 10/01/2016

Subject's Last Name INSTRUMENT

Subject's First Name TEST

Subject's Middle Initial

Subject's Date of Birth 00/00/0000

Subject's Sex Male

Driver's License Number

Driver's License Expiration 00/00/0000

Driver's License State

Court Name DFS

End Date 29 Apr 2015 End Time 11:01

Result Time 11:00

Result Date 29 Apr 2015

Result 0.00

Data Type DIAG

Sample Value Pass

Sample Time 10:54

Data Type BLK

Sample Value 0.000

Sample Time 10:55

Data Type CHK

Sample Value 0.100

Sample Time 10:56

Data Type BLK

Sample Value 0.000

Sample Time 10:57

Data Type SUBJ

Sample Value 0.000

Sample Time 10:58

Data Type BLK

Sample Value 0.000

Sample Time 10:59

Data Type SUBJ

Sample Value 0.000

Sample Time 11:00

Data Type BLK

Sample Value 0.000

Sample Time 11:01

Standard Type Dry Gas Std

Standard Value 0.099

Standard Lot Number AG406301-17

Standard Expiration Date 03/04/2016

Tank Pressure 1080

Barometric Pressure 755 mmHg

Blow Sample Number 1 Blow Duration 3.67 sec

Blow Volume 1716 cc End-of-Blow Time 10:58

Blow Sample Number 2 Blow Duration 3.68 sec

Blow Volume 1742 cc End-of-Blow Time 11:00

Tamper Evident Stamp fb2617ce

Test Status Code 0

Test Status Success