

TLN

Department of Forensic Science
Breath Alcohol Section
Instrument Location History
1/1/1990 To 5/12/2015

Instrument Serial Number: 010453

As of 12-May-15

Start Date	End Date	Agency Name
12-May-15		Poquoson PD
29-Apr-15	12-May-15	DFS Central Lab
05-Dec-14	29-Apr-15	Richmond County Sheriff's Office
30-Oct-14	05-Dec-14	DFS Central Lab
29-May-14	30-Oct-14	Campbell County SO
20-May-14	29-May-14	DFS Central Lab
27-Feb-14	20-May-14	Clifton Forge PD
20-Feb-14	27-Feb-14	DFS Central Lab
06-Aug-09	20-Feb-14	Clifton Forge PD
22-Apr-09	06-Aug-09	DFS Central Lab
06-Nov-08	22-Apr-09	Chase City PD
19-May-08	06-Nov-08	DFS Central Lab

TLN

Department of Forensic Science
Breath Alcohol Section
Instrument Maintenance History
5/12/2014 To 5/12/2015

Instrument Serial Number: 010453 As of 12-May-15

Maintenance Date	Certification Date	Technician Initials	Remarks
12-May-15	01-May-15	TLN (25387)	REPLACED UPS BACK-UP BATTERY.
01-May-15	01-May-15	DBH (21250)	CERTIFIED.
07-Nov-14	07-Nov-14	SED (24178)	CERTIFIED.
23-Jul-14	22-May-14	DBH (21250)	REPLACED UPS BACK-UP BATTERY.
22-May-14	22-May-14	DBH (21250)	CERTIFIED.

TJA

INTOX EC/IR II Quality Assurance Worksheet

Instrument Serial Number: 010453 Worksheet Start Date: 5/12/2015
 Location: Poquoson PD
 Address: 774 Poquoson Ave Poquoson VA 23662
 DFS Technician: Timmy Neece, Jr License No.: 25387
 Laboratory On-Site

Site Specification: No detrimental environmental conditions exist.

Instrument Barometer (mm HG): 760 Reference Barometer (mm HG): 762
 Reference Barometer(RB)Serial #: 009109 RB Calibration Due: 7/28/2015

Measurement Assurance Check


Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.308
0.300	0.300	0.291	0.309	Sample 2	0.307
Precision		sample min	sample max	Sample 3	0.306
0.002		0.306	0.308		

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.101
0.100	0.100	0.097	0.103	Sample 2	0.101
Precision		sample min	sample max	Sample 3	0.101
0		0.101	0.101		

Dry gas standard Lot No. (with tank no.) AG400202-33

- Replaced dry gas standard (+O-ring)
- Installed at Location
- Removed to DFS-Central

Supplies
 Mouthpieces
 Certificates of Analysis
 Operator Worksheet
 Other:



Notes:

Replaced UPS back-up battery.

Instrument Serial Number

010453

Certification Date

Calibrated

Certified

Measurement Assurance Check

Instrument Test

Troubleshooting/Maintenance Worksheet Completed

(Successful completion denotes satisfactory condition of the item.)

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.250		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.150		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.080		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.020		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

All measurements are in g/210L

Estimation of Uncertainty of Measurement and traceability records are located within the Breath Alcohol Section

DFS Technician

[Signature]

Date

5/12/15

Issuing Analyst

[Signature]

Date

5/12/15

JPD

TLN

Intox EC/IR-II: Accuracy Check

Department of Forensic Science

Serial Number: 010453 Test Number: 451

Test Date: 05/12/2015 Test Time: 11:27 EDT

Dry Gas Target: 0.300

Lot Number: AG428003-02 Exp Date: 10/07/2016

Tank Pressure: 276 psi Barometric Pressure: 760 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	11:28
CHK	0.308	11:29
BLK	0.000	11:31
CHK	0.307	11:32
BLK	0.000	11:33
CHK	0.306	11:34

Test Status: *Success*

Calibration CRC: 60D5B217

TZ

Intox EC/IR-II: Accuracy Check

Department of Forensic Science

Serial Number: 010453 Test Number: 452

Test Date: 05/12/2015 Test Time: 11:38 EDT

Dry Gas Target: 0.100

Lot Number: AG400202-33 Exp Date: 01/02/2016

Tank Pressure: 183 psi Barometric Pressure: 760 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	11:38
CHK	0.101	11:39
BLK	0.000	11:41
CHK	0.101	11:41
BLK	0.000	11:43
CHK	0.101	11:43

Test Status: *Success*

Calibration CRC: 60D5B217

TLN



COMMONWEALTH OF VIRGINIA
DEPARTMENT OF FORENSIC SCIENCE

CERTIFICATE OF BLOOD ALCOHOL ANALYSIS
AS DETERMINED BY A CHEMICAL TEST OF THE ACCUSED'S BREATH

NAME OF ACCUSED INSTRUMENT, TEST,	NAME OF COURT DFS
---	-----------------------------

BREATH ANALYSIS

SAMPLE EXAMINED AND TEST CONDUCTED BY NEECE, JR., TIMMY, L		AGENCY DFS Central Lab
DFS LICENSE NUMBER 25387	LICENSE EXPIRES 10/01/2016	DATE TEST CONDUCTED 05/12/2015
TEST EQUIPMENT NUMBER 010453		

RESULTS: TIME SAMPLE TAKEN 11:52 EDT

SAMPLE'S ALCOHOL CONTENT 0.00 GRAMS PER 210 LITERS OF BREATH

ATTEST:

I CERTIFY THAT THE ABOVE IS AN ACCURATE RECORD OF THE TEST CONDUCTED; THAT THE TEST WAS CONDUCTED WITH THE TYPE OF EQUIPMENT AND IN ACCORDANCE WITH THE METHODS APPROVED BY THE DEPARTMENT OF FORENSIC SCIENCE; THAT THE TEST WAS CONDUCTED IN ACCORDANCE WITH THE DEPARTMENT'S SPECIFICATIONS; THAT PRIOR TO ADMINISTRATION OF THE TEST THE ACCUSED WAS ADVISED OF HIS RIGHT TO OBSERVE THE PROCESS AND SEE THE BLOOD ALCOHOL READING ON THE EQUIPMENT USED TO PERFORM THE BREATH TEST, AND THAT I POSSESS A VALID LICENSE TO CONDUCT SUCH TEST, GIVEN UNDER MY HAND THIS _____ DAY OF _____, 20 _____.

BREATH TEST OPERATOR

I HAVE RECEIVED A COPY OF THIS CERTIFICATE OF ANALYSIS _____
SUBJECT'S SIGNATURE

SUBJECT REFUSED TO SIGN FOR COPY OF CERTIFICATE OF ANALYSIS _____
OPERATOR'S SIGNATURE

TBN



IntoxNet MIS Report

Report Generated 12 May 2015 at 14:21

Test Results

Instrument Serial Number 010453

Test # 000453 Subject Test

Test Location 1 Department of
Test Date 12 May 2015

Test Location 2 Forensic Science
Test Time 11:46
Remote/Local Local

Test Location 3
System Check Passed

Operator's Last Name NEECE,JR.
Agency DFS Central Lab

Operator's First Name TIMMY

Operator's Middle Initial L

Card Serial Number 125387

License Number 25387

Effective Date 10/01/2014

Expiration Date 10/01/2016

Subject's Last Name INSTRUMENT

Subject's First Name TEST

Subject's Middle Initial

Subject's Date of Birth 00/00/0000

Subject's Sex Male

Driver's License Number

Driver's License Expiration 00/00/0000

Driver's License State

Court Name DFS

End Date 12 May 2015 End Time 11:53

Result Time 11:52

Result Date 12 May 2015 Result 0.00

Data Type DIAG

Sample Value Pass

Sample Time 11:46

Data Type BLK

Sample Value 0.000

Sample Time 11:47

Data Type CHK

Sample Value 0.101

Sample Time 11:47

Data Type BLK

Sample Value 0.000

Sample Time 11:49

Data Type SUBJ

Sample Value 0.000

Sample Time 11:50

Data Type BLK

Sample Value 0.000

Sample Time 11:51

Data Type SUBJ

Sample Value 0.000

Sample Time 11:52

Data Type BLK

Sample Value 0.000

Sample Time 11:53

Standard Type Dry Gas Std

Standard Value 0.100

Standard Lot Number AG400202-33

Standard Expiration Date 01/02/2016

Tank Pressure 178

Barometric Pressure 760 mmHg

Blow Sample Number 1 Blow Duration 3.47 sec

Blow Volume 1776 cc End-of-Blow Time 11:50

Blow Sample Number 2 Blow Duration 3.84 sec

Blow Volume 2051 cc End-of-Blow Time 11:52

Tamper Evident Stamp 1181467f

Test Status Code 0

Test Status Success