

DBJ
Tb

Department of Forensic Science
Breath Alcohol Section
Instrument Location History
1/1/1990 To 5/26/2015

Instrument Serial Number: 010457

As of 26-May-15

Start Date	End Date	Agency Name
26-May-15		DFS Central Lab
10-Dec-14	26-May-15	Smithfield PD
18-Nov-14	10-Dec-14	DFS Central Lab
11-Jun-14	18-Nov-14	Colonial Heights PD
04-Jun-12	11-Jun-14	DFS Central Lab
24-May-12	04-Jun-12	Virginia Beach PD
10-May-12	24-May-12	DFS Central Lab
07-Mar-11	10-May-12	Chesapeake PD
22-Apr-09	07-Mar-11	DFS Central Lab
12-Nov-08	22-Apr-09	Buena Vista PD
08-May-08	12-Nov-08	DFS Central Lab

DS11
Tln

INTOX EC/IR II Quality Assurance Worksheet

Instrument Serial Number: 010457 Worksheet Start Date: 5/26/2015

Location: Smithfield PD

Address: 913 S Church St Smithfield VA 23430

DFS Technician: Donald Hall License No.: 21250

Laboratory On-Site

Site Specification: No detrimental environmental conditions exist.

Instrument Barometer (mm HG): 765 Reference Barometer (mm HG): 767

Reference Barometer(RB)Serial #: 009109 RB Calibration Due: 7/28/2015

Measurement Assurance Check

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.301
0.300	0.302	0.292	0.311	Sample 2	0.301
Precision		sample min	sample max	Sample 3	0.300
0.001		0.300	0.301		

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.100
0.100	0.101	0.097	0.104	Sample 2	0.100
Precision		sample min	sample max	Sample 3	0.100
0		0.100	0.100		

Dry gas standard Lot No. (with tank no.) AG400202-35

- Replaced dry gas standard (+O-ring)
 - Installed at Location
 - Removed to DFS-Central
- Supplies
Mouthpieces
Certificates of Analysis
Operator Worksheet
Other:

Notes:

Instrument Serial Number 010457

Certification Date

Calibrated

Certified

Measurement Assurance Check

Instrument Test

Troubleshooting/Maintenance Worksheet Completed

(Successful completion denotes satisfactory condition of the item.)

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.250		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.150		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.080		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.020		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

All measurements are in g/210L

Estimation of Uncertainty of Measurement and traceability records are located within the Breath Alcohol Section

DFS Technician

David B Hall

Date

5/26/15

Issuing Analyst

J. V. 2

Date

5/27/15

JBT
TLN



COMMONWEALTH OF VIRGINIA
DEPARTMENT OF FORENSIC SCIENCE

CERTIFICATE OF BLOOD ALCOHOL ANALYSIS
AS DETERMINED BY A CHEMICAL TEST OF THE ACCUSED'S BREATH

NAME OF ACCUSED INSTRUMENT, TEST,	NAME OF COURT DFS
---	-----------------------------

BREATH ANALYSIS

SAMPLE EXAMINED AND TEST CONDUCTED BY HALL, DONALD, B		AGENCY DFS Central Lab
DFS LICENSE NUMBER 21250	LICENSE EXPIRES 10/01/2016	DATE TEST CONDUCTED 05/26/2015
TEST EQUIPMENT NUMBER 010457		

RESULTS: TIME SAMPLE TAKEN 12:13 EDT
SAMPLE'S ALCOHOL CONTENT 0.00 GRAMS PER 210 LITERS OF BREATH

ATTEST:

I CERTIFY THAT THE ABOVE IS AN ACCURATE RECORD OF THE TEST CONDUCTED; THAT THE TEST WAS CONDUCTED WITH THE TYPE OF EQUIPMENT AND IN ACCORDANCE WITH THE METHODS APPROVED BY THE DEPARTMENT OF FORENSIC SCIENCE; THAT THE TEST WAS CONDUCTED IN ACCORDANCE WITH THE DEPARTMENT'S SPECIFICATIONS; THAT PRIOR TO ADMINISTRATION OF THE TEST THE ACCUSED WAS ADVISED OF HIS RIGHT TO OBSERVE THE PROCESS AND SEE THE BLOOD ALCOHOL READING ON THE EQUIPMENT USED TO PERFORM THE BREATH TEST, AND THAT I POSSESS A VALID LICENSE TO CONDUCT SUCH TEST, GIVEN UNDER MY HAND THIS _____ DAY OF _____, 20_____.

BREATH TEST OPERATOR

I HAVE RECEIVED A COPY OF THIS CERTIFICATE OF ANALYSIS _____ SUBJECT'S SIGNATURE

SUBJECT REFUSED TO SIGN FOR COPY OF CERTIFICATE OF ANALYSIS _____ OPERATOR'S SIGNATURE

DGH
Tbr



IntoxNet MIS Report

Report Generated 26 May 2015 at 17:04

Test Results

Instrument Serial Number 010457

Test # 000702 Subject Test

Test Location 1 Department of
Test Date 26 May 2015

Test Location 2 Forensic Science
Test Time 12:06
Remote/Local Local

Test Location 3
System Check Passed

Operator's Last Name HALL
Agency DFS Central Lab

Operator's First Name DONALD

Operator's Middle Initial B

Card Serial Number 121250

License Number 21250

Effective Date 10/01/2014

Expiration Date 10/01/2016

Subject's Last Name INSTRUMENT

Subject's First Name TEST

Subject's Middle Initial

Subject's Date of Birth 00/00/0000

Subject's Sex Male

Driver's License Number

Driver's License Expiration 00/00/0000

Driver's License State

Court Name DFS

End Date 26 May 2015 End Time 12:14

Result Time 12:13

Result Date 26 May 2015 Result 0.00

Data Type DIAG

Sample Value Pass

Sample Time 12:06

Data Type BLK

Sample Value 0.000

Sample Time 12:07

Data Type CHK

Sample Value 0.100

Sample Time 12:08

Data Type BLK

Sample Value 0.000

Sample Time 12:09

Data Type SUBJ

Sample Value 0.000

Sample Time 12:10

Data Type BLK

Sample Value 0.000

Sample Time 12:11

Data Type SUBJ

Sample Value 0.000

Sample Time 12:13

Data Type BLK

Sample Value 0.000

Sample Time 12:13

Standard Type Dry Gas Std

Standard Value 0.101

Standard Lot Number AG400202-35

Standard Expiration Date 01/02/2016

Tank Pressure 502

Barometric Pressure 765 mmHg

Blow Sample Number 1 Blow Duration 3.98 sec

Blow Volume 2205 cc End-of-Blow Time 12:10

Blow Sample Number 2 Blow Duration 3.77 sec

Blow Volume 1963 cc End-of-Blow Time 12:13

Tamper Evident Stamp 9da3fccb

Test Status Code 0

Test Status Success

DBH
TLW

Intox EC/IR-II: Accuracy Check

Department of Forensic Science

Serial Number: 010457 Test Number: 703

Test Date: 05/26/2015 Test Time: 12:16 EDT

Dry Gas Target: 0.101

Lot Number: AG400202-35 Exp Date: 01/02/2016

Tank Pressure: 492 psi Barometric Pressure: 765 mmHg

System Check: Passed

Test	g/210L	Time
BLK	0.000	12:17
CHK	0.100	12:18
BLK	0.000	12:19
CHK	0.100	12:20
BLK	0.000	12:21
CHK	0.100	12:22

Test Status: Success

Calibration CRC: 976995F8

DBH
Tlw

Intox EC/IR-II: Accuracy Check

Department of Forensic Science

Serial Number: 010457 Test Number: 704

Test Date: 05/26/2015 Test Time: 12:25 EDT

Dry Gas Target: 0.302

Lot Number: AG506501-10 Exp Date: 03/06/2017

Tank Pressure: 144 psi Barometric Pressure: 765 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	12:25
CHK	0.301	12:26
BLK	0.000	12:28
CHK	0.301	12:28
BLK	0.000	12:30
CHK	0.300	12:31

Test Status: *Success*

Calibration CRC: 976995F8