

Department of Forensic Science
Breath Alcohol Section
Instrument Location History
1/1/1990 To 5/11/2015

*KMS
JFD*

Instrument Serial Number: 010464

As of 11-May-15

Start Date	End Date	Agency Name
05-May-15		DFS Central Lab
02-Dec-14	05-May-15	Henry County Sheriff's Office
22-Jul-14	02-Dec-14	DFS Central Lab
16-Jul-14	22-Jul-14	Loudoun County SO
18-Jun-14	16-Jul-14	DFS Central Lab
07-Jan-14	18-Jun-14	Franklin County SO
28-Dec-12	07-Jan-14	DFS Central Lab
14-Dec-11	28-Dec-12	Fairfax County PD
13-Oct-11	14-Dec-11	DFS Central Lab
20-Sep-10	13-Oct-11	Loudoun County SO
05-Aug-09	20-Sep-10	DFS Central Lab
17-Oct-08	05-Aug-09	Amelia County SO
08-May-08	17-Oct-08	DFS Central Lab

AMS
JrD

INTOX EC/IR II Quality Assurance Worksheet

Instrument Serial Number 010464 Worksheet Start Date 5/5/2015

Location Henry County

Address 3160 Kings Mountain Rd., Martinsville, VA 24112

DFS Technician Heather Stanton License No. 18910

Laboratory On-Site

Site Specification: No detrimental environmental conditions exist.

Instrument Barometer (mm HG) 743 Reference Barometer (mm HG) 743

Reference Barometer(RB)Serial # 009113 RB Calibration Due 4/16/2016

Measurement Assurance Check

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.300
0.300	0.293	0.284	0.301	Sample 2	0.300
Precision		sample min	sample max	Sample 3	0.300
0		0.300	0.300		

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.099
0.100	0.098	0.095	0.101	Sample 2	0.099
Precision		sample min	sample max	Sample 3	0.099
0		0.099	0.099		


Dry gas standard Lot No. (with tank no.) AG332301-25

Replaced dry gas standard (+O-ring)

Installed at Location

Removed to DFS-Central

Supplies
Mouthpieces
Certificates of Analysis
Operator Worksheet
Other:



Notes:

Instrument Serial Number

010464

Certification Date

05/11/15

Calibrated

Certified

Measurement Assurance Check

Instrument Test

Troubleshooting/Maintenance Worksheet Completed

(Successful completion denotes satisfactory condition of the item.)

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.250		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.150		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.080		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.020		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

All measurements are in g/210L

Estimation of Uncertainty of Measurement and traceability records are located within the Breath Alcohol Section

DFS Technician

Heath M. Starks

Date

5/11/15

Issuing Analyst

Jennie Duffy

Date

5/11/15

AMS
JPD



COMMONWEALTH OF VIRGINIA
DEPARTMENT OF FORENSIC SCIENCE

CERTIFICATE OF BLOOD ALCOHOL ANALYSIS
AS DETERMINED BY A CHEMICAL TEST OF THE ACCUSED'S BREATH

NAME OF ACCUSED INSTRUMENT, TEST,	NAME OF COURT DFS
--	--------------------------

BREATH ANALYSIS

SAMPLE EXAMINED AND TEST CONDUCTED BY STANTON, HEATHER, M		AGENCY DFS Central Lab
DPS LICENSE NUMBER 18910	LICENSE EXPIRES 10/01/2016	DATE TEST CONDUCTED 05/05/2015
TEST EQUIPMENT NUMBER 010464		

RESULTS: TIME SAMPLE TAKEN 15:02 EDT

SAMPLE'S ALCOHOL CONTENT 0.00 GRAMS PER 210 LITERS OF BREATH

ATTEST:

I CERTIFY THAT THE ABOVE IS AN ACCURATE RECORD OF THE TEST CONDUCTED; THAT THE TEST WAS CONDUCTED WITH THE TYPE OF EQUIPMENT AND IN ACCORDANCE WITH THE METHODS APPROVED BY THE DEPARTMENT OF FORENSIC SCIENCE; THAT THE TEST WAS CONDUCTED IN ACCORDANCE WITH THE DEPARTMENT'S SPECIFICATIONS; THAT PRIOR TO ADMINISTRATION OF THE TEST THE ACCUSED WAS ADVISED OF HIS RIGHT TO OBSERVE THE PROCESS AND SEE THE BLOOD ALCOHOL READING ON THE EQUIPMENT USED TO PERFORM THE BREATH TEST, AND THAT I POSSESS A VALID LICENSE TO CONDUCT SUCH TEST, GIVEN UNDER MY HAND THIS _____ DAY OF _____, 20_____.

BREATH TEST OPERATOR

I HAVE RECEIVED A COPY OF THIS CERTIFICATE OF ANALYSIS _____ SUBJECT'S SIGNATURE

SUBJECT REFUSED TO SIGN FOR COPY OF CERTIFICATE OF ANALYSIS _____ OPERATOR'S SIGNATURE

FOR COURT



IntoxNet MIS Report
Report Generated 07 May 2015 at 08:05

*HMS
JPD*

Test Results

Instrument Serial Number 010464

Test # 001378 Subject Test

Test Location 1 Department of
Test Date 05 May 2015

Test Location 2 Forensic Science
Test Time 14:56
Remote/Local Local

Test Location 3
System Check Passed

Operator's Last Name STANTON
Agency DFS Central Lab

Operator's First Name HEATHER

Operator's Middle Initial M

Card Serial Number 118910

License Number 18910

Effective Date 10/01/2014

Expiration Date 10/01/2016

Subject's Last Name INSTRUMENT

Subject's First Name TEST

Subject's Middle Initial

Subject's Date of Birth 00/00/0000

Subject's Sex Male

Driver's License Number

Driver's License Expiration 00/00/0000

Driver's License State

Court Name DFS

End Date 05 May 2015 End Time 15:04

Result Time 15:02

Result Date 05 May 2015 Result 0.00

Data Type DIAG

Sample Value Pass

Sample Time 14:56

Data Type BLK

Sample Value 0.000

Sample Time 14:57

Data Type CHK

Sample Value 0.100

Sample Time 14:58

Data Type BLK

Sample Value 0.000

Sample Time 14:59

Data Type SUBJ

Sample Value 0.000

Sample Time 15:00

Data Type BLK

Sample Value 0.000

Sample Time 15:01

Data Type SUBJ

Sample Value 0.000

Sample Time 15:02

Data Type BLK

Sample Value 0.000

Sample Time 15:03

Standard Type Dry Gas Std

Standard Value 0.098

Standard Lot Number AG332301-25

Standard Expiration Date 11/19/2015

Tank Pressure 634

Barometric Pressure 743 mmHg

Blow Sample Number 1 Blow Duration 3.41 sec

Blow Volume 1737 cc End-of-Blow Time 15:00

Blow Sample Number 2 Blow Duration 3.35 sec

Blow Volume 1763 cc End-of-Blow Time 15:02

Tamper Evident Stamp cd911388

Test Status Code 0

Test Status Success

Intox EC/IR-II: Accuracy Check

Handwritten initials:
JPD

Department of Forensic Science

Serial Number: 010464 Test Number: 1379
Test Date: 05/05/2015 Test Time: 15:06 EDT
Dry Gas Target: 0.098
Lot Number: AG332301-25 Exp Date: 11/19/2015
Tank Pressure: 629 psi Barometric Pressure: 743 mmHg
System Check: *Passed*

Test	g/210L	Time
BLK	0.000	15:06
CHK	0.099	15:07
BLK	0.000	15:09
CHK	0.099	15:10
BLK	0.000	15:11
CHK	0.099	15:12

Test Status: *Success*

Calibration CRC: 5216D8E7

HMS
JPP

Intox EC/IR-II: Accuracy Check

Department of Forensic Science

Serial Number: 010464 Test Number: 1380

Test Date: 05/05/2015 Test Time: 15:15 EDT

Dry Gas Target: 0.293

Lot Number: AG506501-08 Exp Date: 03/06/2017

Tank Pressure: 497 psi Barometric Pressure: 742 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	15:16
CHK	0.300	15:17
BLK	0.000	15:18
CHK	0.300	15:19
BLK	0.000	15:20
CHK	0.300	15:21

Test Status: *Success*

Calibration CRC: 5216D8E7