

Department of Forensic Science
Breath Alcohol Section
Instrument Location History
1/1/1990 To 4/30/2015

LMS
JPD

Instrument Serial Number: 010595

As of 30-Apr-15

Start Date	End Date	Agency Name
30-Apr-15		DFS Central Lab
19-Nov-14	30-Apr-15	Prince George County PD
05-Nov-14	19-Nov-14	DFS Central Lab
03-Jun-14	05-Nov-14	Craig County SO
15-Jan-14	03-Jun-14	DFS Central Lab
12-Dec-08	15-Jan-14	Danville PD
15-Aug-08	12-Dec-08	DFS Central Lab

LMS
JPD

INTOX EC/IR II Quality Assurance Worksheet

Instrument Serial Number: 010595 Worksheet Start Date: 4/30/2015

Location: Prince George County Police Department

Address: 6600 Courthouse Rd., Prince George, VA 23875

DFS Technician: Heather Stanton License No.: 18910

Laboratory On-Site

Site Specification: No detrimental environmental conditions exist.

Instrument Barometer (mm HG): 750 Reference Barometer (mm HG): 751

Reference Barometer(RB)Serial #: 009113 RB Calibration Due: 4/16/2016

Measurement Assurance Check

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.295
0.300	0.296	0.287	0.304	Sample 2	0.295
Precision		sample min	sample max	Sample 3	0.295
0		0.295	0.295		

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.097
0.100	0.099	0.096	0.102	Sample 2	0.097
Precision		sample min	sample max	Sample 3	0.098
0.001		0.097	0.098		


Dry gas standard Lot No. (with tank no.) AG320502-04

Replaced dry gas standard (+O-ring)

Installed at Location

Removed to DFS-Central

Supplies
Mouthpieces
Certificates of Analysis
Operator Worksheet
Other:



Notes:

Instrument Serial Number

010595

Certification Date

Calibrated

Certified

Measurement Assurance Check

Instrument Test

Troubleshooting/Maintenance Worksheet Completed

(Successful completion denotes satisfactory condition of the item.)

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.250		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.150		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.080		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.020		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

All measurements are in g/210L

Estimation of Uncertainty of Measurement and traceability records are located within the Breath Alcohol Section

DFS Technician

Heath A Stark

Date

5/8/15

Issuing Analyst

Jennie Duffy

Date

5/11/15

LMS
JPD



COMMONWEALTH OF VIRGINIA
DEPARTMENT OF FORENSIC SCIENCE

CERTIFICATE OF BLOOD ALCOHOL ANALYSIS
AS DETERMINED BY A CHEMICAL TEST OF THE ACCUSED'S BREATH

NAME OF ACCUSED INSTRUMENT, TEST,	NAME OF COURT DFS
---	-----------------------------

BREATH ANALYSIS

SAMPLE EXAMINED AND TEST CONDUCTED BY STANTON, HEATHER, M		AGENCY DFS Central Lab
DFS LICENSE NUMBER 18910	LICENSE EXPIRES 10/01/2016	DATE TEST CONDUCTED 04/30/2015
TEST EQUIPMENT NUMBER 010595		

RESULTS: TIME SAMPLE TAKEN 12:05 EDT

SAMPLE'S ALCOHOL CONTENT 0.00 GRAMS PER 210 LITERS OF BREATH

ATTEST:

I CERTIFY THAT THE ABOVE IS AN ACCURATE RECORD OF THE TEST CONDUCTED; THAT THE TEST WAS CONDUCTED WITH THE TYPE OF EQUIPMENT AND IN ACCORDANCE WITH THE METHODS APPROVED BY THE DEPARTMENT OF FORENSIC SCIENCE; THAT THE TEST WAS CONDUCTED IN ACCORDANCE WITH THE DEPARTMENT'S SPECIFICATIONS; THAT PRIOR TO ADMINISTRATION OF THE TEST THE ACCUSED WAS ADVISED OF HIS RIGHT TO OBSERVE THE PROCESS AND SEE THE BLOOD ALCOHOL READING ON THE EQUIPMENT USED TO PERFORM THE BREATH TEST, AND THAT I POSSESS A VALID LICENSE TO CONDUCT SUCH TEST, GIVEN UNDER MY HAND THIS _____ DAY OF _____, 20_____.

BREATH TEST OPERATOR

I HAVE RECEIVED A COPY OF THIS CERTIFICATE OF ANALYSIS _____ SUBJECT'S SIGNATURE

SUBJECT REFUSED TO SIGN FOR COPY OF CERTIFICATE OF ANALYSIS _____ OPERATOR'S SIGNATURE



IntoxNet MIS Report
Report Generated 30 Apr 2015 at 12:41

*LMS
JPD*

Test Results

Instrument Serial Number 010595

Test # 001692 Subject Test

Test Location 1 Department of
Test Date 30 Apr 2015

Test Location 2 Forensic Science
Test Time 11:59
Remote/Local Local

Test Location 3
System Check Passed

Operator's Last Name STANTON
Agency DFS Central Lab

Operator's First Name HEATHER

Operator's Middle Initial M

Card Serial Number 118910

License Number 18910

Effective Date 10/01/2014

Expiration Date 10/01/2016

Subject's Last Name INSTRUMENT

Subject's First Name TEST

Subject's Middle Initial

Subject's Date of Birth 00/00/0000

Subject's Sex Male

Driver's License Number

Driver's License Expiration 00/00/0000

Driver's License State

Court Name DFS

End Date 30 Apr 2015 End Time 12:06

Result Time 12:05

Result Date 30 Apr 2015 Result 0.00

Data Type DIAG

Sample Value Pass

Sample Time 11:59

Data Type BLK

Sample Value 0.000

Sample Time 12:00

Data Type CHK

Sample Value 0.098

Sample Time 12:01

Data Type BLK

Sample Value 0.000

Sample Time 12:02

Data Type SUBJ

Sample Value 0.000

Sample Time 12:03

Data Type BLK

Sample Value 0.000

Sample Time 12:04

Data Type SUBJ

Sample Value 0.000

Sample Time 12:05

Data Type BLK

Sample Value 0.000

Sample Time 12:06

Standard Type Dry Gas Std

Standard Value 0.099

Standard Lot Number AG320502-04

Standard Expiration Date 07/24/2015

Tank Pressure 463

Barometric Pressure 750 mmHg

Blow Sample Number 1 Blow Duration 3.17 sec

Blow Volume 1796 cc End-of-Blow Time 12:03

Blow Sample Number 2 Blow Duration 3.12 sec

Blow Volume 1750 cc End-of-Blow Time 12:05

Tamper Evident Stamp 3a8b09af

Test Status Code 0

Test Status Success

AMS
JPD

Intox EC/IR-II: Accuracy Check

Department of Forensic Science

Serial Number: 010595 Test Number: 1693

Test Date: 04/30/2015 Test Time: 12:07 EDT

Dry Gas Target: 0.099

Lot Number: AG320502-04 Exp Date: 07/24/2015

Tank Pressure: 463 psi Barometric Pressure: 750 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	12:08
CHK	0.097	12:08
BLK	0.000	12:10
CHK	0.097	12:11
BLK	0.000	12:12
CHK	0.098	12:13

Test Status: *Success*

Calibration CRC: 00C55F9A

LWS
JPD

Intox EC/IR-II: Accuracy Check

Department of Forensic Science

Serial Number: 010595 Test Number: 1694

Test Date: 04/30/2015 Test Time: 12:18 EDT

Dry Gas Target: 0.296

Lot Number: AG506501-09 Exp Date: 03/06/2017

Tank Pressure: 80 psi Barometric Pressure: 750 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	12:19
CHK	0.080	12:20

Test Status: *Out of tolerance*

Calibration CRC: 00C55F9A

LMS
JPD

Intox EC/IR-II: Accuracy Check

Department of Forensic Science

Serial Number: 010595 Test Number: 1695

Test Date: 04/30/2015 Test Time: 12:23 EDT

Dry Gas Target: 0.296

Lot Number: AG506501-08 Exp Date: 03/06/2017

Tank Pressure: 522 psi Barometric Pressure: 750 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	12:24
CHK	0.295	12:25
BLK	0.000	12:27
CHK	0.295	12:28
BLK	0.000	12:29
CHK	0.295	12:31

Test Status: *Success*

Calibration CRC: 00C55F9A