

Department of Forensic Science
Breath Alcohol Section
Instrument Location History
1/1/1990 To 6/26/2015

JMS
mmB

Instrument Serial Number: 010534

As of 26-Jun-15

Start Date	End Date	Agency Name
23-Jun-15		DFS Central Lab
21-Jan-15	23-Jun-15	Smyth County SO
15-Jan-15	21-Jan-15	DFS Central Lab
12-Aug-14	15-Jan-15	Arlington County PD
18-Jun-14	12-Aug-14	DFS Central Lab
24-Apr-12	18-Jun-14	Henry County Sheriff's Office
25-Jan-12	24-Apr-12	DFS Central Lab
20-Nov-08	25-Jan-12	Rockingham County SO
16-Jun-08	20-Nov-08	DFS Central Lab

**INTOX EC/IR II
Quality Assurance Worksheet**

AMS
mmB

Instrument Serial Number: 010534 Worksheet Start Date: 6/23/2015
 Location: Smyth County
 Address: 819 Matson Dr., Marion, VA 24354
 DFS Technician: Heather Stanton License No.: 18910
 Laboratory On-Site

Site Specification: No detrimental environmental conditions exist.

Instrument Barometer (mm HG): 705 Reference Barometer (mm HG): 708
 Reference Barometer(RB)Serial #: 009113 RB Calibration Due: 4/16/2016

Measurement Assurance Check

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.286
0.300	0.278	0.269	0.286	Sample 2	0.286
Precision		sample min	sample max	Sample 3	0.285
0.001		0.285	0.286		

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.095
0.100	0.093	0.090	0.096	Sample 2	0.095
Precision		sample min	sample max	Sample 3	0.095
0		0.095	0.095		

Dry gas standard Lot No. (with tank no.) AG406301-30

- Replaced dry gas standard (+O-ring)
 - Installed at Location
 - Removed to DFS-Central
- Supplies
 Mouthpieces
 Certificates of Analysis
 Operator Worksheet
 Other:

Notes:

Instrument Serial Number

010534

Certification Date

06/26/15

Calibrated

Certified

Measurement Assurance Check

Instrument Test

Troubleshooting/Maintenance Worksheet Completed

(Successful completion denotes satisfactory condition of the item.)

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.250		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.150		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.080		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.020		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

All measurements are in g/210L

Estimation of Uncertainty of Measurement and traceability records are located within the Breath Alcohol Section

DFS Technician

Heath W. Stark

Date

6/26/15

Issuing Analyst

Michelle Bell

Date

6/26/15

JPO



COMMONWEALTH OF VIRGINIA
DEPARTMENT OF FORENSIC SCIENCE

HMS
mmb

CERTIFICATE OF BLOOD ALCOHOL ANALYSIS
AS DETERMINED BY A CHEMICAL TEST OF THE ACCUSED'S BREATH

NAME OF ACCUSED INSTRUMENT, TEST,	NAME OF COURT DFS
--	--------------------------

BREATH ANALYSIS

SAMPLE EXAMINED AND TEST CONDUCTED BY STANTON, HEATHER, M		AGENCY DFS Central Lab
DFS LICENSE NUMBER 18910	LICENSE EXPIRES 10/01/2016	DATE TEST CONDUCTED 06/23/2015
TEST EQUIPMENT NUMBER 010534		

RESULTS: TIME SAMPLE TAKEN 12:40 EDT

SAMPLE'S ALCOHOL CONTENT 0.00 GRAMS PER 210 LITERS OF BREATH

ATTEST:

I CERTIFY THAT THE ABOVE IS AN ACCURATE RECORD OF THE TEST CONDUCTED; THAT THE TEST WAS CONDUCTED WITH THE TYPE OF EQUIPMENT AND IN ACCORDANCE WITH THE METHODS APPROVED BY THE DEPARTMENT OF FORENSIC SCIENCE; THAT THE TEST WAS CONDUCTED IN ACCORDANCE WITH THE DEPARTMENT'S SPECIFICATIONS; THAT PRIOR TO ADMINISTRATION OF THE TEST THE ACCUSED WAS ADVISED OF HIS RIGHT TO OBSERVE THE PROCESS AND SEE THE BLOOD ALCOHOL READING ON THE EQUIPMENT USED TO PERFORM THE BREATH TEST, AND THAT I POSSESS A VALID LICENSE TO CONDUCT SUCH TEST, GIVEN UNDER MY HAND THIS _____ DAY OF _____, 20_____.

BREATH TEST OPERATOR

I HAVE RECEIVED A COPY OF THIS CERTIFICATE OF ANALYSIS _____ SUBJECT'S SIGNATURE

SUBJECT REFUSED TO SIGN FOR COPY OF CERTIFICATE OF ANALYSIS _____ OPERATOR'S SIGNATURE

FOR COURT



IntoxNet MIS Report
Report Generated 24 Jun 2015 at 09:16

Handwritten initials: HWS
mmB

Test Results

Instrument Serial Number 010534

Test # 002184 Subject Test

Test Location 1 Department of
Test Date 23 Jun 2015

Test Location 2 Forensic Science
Test Time 12:34
Remote/Local Local

Test Location 3
System Check Passed

Operator's Last Name STANTON
Agency DFS Central Lab

Operator's First Name HEATHER

Operator's Middle Initial M

Card Serial Number 118910

License Number 18910

Effective Date 10/01/2014

Expiration Date 10/01/2016

Subject's Last Name INSTRUMENT

Subject's First Name TEST

Subject's Middle Initial

Subject's Date of Birth 00/00/0000

Subject's Sex Male

Driver's License Number

Driver's License Expiration 00/00/0000

Driver's License State

Court Name DFS

End Date 23 Jun 2015 End Time 12:41

Result Time 12:40

Result Date 23 Jun 2015 Result 0.00

Data Type DIAG

Sample Value Pass

Sample Time 12:35

Data Type BLK

Sample Value 0.000

Sample Time 12:35

Data Type CHK

Sample Value 0.095

Sample Time 12:36

Data Type BLK

Sample Value 0.000

Sample Time 12:37

Data Type SUBJ

Sample Value 0.000

Sample Time 12:38

Data Type BLK

Sample Value 0.000

Sample Time 12:39

Data Type SUBJ

Sample Value 0.000

Sample Time 12:40

Data Type BLK

Sample Value 0.000

Sample Time 12:41

Standard Type Dry Gas Std

Standard Value 0.093

Standard Lot Number AG406301-30

Standard Expiration Date 03/04/2016

Tank Pressure 899

Barometric Pressure 705 mmHg

Blow Sample Number 1 Blow Duration 3.53 sec

Blow Volume 1756 cc End-of-Blow Time 12:38

Blow Sample Number 2 Blow Duration 3.57 sec

Blow Volume 1735 cc End-of-Blow Time 12:40

Tamper Evident Stamp 26f5fa71

Test Status Code 0

Test Status Success

~~MMB~~
mmB

Intox EC/IR-II: Accuracy Check

Department of Forensic Science

Serial Number: 010534 Test Number: 2185
Test Date: 06/23/2015 Test Time: 12:44 EDT

Dry Gas Target: 0.093

Lot Number: AG406301-30 Exp Date: 03/04/2016

Tank Pressure: 889 psi Barometric Pressure: 705 mmHg
System Check: *Passed*

Test	g/210L	Time
BLK	0.000	12:44
CHK	0.095	12:45
BLK	0.000	12:47
CHK	0.095	12:47
BLK	0.000	12:49
CHK	0.095	12:49

Test Status: *Success*

Calibration CRC: 90196AE6

HMS
MMB

Intox EC/IR-II: Accuracy Check

Department of Forensic Science

Serial Number: 010534 Test Number: 2186

Test Date: 06/23/2015 Test Time: 12:52 EDT

Dry Gas Target: 0.278

Lot Number: AG506501-11 Exp Date: 03/06/2017

Tank Pressure: 492 psi Barometric Pressure: 705 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	12:52
CHK	0.286	12:53
BLK	0.000	12:55
CHK	0.286	12:56
BLK	0.000	12:57
CHK	0.285	12:58

Test Status: *Success*

Calibration CRC: 90196AE6