

mm B

Department of Forensic Science  
Breath Alcohol Section  
Instrument Location History  
1/1/1990 To 5/29/2015

Instrument Serial Number: 010537

As of 29-May-15

Start Date	End Date	Agency Name
28-May-15		Hampton PD
15-May-15	28-May-15	DFS Central Lab
09-Dec-14	15-May-15	Stafford County SO
19-Nov-14	09-Dec-14	DFS Central Lab
12-Jun-14	19-Nov-14	Prince George County PD
30-May-14	12-Jun-14	DFS Central Lab
03-Apr-14	30-May-14	Bedford County SO
14-Mar-14	03-Apr-14	DFS Central Lab
03-May-10	14-Mar-14	Portsmouth PD
27-Apr-10	03-May-10	DFS Central Lab
12-Mar-09	27-Apr-10	Portsmouth PD
16-Jun-08	12-Mar-09	DFS Central Lab

# INTOX EC/IR II Quality Assurance Worksheet

Instrument Serial Number 010537 Worksheet Start Date 5/28/2015

Location Hampton Jail

Address 236 N King St Hampton VA 23669

DFS Technician Michelle Billeter License No. 26417

Laboratory  On-Site

Site Specification: No detrimental environmental conditions exist.

Instrument Barometer (mm HG) 763 Reference Barometer (mm HG) 766

Reference Barometer(RB)Serial # 009111 RB Calibration Due 5/18/2016

### Measurement Assurance Check


Standard (sea level)	PA Target	minimum	maximum	Sample 1	Sample 2	Sample 3
0.300	0.301	0.291	0.310	0.305	0.305	0.305
<b>Precision</b>		<b>sample min</b>	<b>sample max</b>			
0		0.305	0.305			

Standard (sea level)	PA Target	minimum	maximum	Sample 1	Sample 2	Sample 3
0.100	0.100	0.097	0.103	0.100	0.100	0.100
<b>Precision</b>		<b>sample min</b>	<b>sample max</b>			
0		0.100	0.100			

Dry gas standard Lot No. (with tank no.) AG407801-11

- Replaced dry gas standard (+O-ring)
- Installed at Location
- Removed to DFS-Central

- Supplies
- Mouthpieces
- Certificates of Analysis
- Operator Worksheet
- Other:



Notes:

Instrument Serial Number

11010537

Certification Date

05/29/15

Calibrated

Certified

Measurement Assurance Check

Instrument Test

Troubleshooting/Maintenance Worksheet Completed

(Successful completion denotes satisfactory condition of the item.)

<b>Standard (sea level)</b>	<b>PA Target</b>	<b>minimum</b>	<b>maximum</b>	<b>Sample 1</b>
0.250		0.000	0.000	<b>Sample 2</b>
<b>Precision</b>		<b>sample min</b>	<b>sample max</b>	<b>Sample 3</b>
0		0.000	0.000	

<b>Standard (sea level)</b>	<b>PA Target</b>	<b>minimum</b>	<b>maximum</b>	<b>Sample 1</b>
0.150		0.000	0.000	<b>Sample 2</b>
<b>Precision</b>		<b>sample min</b>	<b>sample max</b>	<b>Sample 3</b>
0		0.000	0.000	

<b>Standard (sea level)</b>	<b>PA Target</b>	<b>minimum</b>	<b>maximum</b>	<b>Sample 1</b>
0.080		-0.003	0.003	<b>Sample 2</b>
<b>Precision</b>		<b>sample min</b>	<b>sample max</b>	<b>Sample 3</b>
0		0.000	0.000	

<b>Standard (sea level)</b>	<b>PA Target</b>	<b>minimum</b>	<b>maximum</b>	<b>Sample 1</b>
0.020		-0.003	0.003	<b>Sample 2</b>
<b>Precision</b>		<b>sample min</b>	<b>sample max</b>	<b>Sample 3</b>
0		0.000	0.000	

All measurements are in g/210L

Estimation of Uncertainty of Measurement and traceability records are located within the Breath Alcohol Section

DFS Technician

*Michelle Bellis*

Date

5/29/15

Issuing Analyst

*Michelle Bellis*

Date

5/29/15

*JPP*

mmb

**Intox EC/IR-II: Accuracy Check**

*Department of Forensic Science*

Serial Number: 010537      Test Number: 1808

Test Date: 05/28/2015      Test Time: 14:06 EDT

Dry Gas Target: 0.301

Lot Number: AG506501-06      Exp Date: 03/06/2017

Tank Pressure: 438 psi      Barometric Pressure: 763 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	14:07
CHK	0.305	14:08
BLK	0.000	14:10
CHK	0.305	14:11
BLK	0.000	14:12
CHK	0.305	14:13

Test Status: *Success*

Calibration CRC: 6B8E7224

mms

**Intox EC/IR-II: Accuracy Check**

*Department of Forensic Science*

Serial Number: 010537      Test Number: 1809

Test Date: 05/28/2015      Test Time: 14:16 EDT

Dry Gas Target: 0.100

Lot Number: AG407801-11      Exp Date: 03/19/2016

Tank Pressure: 497 psi      Barometric Pressure: 763 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	14:17
CHK	0.100	14:18
BLK	0.000	14:19
CHK	0.100	14:20
BLK	0.000	14:22
CHK	0.100	14:22

Test Status: *Success*

Calibration CRC: 6B8E7224

mmb



COMMONWEALTH OF VIRGINIA  
DEPARTMENT OF FORENSIC SCIENCE

CERTIFICATE OF BLOOD ALCOHOL ANALYSIS  
AS DETERMINED BY A CHEMICAL TEST OF THE ACCUSED'S BREATH

NAME OF ACCUSED <b>INSTRUMENT, TEST,</b>	NAME OF COURT <b>DFS</b>
---	-----------------------------

BREATH ANALYSIS

SAMPLE EXAMINED AND TEST CONDUCTED BY <b>BILLETER, MICHELLE, M</b>		AGENCY <b>DFS Central Lab</b>
DFS LICENSE NUMBER <b>26417</b>	LICENSE EXPIRES <b>10/01/2016</b>	DATE TEST CONDUCTED <b>05/28/2015</b>
TEST EQUIPMENT NUMBER <b>010537</b>		

RESULTS: TIME SAMPLE TAKEN 14:30 EDT

SAMPLE'S ALCOHOL CONTENT 0.00 GRAMS PER 210 LITERS OF BREATH

ATTEST:

I CERTIFY THAT THE ABOVE IS AN ACCURATE RECORD OF THE TEST CONDUCTED; THAT THE TEST WAS CONDUCTED WITH THE TYPE OF EQUIPMENT AND IN ACCORDANCE WITH THE METHODS APPROVED BY THE DEPARTMENT OF FORENSIC SCIENCE; THAT THE TEST WAS CONDUCTED IN ACCORDANCE WITH THE DEPARTMENT'S SPECIFICATIONS; THAT PRIOR TO ADMINISTRATION OF THE TEST THE ACCUSED WAS ADVISED OF HIS RIGHT TO OBSERVE THE PROCESS AND SEE THE BLOOD ALCOHOL READING ON THE EQUIPMENT USED TO PERFORM THE BREATH TEST, AND THAT I POSSESS A VALID LICENSE TO CONDUCT SUCH TEST, GIVEN UNDER MY HAND THIS \_\_\_\_\_ DAY OF \_\_\_\_\_, 20\_\_\_\_\_.

BREATH TEST OPERATOR
----------------------

I HAVE RECEIVED A COPY OF THIS CERTIFICATE OF ANALYSIS \_\_\_\_\_ SUBJECT'S SIGNATURE

SUBJECT REFUSED TO SIGN FOR COPY OF CERTIFICATE OF ANALYSIS \_\_\_\_\_ OPERATOR'S SIGNATURE

mmb



# IntoxNet MIS Report

Report Generated 29 May 2015 at 09:08

## Test Results

**Instrument Serial Number 010537**

**Test # 001810**      **Subject Test**

**Test Location 1** Department of  
Test Date 28 May 2015

**Test Location 2** Forensic Science  
Test Time 14:24  
Remote/Local Local

**Test Location 3**  
System Check Passed

**Operator's Last Name** BILLETER  
Agency DFS Central Lab

**Operator's First Name** MICHELLE

**Operator's Middle Initial** M

**Card Serial Number** 126417

**License Number** 26417

**Effective Date** 10/01/2014

**Expiration Date** 10/01/2016

**Subject's Last Name** INSTRUMENT

**Subject's First Name** TEST

**Subject's Middle Initial**

**Subject's Date of Birth** 00/00/0000

**Subject's Sex** Male

**Driver's License Number**

**Driver's License Expiration** 00/00/0000

**Driver's License State**

**Court Name** DFS

**End Date** 28 May 2015      **End Time** 14:31

**Result Time** 14:30

**Result Date** 28 May 2015      **Result** 0.00

**Data Type** DIAG

**Sample Value** Pass

**Sample Time** 14:24

**Data Type** BLK

**Sample Value** 0.000

**Sample Time** 14:25

**Data Type** CHK

**Sample Value** 0.100

**Sample Time** 14:25

**Data Type** BLK

**Sample Value** 0.000

**Sample Time** 14:27

**Data Type** SUBJ

**Sample Value** 0.000

**Sample Time** 14:28

**Data Type** BLK

**Sample Value** 0.000

**Sample Time** 14:29

**Data Type** SUBJ

**Sample Value** 0.000

**Sample Time** 14:30

**Data Type** BLK

**Sample Value** 0.000

**Sample Time** 14:31

**Standard Type** Dry Gas Std

**Standard Value** 0.100

**Standard Lot Number** AG407801-11

**Standard Expiration Date** 03/19/2016

**Tank Pressure** 492

**Barometric Pressure** 763 mmHg

**Blow Sample Number** 1      **Blow Duration** 4.30 sec

**Blow Volume** 1867 cc      **End-of-Blow Time** 14:28

**Blow Sample Number** 2      **Blow Duration** 3.83 sec

**Blow Volume** 1739 cc      **End-of-Blow Time** 14:30

**Tamper Evident Stamp** 9835778d

**Test Status Code** 0

**Test Status** Success