

DBH
Jed

Department of Forensic Science
Breath Alcohol Section
Instrument Location History
1/1/1990 To 6/17/2015

Instrument Serial Number: 010554

As of 17-Jun-15

Start Date	End Date	Agency Name
16-Jun-15		DFS Central Lab
22-Jan-15	16-Jun-15	Lee County Sheriff's Office
06-Jan-15	22-Jan-15	DFS Central Lab
22-Jul-14	06-Jan-15	Caroline County SO
10-Jul-14	22-Jul-14	DFS Central Lab
06-Mar-09	10-Jul-14	Smithfield PD
23-Jun-08	06-Mar-09	DFS Central Lab

DBH
Sed

INTOX EC/IR II Quality Assurance Worksheet

Instrument Serial Number: 010554 Worksheet Start Date: 6/16/2015

Location: Lee CO SO

Address: 33640 Main St Jonesville VA 24263

DFS Technician: Donald Hall License No.: 21250

Laboratory On-Site

Site Specification: No detrimental environmental conditions exist.

Instrument Barometer (mm HG): 724 Reference Barometer (mm HG): 725

Reference Barometer(RB)Serial #: 009109 RB Calibration Due: 7/28/2015

Measurement Assurance Check

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.287
0.300	0.286	0.277	0.294	Sample 2	0.287
Precision		sample min	sample max	Sample 3	0.286
0.001		0.286	0.287		

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.095
0.100	0.095	0.092	0.098	Sample 2	0.095
Precision		sample min	sample max	Sample 3	0.095
0		0.095	0.095		


Dry gas standard Lot No. (with tank no.) AG400202-31

Replaced dry gas standard (+O-ring)

Installed at Location

Removed to DFS-Central

Supplies
Mouthpieces
Certificates of Analysis
Operator Worksheet
Other:



Notes:

Instrument Serial Number

010554

Certification Date

Calibrated

Certified

Measurement Assurance Check

Instrument Test

Troubleshooting/Maintenance Worksheet Completed

(Successful completion denotes satisfactory condition of the item.)

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.250		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.150		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.080		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.020		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

All measurements are in g/210L

Estimation of Uncertainty of Measurement and traceability records are located within the Breath Alcohol Section

DFS Technician

Donald B Hall

Date

6/17/15

Issuing Analyst

Jed
TLW

Date

6/19/15

DBH
Sud



COMMONWEALTH OF VIRGINIA
DEPARTMENT OF FORENSIC SCIENCE

CERTIFICATE OF BLOOD ALCOHOL ANALYSIS
AS DETERMINED BY A CHEMICAL TEST OF THE ACCUSED'S BREATH

NAME OF ACCUSED INSTRUMENT, TEST,	NAME OF COURT DFS
---	-----------------------------

BREATH ANALYSIS

SAMPLE EXAMINED AND TEST CONDUCTED BY HALL, DONALD, B		AGENCY DFS Central Lab
DFS LICENSE NUMBER 21250	LICENSE EXPIRES 10/01/2016	DATE TEST CONDUCTED 06/16/2015
TEST EQUIPMENT NUMBER 010554	[REDACTED]	

RESULTS: TIME SAMPLE TAKEN 11:42 EDT

SAMPLE'S ALCOHOL CONTENT 0.00 GRAMS PER 210 LITERS OF BREATH

ATTEST:

I CERTIFY THAT THE ABOVE IS AN ACCURATE RECORD OF THE TEST CONDUCTED; THAT THE TEST WAS CONDUCTED WITH THE TYPE OF EQUIPMENT AND IN ACCORDANCE WITH THE METHODS APPROVED BY THE DEPARTMENT OF FORENSIC SCIENCE; THAT THE TEST WAS CONDUCTED IN ACCORDANCE WITH THE DEPARTMENT'S SPECIFICATIONS; THAT PRIOR TO ADMINISTRATION OF THE TEST THE ACCUSED WAS ADVISED OF HIS RIGHT TO OBSERVE THE PROCESS AND SEE THE BLOOD ALCOHOL READING ON THE EQUIPMENT USED TO PERFORM THE BREATH TEST, AND THAT I POSSESS A VALID LICENSE TO CONDUCT SUCH TEST, GIVEN UNDER MY HAND THIS _____ DAY OF _____, 20 _____.

BREATH TEST OPERATOR

I HAVE RECEIVED A COPY OF THIS CERTIFICATE OF ANALYSIS _____ SUBJECT'S SIGNATURE

SUBJECT REFUSED TO SIGN FOR COPY OF CERTIFICATE OF ANALYSIS _____ OPERATOR'S SIGNATURE



IntoxNet MIS Report

Report Generated 16 Jun 2015 at 12:03

JPH
ml

Test Results

Instrument Serial Number 010554

Test # 001073 Subject Test

Test Location 1 Department of
Test Date 16 Jun 2015

Test Location 2 Forensic Science
Test Time 11:35
Remote/Local Local

Test Location 3
System Check Passed

Operator's Last Name HALL
Agency DFS Central Lab

Operator's First Name DONALD

Operator's Middle Initial B

Card Serial Number 121250

License Number 21250

Effective Date 10/01/2014

Expiration Date 10/01/2016

Subject's Last Name INSTRUMENT

Subject's First Name TEST

Subject's Middle Initial

Subject's Date of Birth 00/00/0000

Subject's Sex Male

Driver's License Number

Driver's License Expiration 00/00/0000

Driver's License State

Court Name DFS

End Date 16 Jun 2015 End Time 11:43

Result Time 11:42

Result Date 16 Jun 2015

Result 0.00

Data Type DIAG

Sample Value Pass

Sample Time 11:36

Data Type BLK

Sample Value 0.000

Sample Time 11:37

Data Type CHK

Sample Value 0.096

Sample Time 11:37

Data Type BLK

Sample Value 0.000

Sample Time 11:39

Data Type SUBJ

Sample Value 0.000

Sample Time 11:39

Data Type BLK

Sample Value 0.000

Sample Time 11:40

Data Type SUBJ

Sample Value 0.000

Sample Time 11:42

Data Type BLK

Sample Value 0.000

Sample Time 11:43

Standard Type Dry Gas Std

Standard Value 0.095

Standard Lot Number AG400202-31

Standard Expiration Date 01/02/2016

Tank Pressure 825

Barometric Pressure 724 mmHg

Blow Sample Number 1 Blow Duration 3.64 sec

Blow Volume 1931 cc End-of-Blow Time 11:39

Blow Sample Number 2 Blow Duration 3.70 sec

Blow Volume 1822 cc End-of-Blow Time 11:42

Tamper Evident Stamp 709e0c30

Test Status Code 0

Test Status Success

105H

nd

INTOX EC/IR-11: Accuracy Check

Department of Forensic Science

Serial Number: 010554 Test Number: 1074

Test Date: 06/16/2015 Test Time: 11:44 EDT

Dry Gas Target: 0.095

Lot Number: AG400202-31 Exp Date: 01/02/2016

Tank Pressure: 821 psi Barometric Pressure: 724 mmHg

System Check: Passed

Test	g/210L	Time
BLK	0.000	11:45
CHK	0.095	11:46
BLK	0.000	11:48
CHK	0.095	11:48
BLK	0.000	11:50
CHK	0.095	11:50

Test Status: Success

Calibration CRC: 91E4234E

DBT
Ad

Intox EC/IR-11: Accuracy Check

Department of Forensic Science

Serial Number: 010554 Test Number: 1075

Test Date: 06/16/2015 Test Time: 11:53 EDT

Dry Gas Target: 0.286

Lot Number: AG506501-12 Exp Date: 03/06/2017

Tank Pressure: 232 psi Barometric Pressure: 724 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	11:53
CHK	0.287	11:54
BLK	0.000	11:56
CHK	0.287	11:56
BLK	0.000	11:58
CHK	0.286	11:59

Test Status: *Success*

Calibration CRC: 91E4234E