

Department of Forensic Science  
Breath Alcohol Section  
Instrument Location History  
1/1/1990 To 6/10/2015

*AMS*  
*WBL*

Instrument Serial Number: 010560

As of 10-Jun-15

Start Date	End Date	Agency Name
10-Jun-15		DFS Central Lab
04-Mar-15	10-Jun-15	Prince William County PD
18-Feb-15	04-Mar-15	DFS Central Lab
10-Sep-14	18-Feb-15	Carroll County SO
20-Aug-14	10-Sep-14	DFS Central Lab
12-Mar-14	20-Aug-14	Bristol PD
06-Mar-14	12-Mar-14	DFS Central Lab
01-Apr-13	06-Mar-14	Arlington County PD
25-Feb-13	01-Apr-13	DFS Central Lab
14-Apr-09	25-Feb-13	Arlington County PD
23-Jun-08	14-Apr-09	DFS Central Lab

# INTOX EC/IR II Quality Assurance Worksheet

*HMS*  
*MSK*

Instrument Serial Number: 010560      Worksheet Start Date: 6/10/2015

Location: Prince William County Adult Detention Center

Address: 9320 Lee Ave., Manassas, VA 20110

DFS Technician: Heather Stanton      License No.: 18910

Laboratory       On-Site

Site Specification: No detrimental environmental conditions exist.

Instrument Barometer (mm HG): 751      Reference Barometer (mm HG): 753

Reference Barometer(RB)Serial #: 009113      RB Calibration Due: 4/16/2016

Measurement Assurance Check

<b>Standard (sea level)</b>	<b>PA Target</b>	<b>minimum</b>	<b>maximum</b>	<b>Sample 1</b>
0.300		-0.003	0.003	<b>Sample 2</b>
<b>Precision</b>		<b>sample min</b>	<b>sample max</b>	<b>Sample 3</b>
0		0.000	0.000	

<b>Standard (sea level)</b>	<b>PA Target</b>	<b>minimum</b>	<b>maximum</b>	<b>Sample 1</b>
0.100		-0.003	0.003	<b>Sample 2</b>
<b>Precision</b>		<b>sample min</b>	<b>sample max</b>	<b>Sample 3</b>
0		0.000	0.000	


Dry gas standard Lot No. (with tank no.): AG404902-06

Replaced dry gas standard (+O-ring)

Installed at Location

Removed to DFS-Central

Supplies  
Mouthpieces  
Certificates of Analysis  
Operator Worksheet  
Other:



Notes:

Unable to perform instrument test and measurement assurance checks due to instrument being disabled.

Instrument Serial Number

010560

Certification Date

06/10/15

Calibrated

Certified

Measurement Assurance Check

Instrument Test

Troubleshooting/Maintenance Worksheet Completed

(Successful completion denotes satisfactory condition of the item.)

<b>Standard (sea level)</b>	<b>PA Target</b>	<b>minimum</b>	<b>maximum</b>	<b>Sample 1</b>
0.250		0.000	0.000	<b>Sample 2</b>
<b>Precision</b>		<b>sample min</b>	<b>sample max</b>	<b>Sample 3</b>
0		0.000	0.000	

<b>Standard (sea level)</b>	<b>PA Target</b>	<b>minimum</b>	<b>maximum</b>	<b>Sample 1</b>
0.150		0.000	0.000	<b>Sample 2</b>
<b>Precision</b>		<b>sample min</b>	<b>sample max</b>	<b>Sample 3</b>
0		0.000	0.000	

<b>Standard (sea level)</b>	<b>PA Target</b>	<b>minimum</b>	<b>maximum</b>	<b>Sample 1</b>
0.080		-0.003	0.003	<b>Sample 2</b>
<b>Precision</b>		<b>sample min</b>	<b>sample max</b>	<b>Sample 3</b>
0		0.000	0.000	

<b>Standard (sea level)</b>	<b>PA Target</b>	<b>minimum</b>	<b>maximum</b>	<b>Sample 1</b>
0.020		-0.003	0.003	<b>Sample 2</b>
<b>Precision</b>		<b>sample min</b>	<b>sample max</b>	<b>Sample 3</b>
0		0.000	0.000	

All measurements are in g/210L

Estimation of Uncertainty of Measurement and traceability records are located within the Breath Alcohol Section

DFS Technician

*Heath W. Smith*

Date

*6/10/15*

Issuing Analyst

*Melissa Kennedy*

Date

*6/11/2015*