

Department of Forensic Science
Breath Alcohol Section
Instrument Location History
1/1/1990 To 6/9/2015

JMS
JPD

Instrument Serial Number: 010562

As of 09-Jun-15

Start Date	End Date	Agency Name
09-Jun-15		Franklin PD
19-May-15	09-Jun-15	DFS Central Lab
16-Dec-14	19-May-15	Henrico County PD
04-Dec-14	16-Dec-14	DFS Central Lab
26-Jun-14	04-Dec-14	Culpeper PD
11-Jun-14	26-Jun-14	DFS Central Lab
05-Feb-13	11-Jun-14	New River Valley Regional Jail
26-Apr-12	05-Feb-13	DFS Central Lab
12-Mar-09	26-Apr-12	Norton PD
23-Jun-08	12-Mar-09	DFS Central Lab

Handwritten initials: JPD

INTOX EC/IR II Quality Assurance Worksheet

Instrument Serial Number: 010562 Worksheet Start Date: 6/9/2015

Location: Franklin Police Department

Address: 1018 Pretlow St., Franklin, VA 23851

DFS Technician: Heather Stanton License No.: 18910

Laboratory On-Site

Site Specification: No detrimental environmental conditions exist.

Instrument Barometer (mm HG): 754 Reference Barometer (mm HG): 757

Reference Barometer(RB)Serial #: 009113 RB Calibration Due: 4/16/2016

Measurement Assurance Check


Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.301
0.300	0.298	0.289	0.306	Sample 2	0.301
Precision		sample min	sample max	Sample 3	0.300
0.001		0.300	0.301		

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.099
0.100	0.099	0.096	0.102	Sample 2	0.099
Precision		sample min	sample max	Sample 3	0.099
0		0.099	0.099		

Dry gas standard Lot No. (with tank no.) AG509101-36

- Replaced dry gas standard (+O-ring)
- Installed at Location
- Removed to DFS-Central

- Supplies
- Mouthpieces
- Certificates of Analysis
- Operator Worksheet
- Other:



Notes:

Instrument Serial Number

010562

Certification Date

Calibrated

Certified

Measurement Assurance Check

Instrument Test

Troubleshooting/Maintenance Worksheet Completed

(Successful completion denotes satisfactory condition of the item.)

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.250		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.150		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.080		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.020		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

All measurements are in g/210L

Estimation of Uncertainty of Measurement and traceability records are located within the Breath Alcohol Section

DFS Technician Heath W. Stach Date 6/9/15

Issuing Analyst Jennie Duffy Date 6/10/15

tho
JPD

Intox EC/IR-II: Accuracy Check

Department of Forensic Science

Serial Number: 010562 Test Number: 1158

Test Date: 06/09/2015 Test Time: 10:49 EDT

Dry Gas Target: 0.298

Lot Number: AG506501-08 Exp Date: 03/06/2017

Tank Pressure: 188 psi Barometric Pressure: 754 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	10:49
CHK	0.301	10:50
BLK	0.000	10:52
CHK	0.301	10:53
BLK	0.000	10:54
CHK	0.300	10:55

Test Status: *Success*

Calibration CRC: CE6A8082

HMS
JPD

intox EC/IR-II: Accuracy Check

Department of Forensic Science

Serial Number: 010562 Test Number: 1159

Test Date: 06/09/2015 Test Time: 10:59 EDT

Dry Gas Target: 0.099

Lot Number: AG509101-36 Exp Date: 04/01/2017

Tank Pressure: 1115 psi Barometric Pressure: 754 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	11:00
CHK	0.099	11:00
BLK	0.000	11:02
CHK	0.099	11:03
BLK	0.000	11:04
CHK	0.099	11:05

Test Status: *Success*

Calibration CRC: CE6A8082

HMS
JPD



COMMONWEALTH OF VIRGINIA
DEPARTMENT OF FORENSIC SCIENCE

CERTIFICATE OF BLOOD ALCOHOL ANALYSIS
AS DETERMINED BY A CHEMICAL TEST OF THE ACCUSED'S BREATH

NAME OF ACCUSED INSTRUMENT, TEST,	NAME OF COURT DFS
---	-----------------------------

BREATH ANALYSIS

SAMPLE EXAMINED AND TEST CONDUCTED BY STANTON, HEATHER, M		AGENCY DFS Central Lab
DFS LICENSE NUMBER 18910	LICENSE EXPIRES 10/01/2016	DATE TEST CONDUCTED 06/09/2015
TEST EQUIPMENT NUMBER 010562	[REDACTED]	

RESULTS: TIME SAMPLE TAKEN 11:12 EDT

SAMPLE'S ALCOHOL CONTENT 0.00 GRAMS PER 210 LITERS OF BREATH

ATTEST:

I CERTIFY THAT THE ABOVE IS AN ACCURATE RECORD OF THE TEST CONDUCTED; THAT THE TEST WAS CONDUCTED WITH THE TYPE OF EQUIPMENT AND IN ACCORDANCE WITH THE METHODS APPROVED BY THE DEPARTMENT OF FORENSIC SCIENCE; THAT THE TEST WAS CONDUCTED IN ACCORDANCE WITH THE DEPARTMENT'S SPECIFICATIONS; THAT PRIOR TO ADMINISTRATION OF THE TEST THE ACCUSED WAS ADVISED OF HIS RIGHT TO OBSERVE THE PROCESS AND SEE THE BLOOD ALCOHOL READING ON THE EQUIPMENT USED TO PERFORM THE BREATH TEST, AND THAT I POSSESS A VALID LICENSE TO CONDUCT SUCH TEST, GIVEN UNDER MY HAND THIS _____ DAY OF _____, 20_____.

BREATH TEST OPERATOR

I HAVE RECEIVED A COPY OF THIS CERTIFICATE OF ANALYSIS _____ SUBJECT'S SIGNATURE

SUBJECT REFUSED TO SIGN FOR COPY OF CERTIFICATE OF ANALYSIS _____ OPERATOR'S SIGNATURE

FOR COURT



IntoxNet MIS Report
Report Generated 09 Jun 2015 at 14:58

Handwritten initials: HNS, JPD

Test Results

Instrument Serial Number 010562

Test # 001160 Subject Test
Test Location 1 Department of
Test Date 09 Jun 2015

Test Location 2 Forensic Science
Test Time 11:06
Remote/Local Local

Test Location 3
System Check Passed

Operator's Last Name STANTON
Agency DFS Central Lab
Card Serial Number 118910
Subject's Last Name INSTRUMENT
Subject's Middle Initial
Driver's License Number
Driver's License State

Operator's First Name HEATHER
Operator's Middle Initial M
License Number 18910
Effective Date 10/01/2014
Expiration Date 10/01/2016
Subject's First Name TEST
Subject's Date of Birth 00/00/0000
Subject's Sex Male
Driver's License Expiration 00/00/0000
Court Name DFS

End Date 09 Jun 2015 End Time 11:13
Data Type DIAG
Data Type BLK
Data Type CHK
Data Type BLK
Data Type SUBJ
Data Type BLK
Data Type SUBJ
Data Type BLK

Result Time 11:12
Sample Value Pass
Sample Value 0.000
Sample Value 0.099
Sample Value 0.000
Sample Value 0.000
Sample Value 0.000
Sample Value 0.000
Sample Value 0.000

Result Date 09 Jun 2015 Result 0.00
Sample Time 11:06
Sample Time 11:07
Sample Time 11:08
Sample Time 11:09
Sample Time 11:10
Sample Time 11:11
Sample Time 11:12
Sample Time 11:13

Standard Type Dry Gas Std
Standard Lot Number AG509101-36
Tank Pressure 1115

Standard Value 0.099
Standard Expiration Date 04/01/2017
Barometric Pressure 754 mmHg

Blow Sample Number 1 Blow Duration 3.27 sec
Blow Sample Number 2 Blow Duration 3.40 sec
Tamper Evident Stamp d11347d2
Test Status Success

Blow Volume 1767 cc End-of-Blow Time 11:10
Blow Volume 1780 cc End-of-Blow Time 11:12
Test Status Code 0