

DBH
JPO

Department of Forensic Science
Breath Alcohol Section
Instrument Location History
1/1/1990 To 6/12/2015

Instrument Serial Number: 010614

As of 12-Jun-15

Start Date	End Date	Agency Name
11-Jun-15		DFS Central Lab
15-Jan-15	11-Jun-15	Blackstone PD
23-Dec-14	15-Jan-15	DFS Central Lab
30-Jul-14	23-Dec-14	Richmond PD
16-Jul-14	30-Jul-14	DFS Central Lab
23-Feb-12	16-Jul-14	Loudoun County SO
24-Oct-11	23-Feb-12	DFS Central Lab
15-Dec-08	24-Oct-11	Goochland County Sheriff's Office
04-Aug-08	15-Dec-08	DFS Central Lab

DBH
JPD

INTOX EC/IR II Quality Assurance Worksheet

Instrument Serial Number: 010614 Worksheet Start Date: 6/11/2015

Location: Blackstone PD

Address: 101 W Elm St Blackstone VA 23824

DFS Technician: Donald Hall License No.: 21250

Laboratory On-Site

Site Specification: No detrimental environmental conditions exist.

Instrument Barometer (mm HG): 748 Reference Barometer (mm HG): 750

Reference Barometer(RB)Serial #: 009109 RB Calibration Due: 7/28/2015

Measurement Assurance Check


Standard (sea level)	PA Target	minimum	maximum	Sample 1	Sample 2	Sample 3
0.300	0.295	0.286	0.303	0.297	0.297	0.297
Precision		sample min	sample max			
0		0.297	0.297			

Standard (sea level)	PA Target	minimum	maximum	Sample 1	Sample 2	Sample 3
0.100	0.098	0.095	0.101	0.098	0.098	0.098
Precision		sample min	sample max			
0		0.098	0.098			

Dry gas standard Lot No. (with tank no.) AG332301-10

- Replaced dry gas standard (+O-ring)
- Installed at Location
- Removed to DFS-Central

- Supplies
- Mouthpieces
- Certificates of Analysis
- Operator Worksheet
- Other:



Notes:

Instrument Serial Number

010614

Certification Date

Calibrated

Certified

Measurement Assurance Check

Instrument Test

Troubleshooting/Maintenance Worksheet Completed

(Successful completion denotes satisfactory condition of the item.)

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.250		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.150		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.080		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.020		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

All measurements are in g/210L

Estimation of Uncertainty of Measurement and traceability records are located within the Breath Alcohol Section

DFS Technician

Donald B Hall

Date

6/12/15

Issuing Analyst

Jennie Duffy

Date

6/12/15

nmb

BBH
JPP



COMMONWEALTH OF VIRGINIA
DEPARTMENT OF FORENSIC SCIENCE

CERTIFICATE OF BLOOD ALCOHOL ANALYSIS
AS DETERMINED BY A CHEMICAL TEST OF THE ACCUSED'S BREATH

NAME OF ACCUSED INSTRUMENT, TEST,	NAME OF COURT DFS
---------------------------------------------	-----------------------------

BREATH ANALYSIS

SAMPLE EXAMINED AND TEST CONDUCTED BY HALL, DONALD, B		AGENCY DFS Central Lab
DFS LICENSE NUMBER 21250	LICENSE EXPIRES 10/01/2016	DATE TEST CONDUCTED 06/11/2015
TEST EQUIPMENT NUMBER 010614		

RESULTS: TIME SAMPLE TAKEN 09:05 EDT

SAMPLE'S ALCOHOL CONTENT 0.00 GRAMS PER 210 LITERS OF BREATH

ATTEST:

I CERTIFY THAT THE ABOVE IS AN ACCURATE RECORD OF THE TEST CONDUCTED; THAT THE TEST WAS CONDUCTED WITH THE TYPE OF EQUIPMENT AND IN ACCORDANCE WITH THE METHODS APPROVED BY THE DEPARTMENT OF FORENSIC SCIENCE; THAT THE TEST WAS CONDUCTED IN ACCORDANCE WITH THE DEPARTMENT'S SPECIFICATIONS; THAT PRIOR TO ADMINISTRATION OF THE TEST THE ACCUSED WAS ADVISED OF HIS RIGHT TO OBSERVE THE PROCESS AND SEE THE BLOOD ALCOHOL READING ON THE EQUIPMENT USED TO PERFORM THE BREATH TEST, AND THAT I POSSESS A VALID LICENSE TO CONDUCT SUCH TEST, GIVEN UNDER MY HAND THIS _____ DAY OF _____, 20_____.

BREATH TEST OPERATOR

I HAVE RECEIVED A COPY OF THIS CERTIFICATE OF ANALYSIS _____ SUBJECT'S SIGNATURE

SUBJECT REFUSED TO SIGN FOR COPY OF CERTIFICATE OF ANALYSIS _____ OPERATOR'S SIGNATURE

DBF
JPD



IntoxNet MIS Report

Report Generated 11 Jun 2015 at 09:37

Test Results

Instrument Serial Number 010614

Test # 002148 Subject Test

Test Location 1 Department of
Test Date 11 Jun 2015

Test Location 2 Forensic Science
Test Time 08:58
Remote/Local Local

Test Location 3
System Check Passed

Operator's Last Name HALL
Agency DFS Central Lab

Operator's First Name DONALD

Operator's Middle Initial B

Card Serial Number 121250

License Number 21250

Effective Date 10/01/2014

Expiration Date 10/01/2016

Subject's Last Name INSTRUMENT

Subject's First Name TEST

Subject's Middle Initial

Subject's Date of Birth 00/00/0000

Subject's Sex Male

Driver's License Number

Driver's License Expiration 00/00/0000

Driver's License State

Court Name DFS

End Date 11 Jun 2015 End Time 09:06

Result Time 09:05

Result Date 11 Jun 2015 Result 0.00

Data Type DIAG

Sample Value Pass

Sample Time 08:59

Data Type BLK

Sample Value 0.000

Sample Time 09:00

Data Type CHK

Sample Value 0.099

Sample Time 09:00

Data Type BLK

Sample Value 0.000

Sample Time 09:02

Data Type SUBJ

Sample Value 0.000

Sample Time 09:02

Data Type BLK

Sample Value 0.000

Sample Time 09:04

Data Type SUBJ

Sample Value 0.000

Sample Time 09:05

Data Type BLK

Sample Value 0.000

Sample Time 09:06

Standard Type Dry Gas Std

Standard Value 0.098

Standard Lot Number AG332301-10

Standard Expiration Date 11/19/2015

Tank Pressure 1022

Barometric Pressure 748 mmHg

Blow Sample Number 1 Blow Duration 4.37 sec

Blow Volume 2143 cc End-of-Blow Time 09:02

Blow Sample Number 2 Blow Duration 3.96 sec

Blow Volume 1984 cc End-of-Blow Time 09:05

Tamper Evident Stamp eae98bc5

Test Status Code 0

Test Status Success

DBH
JPD

Intox EC/IR-II: Accuracy Check

Department of Forensic Science

Serial Number: 010614 Test Number: 2149
Test Date: 06/11/2015 Test Time: 09:07 EDT
Dry Gas Target: 0.098

Lot Number: AG332301-10 Exp Date: 11/19/2015
Tank Pressure: 1017 psi Barometric Pressure: 748 mmHg
System Check: *Passed*

Test	g/210L	Time
BLK	0.000	09:08
CHK	0.098	09:09
BLK	0.000	09:10
CHK	0.098	09:11
BLK	0.000	09:13
CHK	0.098	09:13

Test Status: *Success*

Calibration CRC: F80233C6

DBH
JFD

Intox EC/IR-II: Accuracy Check

Department of Forensic Science

Serial Number: 010614 Test Number: 2150
Test Date: 06/11/2015 Test Time: 09:16 EDT
Dry Gas Target: 0.295
Lot Number: AG506501-12 Exp Date: 03/06/2017
Tank Pressure: 335 psi Barometric Pressure: 748 mmHg
System Check: *Passed*

Test	g/210L	Time
BLK	0.000	09:17
CHK	0.297	09:17
BLK	0.000	09:19
CHK	0.297	09:20
BLK	0.000	09:22
CHK	0.297	09:22

Test Status: *Success*

Calibration CRC: F80233C6