

DBH
TW

Department of Forensic Science
Breath Alcohol Section
Instrument Location History
1/1/1990 To 6/4/2015

Instrument Serial Number: 010616

As of 04-Jun-15

Start Date	End Date	Agency Name
03-Jun-15		DFS Central Lab
07-Jan-15	03-Jun-15	Colonial Beach PD
16-Dec-14	07-Jan-15	DFS Central Lab
08-Aug-14	16-Dec-14	Henrico County PD
23-Jul-14	08-Aug-14	DFS Central Lab
06-Jan-09	23-Jul-14	King George County Sheriff's Office
15-Aug-08	06-Jan-09	DFS Central Lab

DBH
TLW

INTOX EC/IR II Quality Assurance Worksheet

Instrument Serial Number **010616** Worksheet Start Date **6/3/2015**
 Location **Colonial Beach PD**
 Address **907 McKinney Blvd Colonial Beach VA 22443**
 DFS Technician **Donald Hall** License No. **21250**

Laboratory On-Site

Site Specification: No detrimental environmental conditions exist.

Instrument Barometer (mm HG) **765** Reference Barometer (mm HG) **765**
 Reference Barometer(RB)Serial # **009109** RB Calibration Due **7/28/2015**

Measurement Assurance Check


Standard (sea level)	PA Target	minimum	maximum	Sample 1	Sample 2	Sample 3
0.300	0.302	0.292	0.311	0.303	0.303	0.302
Precision		sample min	sample max			
0.001		0.302	0.303			

Standard (sea level)	PA Target	minimum	maximum	Sample 1	Sample 2	Sample 3
0.100	0.101	0.097	0.104	0.101	0.101	0.101
Precision		sample min	sample max			
0		0.101	0.101			

Dry gas standard Lot No. (with tank no.) **AG407801-15**

- Replaced dry gas standard (+O-ring)
- Installed at Location
- Removed to DFS-Central

- Supplies
- Mouthpieces
- Certificates of Analysis
- Operator Worksheet
- Other:



Notes:



Instrument Serial Number

010616

Certification Date

Calibrated

Certified

Measurement Assurance Check

Instrument Test

Troubleshooting/Maintenance Worksheet Completed

(Successful completion denotes satisfactory condition of the item.)

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.250		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.150		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.080		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.020		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

All measurements are in g/210L

Estimation of Uncertainty of Measurement and traceability records are located within the Breath Alcohol Section

DFS Technician Donald B Hall Date 6/4/15

Issuing Analyst [Signature] Date 6/4/15

JPD

DBH
Tlw



COMMONWEALTH OF VIRGINIA
DEPARTMENT OF FORENSIC SCIENCE

CERTIFICATE OF BLOOD ALCOHOL ANALYSIS
AS DETERMINED BY A CHEMICAL TEST OF THE ACCUSED'S BREATH

NAME OF ACCUSED INSTRUMENT, TEST,	NAME OF COURT DFS
--	--------------------------

BREATH ANALYSIS

SAMPLE EXAMINED AND TEST CONDUCTED BY HALL, DONALD, B		AGENCY DFS Central Lab
DFS LICENSE NUMBER 21250	LICENSE EXPIRES 10/01/2016	DATE TEST CONDUCTED 06/03/2015
TEST EQUIPMENT NUMBER 010616	[REDACTED]	

RESULTS: TIME SAMPLE TAKEN 11:38 EDT

SAMPLE'S ALCOHOL CONTENT 0.00 GRAMS PER 210 LITERS OF BREATH

ATTEST:

I CERTIFY THAT THE ABOVE IS AN ACCURATE RECORD OF THE TEST CONDUCTED; THAT THE TEST WAS CONDUCTED WITH THE TYPE OF EQUIPMENT AND IN ACCORDANCE WITH THE METHODS APPROVED BY THE DEPARTMENT OF FORENSIC SCIENCE; THAT THE TEST WAS CONDUCTED IN ACCORDANCE WITH THE DEPARTMENT'S SPECIFICATIONS; THAT PRIOR TO ADMINISTRATION OF THE TEST THE ACCUSED WAS ADVISED OF HIS RIGHT TO OBSERVE THE PROCESS AND SEE THE BLOOD ALCOHOL READING ON THE EQUIPMENT USED TO PERFORM THE BREATH TEST, AND THAT I POSSESS A VALID LICENSE TO CONDUCT SUCH TEST, GIVEN UNDER MY HAND THIS _____ DAY OF _____, 20_____.

BREATH TEST OPERATOR

I HAVE RECEIVED A COPY OF THIS CERTIFICATE OF ANALYSIS _____ SUBJECT'S SIGNATURE

SUBJECT REFUSED TO SIGN FOR COPY OF CERTIFICATE OF ANALYSIS _____ OPERATOR'S SIGNATURE



IntoxNet MIS Report
Report Generated 03 Jun 2015 at 12:07

DBH
Tlw

Test Results

Instrument Serial Number 010616

<p>Test # 000952 Subject Test</p> <p>Test Location 1 Department of Test Date 03 Jun 2015</p> <p>Operator's Last Name HALL Agency DFS Central Lab</p> <p>Card Serial Number 121250</p> <p>Subject's Last Name INSTRUMENT Subject's Middle Initial</p> <p>Driver's License Number Driver's License State</p> <p>End Date 03 Jun 2015 End Time 11:39</p> <p>Data Type DIAG Sample Value Pass</p> <p>Data Type BLK Sample Value 0.000</p> <p>Data Type CHK Sample Value 0.101</p> <p>Data Type BLK Sample Value 0.000</p> <p>Data Type SUBJ Sample Value 0.000</p> <p>Data Type BLK Sample Value 0.000</p> <p>Data Type SUBJ Sample Value 0.000</p> <p>Data Type BLK Sample Value 0.000</p> <p>Standard Type Dry Gas Std</p> <p>Standard Lot Number AG407801-15</p> <p> Tank Pressure 345</p> <p>Blow Sample Number 1 Blow Duration 3.91 sec</p> <p>Blow Sample Number 2 Blow Duration 5.23 sec</p> <p> Tamper Evident Stamp 10b726d2</p> <p> Test Status Success</p>	<p>Test Location 2 Forensic Science</p> <p> Test Time 11:32</p> <p> Remote/Local Local</p> <p>Operator's First Name DONALD</p> <p> License Number 21250</p> <p> Effective Date 10/01/2014</p> <p>Subject's Date of Birth 00/00/0000</p> <p> Driver's License Expiration 00/00/0000</p> <p> Court Name DFS</p> <p>Result Time 11:38</p> <p>Sample Value 0.000</p> <p>Sample Value 0.101</p> <p>Sample Value 0.000</p> <p>Standard Value 0.101</p> <p>Standard Expiration Date 03/19/2016</p> <p> Barometric Pressure 765 mmHg</p> <p>Blow Volume 1826 cc</p> <p>Blow Volume 2318 cc</p> <p> Test Status Code 0</p>	<p>Test Location 3</p> <p> System Check Passed</p> <p>Operator's Middle Initial B</p> <p>Expiration Date 10/01/2016</p> <p>Subject's First Name TEST</p> <p>Subject's Sex Male</p> <p>Result 0.00</p> <p>Sample Time 11:32</p> <p>Sample Time 11:33</p> <p>Sample Time 11:33</p> <p>Sample Time 11:35</p> <p>Sample Time 11:35</p> <p>Sample Time 11:35</p> <p>Sample Time 11:36</p> <p>Sample Time 11:38</p> <p>Sample Time 11:39</p>
--	--	--

DGH
TLW

Intox EC/IR-II: Accuracy Check

Department of Forensic Science

Serial Number: 010616 Test Number: 953

Test Date: 06/03/2015 Test Time: 11:41 EDT

Dry Gas Target: 0.101

Lot Number: AG407801-15 Exp Date: 03/19/2016

Tank Pressure: 345 psi Barometric Pressure: 765 mmHg

System Check: Passed

Test	g/210L	Time
BLK	0.000	11:42
CHK	0.101	11:42
BLK	0.000	11:44
CHK	0.101	11:45
BLK	0.000	11:46
CHK	0.101	11:47

Test Status: Success

Calibration CRC: E218AE08

DBH
TW

intox EC/IR-II: ACCURACY Check

Department of Forensic Science
Serial Number: 010616 Test Number: 954
Test Date: 06/03/2015 Test Time: 11:49 EDT
Dry Gas Target: 0.302
Lot Number: AG506501-12 Exp Date: 03/06/2017
Tank Pressure: 448 psi Barometric Pressure: 765 mmHg
System Check: Passed

Test	g/210L	Time
BLK	0.000	11:50
CHK	0.303	11:51
BLK	0.000	11:53
CHK	0.303	11:53
BLK	0.000	11:55
CHK	0.302	11:56

Test Status: Success

Calibration CRC: E218AE08