

Department of Forensic Science
Breath Alcohol Section
Instrument Location History
1/1/1990 To 6/26/2015

Handwritten initials:
HMS
mmB

Instrument Serial Number: 010492

As of 26-Jun-15

Start Date	End Date	Agency Name
23-Jun-15		DFS Central Lab
22-Apr-15	23-Jun-15	Blacksburg PD
07-Apr-15	22-Apr-15	DFS Central Lab
12-Dec-14	07-Apr-15	King & Queen Sheriff's Office
02-Dec-14	12-Dec-14	DFS Central Lab
18-Jun-14	02-Dec-14	Henry County Sheriff's Office
28-May-14	18-Jun-14	DFS Central Lab
10-Apr-13	28-May-14	Amherst County SO
28-Mar-13	10-Apr-13	DFS Central Lab
08-May-09	28-Mar-13	Amherst County SO
19-May-08	08-May-09	DFS Central Lab

Handwritten initials and 'mmB'

INTOX EC/IR II Quality Assurance Worksheet

Instrument Serial Number: 010492 Worksheet Start Date: 6/23/2015

Location: Blacksburg Police Department

Address: 200 Clay St. SW, Blacksburg, VA 24060

DFS Technician: Heather Stanton License No.: 18910

Laboratory On-Site

Site Specification: No detrimental environmental conditions exist.

Instrument Barometer (mm HG): 707 Reference Barometer (mm HG): 709

Reference Barometer(RB)Serial #: 009113 RB Calibration Due: 4/16/2016

Measurement Assurance Check

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.300		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.100		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Dry gas standard Lot No. (with tank no.) AG406301-25

- Replaced dry gas standard (+O-ring)
 - Installed at Location
 - Removed to DFS-Central
- Supplies
Mouthpieces
Certificates of Analysis
Operator Worksheet
Other:

Notes:

Unable to perform instrument test and measurement assurance checks due to instrument not warming up.

Instrument Serial Number

010492

Certification Date

[Faded]

Calibrated

Certified

Measurement Assurance Check

Instrument Test

Troubleshooting/Maintenance Worksheet Completed

(Successful completion denotes satisfactory condition of the item.)

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.250		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.150		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.080		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.020		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

All measurements are in g/210L

Estimation of Uncertainty of Measurement and traceability records are located within the Breath Alcohol Section

DFS Technician

Heath W. Stark

Date

6/26/15

Issuing Analyst

Michelle B. [unclear]

Date

6/26/15

Seb

Department of Forensic Science
Breath Alcohol Section
Trouble Call Log
6/22/2015 To 6/26/2015

JMS
mmb

Instrument Serial Number: 010492

26-Jun-15

Date: 22-Jun-15 Time: 3:35 PM Remote: True Tech: HMS (18910) Location: Blacksburg PD

Indication: OPERATOR QUESTION / COMMENT

Problem: OPERATOR STATED THAT THE INSTRUMENT WAS NOT WARMING UP.

Technician Response: SITE VISIT CONDUCTED.

Resolution REMOVED TO DFS-Central
