

DBH
JPD

Department of Forensic Science
Breath Alcohol Section
Instrument Location History
1/1/1990 To 7/7/2015

Instrument Serial Number: 010563 As of 07-Jul-15

Start Date	End Date	Agency Name
07-Jul-15		Virginia Beach PD
19-Jun-15	07-Jul-15	DFS Central Lab
28-Jan-15	19-Jun-15	Alexandria PD
14-Jan-15	28-Jan-15	DFS Central Lab
04-Aug-14	14-Jan-15	Buckingham County SO
22-Jul-14	04-Aug-14	DFS Central Lab
23-Nov-09	22-Jul-14	Caroline County SO
29-Sep-09	23-Nov-09	DFS Central Lab
20-Apr-09	29-Sep-09	Chesapeake PD
23-Jun-08	20-Apr-09	DFS Central Lab

DBH
JPD

INTOX EC/IR II Quality Assurance Worksheet

Instrument Serial Number: 010563 Worksheet Start Date: 7/7/2015

Location: Virginia Beach Jail

Address: 2501 James Madison Blvd Virginia Beach VA 23456

DFS Technician: Donald Hall License No.: 21250

Laboratory On-Site

Site Specification: No detrimental environmental conditions exist.

Instrument Barometer (mm HG): 762 Reference Barometer (mm HG): 763

Reference Barometer(RB)Serial #: 009109 RB Calibration Due: 7/28/2015

Measurement Assurance Check

Standard (sea level)	PA Target	minimum	maximum	Sample 1	Sample 2	Sample 3
0.300	0.301	0.291	0.310	0.305	0.304	0.304
Precision		sample min	sample max			
0.001		0.304	0.305			

Standard (sea level)	PA Target	minimum	maximum	Sample 1	Sample 2	Sample 3
0.100	0.100	0.097	0.103	0.101	0.101	0.101
Precision		sample min	sample max			
0		0.101	0.101			

Dry gas standard Lot No. (with tank no.) AG509101-33

- Replaced dry gas standard (+O-ring)
 - Installed at Location
 - Removed to DFS-Central
- Supplies
- | | |
|--------------------------|-----|
| Mouthpieces | 300 |
| Certificates of Analysis | 200 |
| Operator Worksheet | 200 |
| Other: | |

Notes:

Instrument Serial Number 010563 Certification Date

Calibrated Certified Measurement Assurance Check

Instrument Test Troubleshooting/Maintenance Worksheet Completed

(Successful completion denotes satisfactory condition of the item.)

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.250		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.150		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.080		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.020		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

All measurements are in g/210L
Estimation of Uncertainty of Measurement and traceability records are located within the Breath Alcohol Section

DFS Technician Donald B. Hill Date 7/7/15

Issuing Analyst Jennie D. Hy Date 7/8/15

DBH
JPD

Intox EC/IR-11: Accuracy Check

Department of Forensic Science

Serial Number: 010563 Test Number: 635

Test Date: 07/07/2015 Test Time: 12:11 EDT

Dry Gas Target: 0.301

Lot Number: AG506501-07 Exp Date: 03/06/2017

Tank Pressure: 531 psi Barometric Pressure: 762 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	12:12
CHK	0.305	12:12
BLK	0.000	12:14
CHK	0.304	12:15
BLK	0.000	12:16
CHK	0.304	12:17

Test Status: *Success*

Calibration CRC: 05F416BB

DBH
JPD

Intox EC/IR-II: Accuracy Check

Department of Forensic Science

Serial Number: 010563 Test Number: 636

Test Date: 07/07/2015 Test Time: 12:21 EDT

Dry Gas Target: 0.100

Lot Number: AG509101-33 Exp Date: 04/01/2017

Tank Pressure: 1115 psi Barometric Pressure: 762 mmHg

System Check: Passed

Test	g/210L	Time
BLK	0.000	12:21
CHK	0.101	12:22
BLK	0.000	12:24
CHK	0.101	12:25
BLK	0.000	12:26
CHK	0.101	12:27

Test Status: Success

Calibration CRC: 05F416BB

DBH
JPD



COMMONWEALTH OF VIRGINIA
DEPARTMENT OF FORENSIC SCIENCE

CERTIFICATE OF BLOOD ALCOHOL ANALYSIS
AS DETERMINED BY A CHEMICAL TEST OF THE ACCUSED'S BREATH

NAME OF ACCUSED INSTRUMENT TEST	NAME OF COURT DFS
--	--------------------------

BREATH ANALYSIS

SAMPLE EXAMINED AND TEST CONDUCTED BY HALL, DONALD, R		AGENCY DFS Central Lab
DFS LICENSE NUMBER 21250	LICENSE EXPIRES 10/01/2016	DATE TEST CONDUCTED 07/07/2015
TEST EQUIPMENT NUMBER 010563		

RESULTS: TIME SAMPLE TAKEN 12:39 EDT

SAMPLE'S ALCOHOL CONTENT 0.00 GRAMS PER 210 LITERS OF BREATH

ATTEST:

I CERTIFY THAT THE ABOVE IS AN ACCURATE RECORD OF THE TEST CONDUCTED; THAT THE TEST WAS CONDUCTED WITH THE TYPE OF EQUIPMENT AND IN ACCORDANCE WITH THE METHODS APPROVED BY THE DEPARTMENT OF FORENSIC SCIENCE; THAT THE TEST WAS CONDUCTED IN ACCORDANCE WITH THE DEPARTMENT'S SPECIFICATIONS; THAT PRIOR TO ADMINISTRATION OF THE TEST THE ACCUSED WAS ADVISED OF HIS RIGHT TO OBSERVE THE PROCESS AND SEE THE BLOOD ALCOHOL READING ON THE EQUIPMENT USED TO PERFORM THE BREATH TEST, AND THAT I POSSESS A VALID LICENSE TO CONDUCT SUCH TEST, GIVEN UNDER MY HAND THIS _____ DAY OF _____, 20 _____.

BREATH TEST OPERATOR

I HAVE RECEIVED A COPY OF THIS CERTIFICATE OF ANALYSIS _____ SUBJECT'S SIGNATURE

SUBJECT REFUSED TO SIGN FOR COPY OF CERTIFICATE OF ANALYSIS _____ OPERATOR'S SIGNATURE

DBH
JPD



IntoxNet MIS Report

Report Generated 07 Jul 2015 at 17:21

Test Results

Instrument Serial Number 010563

Test # 000637 Subject Test

Test Location 1 Department of
Test Date 07 Jul 2015

Test Location 2 Forensic Science
Test Time 12:32
Remote/Local Local

Test Location 3
System Check Passed

Operator's Last Name HALL
Agency DFS Central Lab

Operator's First Name DONALD

Operator's Middle Initial B

Card Serial Number 121250

License Number 21250

Effective Date 10/01/2014

Expiration Date 10/01/2016

Subject's Last Name INSTRUMENT

Subject's First Name TEST

Subject's Middle Initial

Subject's Date of Birth 00/00/0000

Subject's Sex Male

Driver's License Number

Driver's License Expiration 00/00/0000

Driver's License State

Court Name DFS

End Date 07 Jul 2015 End Time 12:40

Result Time 12:39

Result Date 07 Jul 2015 Result 0.00

Data Type DIAG

Sample Value Pass

Sample Time 12:32

Data Type BLK

Sample Value 0.000

Sample Time 12:33

Data Type CHK

Sample Value 0.101

Sample Time 12:34

Data Type BLK

Sample Value 0.000

Sample Time 12:35

Data Type SUBJ

Sample Value 0.000

Sample Time 12:36

Data Type BLK

Sample Value 0.000

Sample Time 12:37

Data Type SUBJ

Sample Value 0.000

Sample Time 12:39

Data Type BLK

Sample Value 0.000

Sample Time 12:40

Standard Type Dry Gas Std

Standard Value 0.100

Standard Lot Number AG509101-33

Standard Expiration Date 04/01/2017

Tank Pressure 1120

Barometric Pressure 762 mmHg

Blow Sample Number 1 Blow Duration 3.74 sec

Blow Volume 2012 cc End-of-Blow Time 12:36

Blow Sample Number 2 Blow Duration 3.40 sec

Blow Volume 1935 cc End-of-Blow Time 12:39

Tamper Evident Stamp cf122183

Test Status Code 0

Test Status Success