

Department of Forensic Science
Breath Alcohol Section
Trouble Call Log
7/10/2015 To 7/10/2015

HMS
JPD
MJK

Instrument Serial Number: 010582

10-Jul-15

Date: 10-Jul-15 Time: 9:15 AM Remote: True Tech: HMS (18910) Location: Pittsylvania County SO

Indication: PRINTER NOT FUNCTIONING PROPERLY

Problem: PRINTER IS PRINTING OFF CENTER

Technician Response: AGENCY CONTACTED VIA PHONE, ISSUE RESOLVED AND IS PRINTER FUNCTIONING PROPERLY.

Resolution NO FURTHER ACTION REQUIRED
