

Department of Forensic Science
Breath Alcohol Section
Instrument Location History
1/1/1990 To 7/2/2015

JMS
TLW

Instrument Serial Number: 010603

As of 02-Jul-15

Start Date	End Date	Agency Name
01-Jul-15		DFS Central Lab
28-May-15	01-Jul-15	Rappahannock Shenandoah Warren Regional Jail
19-Dec-14	28-May-15	DFS Central Lab
16-Dec-14	19-Dec-14	Chesterfield County PD
26-Aug-14	16-Dec-14	DFS Central Lab
31-Mar-14	26-Aug-14	Fairfax County PD
25-Sep-13	31-Mar-14	DFS Central Lab
14-Mar-13	25-Sep-13	Fairfax County PD
05-Nov-12	14-Mar-13	DFS Central Lab
24-May-12	05-Nov-12	Virginia Beach PD
27-Mar-12	24-May-12	DFS Central Lab
22-Nov-11	27-Mar-12	Richmond PD
13-Sep-10	22-Nov-11	DFS Central Lab
04-Jun-10	13-Sep-10	Intoximeters, Inc.
12-May-09	04-Jun-10	DFS Central Lab
18-Mar-09	12-May-09	Norfolk PD
04-Aug-08	18-Mar-09	DFS Central Lab

THS
TJW

INTOX EC/IR II Quality Assurance Worksheet

Instrument Serial Number: 010603 Worksheet Start Date: 7/1/2015

Location: Rappahannock Shenandoah Warren Regional Jail

Address: 6601 Winchester Rd Front Royal VA 22630

DFS Technician: Heather Stanton License No.: 18910

Laboratory On-Site

Site Specification: No detrimental environmental conditions exist.

Instrument Barometer (mm HG): [blacked out] Reference Barometer (mm HG): [blacked out]

Reference Barometer(RB)Serial #: 009113 RB Calibration Due: 4/16/2016


Measurement Assurance Check

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.300		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.100		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Dry gas standard Lot No. (with tank no.) AG332301-24

- Replaced dry gas standard (+O-ring)
 - Installed at Location
 - Removed to DFS-Central
- Supplies
Mouthpieces
Certificates of Analysis
Operator Worksheet
Other:



Notes:

Unable to check barometric pressure or perform instrument test and measurement assurance checks due to power loss to instrument.

[Faint, illegible text]

Instrument Serial Number

010603

Certification Date

Calibrated

Certified

Measurement Assurance Check

Instrument Test

Troubleshooting/Maintenance Worksheet Completed

(Successful completion denotes satisfactory condition of the item.)

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.250		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.150		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.080		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.020		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

All measurements are in g/210L

Estimation of Uncertainty of Measurement and traceability records are located within the Breath Alcohol Section

DFS Technician

Heath W. Stach

Date

7/2/15

Issuing Analyst

[Signature]

Date

7/2/15

MSK