



ANSI National Accreditation Board

**SCOPE OF ACCREDITATION TO:
ISO/IEC 17025:2017
ANAB Forensic Testing & Calibration AR 3125:2019**

**Virginia Department of Forensic Science
Breath Alcohol Laboratory**

737 North 5th Street
Richmond, Virginia 23219 USA

FORENSIC CALIBRATION

Expiry Date: 30 September 2022

Certificate Number: FC-0004

Discipline – Toxicology

Component Calibrated¹	Calibration Method	Breath Alcohol Instrument Reporting Range	Expanded Uncertainty of Measurement²	Metrological Traceability
Electrochemical Oxidation ³	Breath Alcohol Procedures Manual - Certification Procedure for the Intox EC/IR II	0.000-0.500 g ethanol/210 L vapor	$\pm 0.003\text{g}$ or 3% whichever is greater ethanol/210L vapor 99.73%, $k=3$	Barometer Certified Reference Material

Note 1: Component calibrated indicates the breath alcohol measuring instrument detector(s) used for measuring breath alcohol. Detector(s) used for qualitative purposes will not be listed on the Scope of Accreditation.

Note 2: The expanded uncertainty is expressed at the stated coverage probability (*e.g.*, 95.45%) and coverage factor (*e.g.*, $k = 2$). The actual measurement uncertainty reported on a certificate of calibration will be equal to or greater than the expanded uncertainty of measurement stated on this scope.

Note 3: This instrument also employs Infrared Spectrometry for qualitative purposes.

Pamela L. Sale
Vice President, Forensics